



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION

Toxicology Section/Breath Alcohol Program

SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Aaron R. Hummel</u>		
Location: <u>GFSO</u>		
Intoxilyzer® Serial Number: <u>80-003057</u> ^{DK}		
Gas Lot Number: <u>659357</u>	Gas Cylinder Number: <u>26</u>	Expiration Date: <u>9/3/11</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>80</u> %		
Test 2: 0. <u>80</u> %		
Test 3: 0. <u>81</u> %		
(ATTACH TEST RECORD)		

Chemical Test Operator Signature

9/2/10
Date

Deb Kashner
Reviewed By

07 Sept 2010
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003057
Location = GFSO 8164.13.00 06/09
09/02/2010 09:37

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:38
02 Std. Gas	0.080	09:38
03 Room Air	0.000	09:39
04 Std. Gas	0.080	09:39
05 Room Air	0.000	09:39
06 Std. Gas	0.081	09:40
07 Room Air	0.000	09:40


Lot No = 659357

Cyl No = 26

Exp Date = 09/03/2011

County = 18

Oper No. = 049907



Operator Signature
AARON HUMMEL

Remarks:

ACA Test

Form 106-I8000

80-003057
P. 106 of 106
JFM