



**ETHANOL BREATH STANDARD CYLINDER REPORT**

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 58282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Lt. Tracy L. Meidinger #258</u>		
Location: <u>UND-PD</u>		
Intoxilyzer® Serial Number: <u>80-004934</u>		
Gas Lot Number: <u>659358</u>	Gas Cylinder Number: <u>31</u>	Expiration Date: <u>09-30-2011</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: <u>0.079</u> %		
Test 2: <del>0.079</del> % <u>.080% TLM</u>		
Test 3: <u>0.079</u> %		
(ATTACH TEST RECORD)		

Lt. Tracy L. Meidinger #258  
Chemical Test Operator Signature

09-07-2010  
Date

Deb Kashur  
Reviewed By

15 Sept 2010  
Date

80-004934  
p. 54 of 56  
FK



**ETHANOL BREATH STANDARD CYLINDER REPORT**  
OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Lt. Tracy L. Meidinger #258</u>		
Location: <u>UND-PD</u>		
Intoxilyzer® Serial Number: <u>80-004934</u>		
Gas Lot Number: <u>659358</u>	Gas Cylinder Number: <u>31</u>	Expiration Date: <u>09-30-2011</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>079</u> %		
Test 2: 0. <u>079</u> %		
Test 3: 0. <u>079</u> %		
(ATTACH TEST RECORD)		

Lt. Tracy L. Meidinger #258  
Chemical Test Operator Signature

09-07-2010  
Date

\_\_\_\_\_  
Reviewed By

\_\_\_\_\_  
Date

80-004934  
P. 55 of 56  
KM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004934  
Location = UNDP      8164.13.00 06/09  
09/07/2010      01:00

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	01:01
02 Std. Gas	0.079	01:01
03 Room Air	0.000	01:02
04 Std. Gas	0.080	01:02
05 Room Air	0.000	01:03
06 Std. Gas	0.079	01:03
07 Room Air	0.000	01:03

Lot No = 659358

Cyl No = 31

Exp Date = 09/03/2011

County = 18

Oper No. = 119409

*Tracy L. Meidinger #258*

Operator Signature  
TRACY L. MEIDINGER

Remarks:

Form 106-I8000

ACA 9-7-10

80-004934  
P. 56 of 56  
KM