



ETHANOL BREATH STANDARD CYLINDER REPORT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

| | | |
|--|--------------------------------|----------------------------------|
| Chemical Test Operator Name: <u>Rosevaren, Danielle M.</u> | | |
| Location: <u>CARR</u> | | |
| Intoxilyzer® Serial Number: <u>80-004944</u> | | |
| Gas Lot Number: <u>709352</u> | Gas Cylinder Number: <u>12</u> | Expiration Date: <u>12/23/11</u> |
| <input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed | | |
| Instrument Test Results (Report to 3 Digits; Example, 0.081%) Test 1: 0. <u>080</u> % Test 2: 0. <u>079</u> % Test 3: 0. <u>079</u> % | | |
| (ATTACH TEST RECORD) | | |

Danielle Rosevaren
 Chemical Test Operator Signature

09/07/10
 Date

Deb Kashner
 Reviewed By

09 Sept 2010
 Date

80-004944
 P. 53 of 54
 KM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = CARR 8164.13.00 06/09
09/07/2010 14:55

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 14:56 |
| 02 Std. Gas | 0.080 | 14:56 |
| 03 Room Air | 0.000 | 14:57 |
| 04 Std. Gas | 0.079 | 14:57 |
| 05 Room Air | 0.000 | 14:58 |
| 06 Std. Gas | 0.079 | 14:58 |
| 07 Room Air | 0.000 | 14:59 |

Lot No = 709352
Cyl No = 12
Exp Date = 12/23/2011
County = 16 Oper No. = 100754

Danielle Roseware

Operator Signature
DANIELLE M. ROSEWAREN

Remarks:

Form 106-I8000

80-004944
P. 54 of 54
JCM