



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>David M. Walker</u>		
Location: <u>Dickinson TSC</u>		
Intoxilyzer® Serial Number: <u>80 - 004947</u>		
Gas Lot Number: <u>659358</u>	Gas Cylinder Number: <u>23</u>	Expiration Date: <u>9-3-11</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: <u>0.082</u> %		
Test 2: <u>0.079</u> %		
Test 3: <u>0.080</u> %		
(ATTACH TEST RECORD)		

David M. Walker
Chemical Test Operator Signature

9-2-10
Date

Deb Kashner
Reviewed By

09 Sept 2010
Date

80-004947
P. 49 of 50
EM


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = DKPD 8164.13.00 06/09
09/07/2010 08:37

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:38
02 Std. Gas	0.080	08:38
03 Room Air	0.000	08:38
04 Std. Gas	0.079	08:39
05 Room Air	0.000	08:39
06 Std. Gas	0.080	08:40
07 Room Air	0.000	08:40

Lot No = 659358
Cyl No = 23
Exp Date = 09/03/2011
County = 45 Oper No. = 068332


Operator Signature
DAVID M. WALLACE

Remarks:

Form 106-I8000

80-004947
P. 50 of 50
DM