

INTOXILYZER 8000  
REPAIR AND MAINTENANCE

Serial No.: 80-004954

Location: TOXL

- |  | Check When Done |
|--|-----------------|
| 1. Is a three-pronged grounded outlet used?  | <u>✓</u>        |
| 2. Is the warm-up time less than 20 minutes?   | <u>✓</u>        |
| 3. Is the breath tube heated?  | <u>✓</u>        |
| 4. Is the diagnostic check complete?   | <u>✓</u>        |
| 5. Is the time, date, and year correct (re-set if necessary)?  | <u>✓</u>        |
| 6. Print test. (Attach test record.)   | <u>✓</u>        |
| 7. Calibration Check. Use Ethanol Breath Standard (EBS). Do three sets in ACA mode. (Attach test records.) |                 |

Lot No. 23411080A1 Cylinder No. 14 Expir. Date 10/01/2013

Test 1 0.081 Test 4 0.081 Test 7 0.081

Test 2 0.080 Test 5 0.081 Test 8 0.081

Test 3 0.080 Test 6 0.080 Test 9 0.081

Average 0.081

- |  |          |
|--|----------|
| 8. Low AC. Use < 0.03 AC in ACA mode. (Attach test record.)                                      | <u>✓</u> |
| 9. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)                              | <u>✓</u> |
| 10. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode. (Attach test record.) | <u>✓</u> |
| 11. RFI check. Run CMS mode. Key radio on first room air. (Attach test record.)                  | <u>✓</u> |

Electrical Repairs:

1. Was the calibration adjusted/performed? (Attach test record.) NO
2. Was the light source changed? NO
3. Was the sample chamber cleaned? NO

Other Maintenance/Repairs:

1. Was the printer cleaned/repaired? NO
2. Was the solenoid cleaned? NO
3. Was new tubing installed? NO
4. Were additional parts replaced/changed/repaired? NO
5. Remarks/Maintenance record: check out after receiving  
from CMI.

07 Sept 2012  
Date

07 Sep. 12  
Date

Deb Kashur  
Field Inspector's Signature

Charles E. Edm  
Reviewed by



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004954  
Location = TOXL      8164.13.00 06/09  
09/06/2012      12:09

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:10
02 Std. Gas	0.081	12:10
03 Room Air	0.000	12:11
04 Std. Gas	0.080	12:11
05 Room Air	0.000	12:12
06 Std. Gas	0.080	12:12
07 Room Air	0.000	12:13

Lot No = 23411080A1

Cyl No = 14

Exp Date = 10/01/2013

County = 08

Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Calibration check 0.080 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004954  
Location = TOXL      8164.13.00 06/09  
09/06/2012      12:13

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:13
02 Std. Gas	0.081	12:14
03 Room Air	0.000	12:14
04 Std. Gas	0.081	12:15
05 Room Air	0.000	12:15
06 Std. Gas	0.080	12:16
07 Room Air	0.000	12:16

Lot No = 23411080A1

Cyl No = 14

Exp Date = 10/01/2013

County = 08

Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Calibration check 0.080 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004954  
Location = TOXL      8164.13.00 06/09  
09/06/2012      14:44

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:45
02 Std. Gas	0.081	14:45
03 Room Air	0.000	14:45
04 Std. Gas	0.081	14:46
05 Room Air	0.000	14:46
06 Std. Gas	0.081	14:47
07 Room Air	0.000	14:47

Lot No = 23411080A1  
Cyl No = 14  
Exp Date = 10/01/2013  
County = 08      Oper No. = 777777

Deb Kashur  
Operator Signature  
DEB KASHUR

Remarks:

Calibration check 0.080AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004954  
Location = TOXL      8164.13.00 06/09  
09/07/2012      08:26

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:27
02 Std. Sol.	0.019	08:27
03 Room Air	0.000	08:28
04 Std. Sol.	0.020	08:29
05 Room Air	0.000	08:29
06 Std. Sol.	0.020	08:30
07 Room Air	0.000	08:30

08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 08

Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

LOW AC 0.020AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004954  
Location = TOXL      8164.13.00 06/09  
09/07/2012      08:31

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:31
02 Std. Sol.	0.300	08:32
03 Room Air	0.000	08:33
04 Std. Sol.	0.299	08:33
05 Room Air	0.000	08:34
06 Std. Sol.	0.298	08:35
07 Room Air	0.000	08:35

08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 08

Oper No. = 777777

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Linearity test      0.300 AC*

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004954  
Location = TOXL      8164.13.00 06/09  
09/07/2012      08:37

Test	AC	Time
01 Room Air	0.000	08:38
02 *Subject Test	INT*	08:38
03 Room Air	0.000	08:39

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 23411080A1  
Cyl No = 14  
Expiration Date = 10/01/2013  
County = 08      Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Interferent check 0.100AC + acetone*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004954  
Location = TOXL      8164.13.00 06/09  
09/07/2012      08:39

Test	AC	Time
01 Diagnostic	OK	08:40
02 Room Air	RFI*	08:40
03 Room Air	0.000	08:40

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 23411080A1  
Cyl No = 14  
Expiration Date = 10/01/2013  
County = 08      Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kashur

Operator Signature  
DEB KASHUR

Remarks:

RFI check

Form 106-I8000