



ETHANOL GAS STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (10-2013)

Chemical Test Operator Name (Print) <i>Darcy W Aberle</i>		
Location <i>Williston LEC</i>	Intoxilyzer® Serial Number <i>80-004950</i>	
Gas Lot Number <i>13314080A2</i>	Gas Cylinder Number <i>51</i>	Gas Expiration Date <i>06/05/16</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. *0.082* AC
 - B. *0.082* AC
 - C. *0.081* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Darcy W Aberle</i>	Date <i>09-08-14</i> ✓
Reviewed By (Crime Laboratory Use Only) <i>Deb Kashur</i>	Date <i>15 Sept 14</i>


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004950
Location = WILL 8164.13.00 06/09
09/08/2014 09:19

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:19
02 Std. Gas	0.082	09:20
03 Room Air	0.000	09:20
04 Std. Gas	0.082	09:20
05 Room Air	0.000	09:21
06 Std. Gas	0.081	09:21
07 Room Air	0.000	09:22

Lot No = 13314080A2
Cyl No = 51
Exp Date = 06/05/2016
County = 53 Oper No. = 010101



Operator Signature
DARCY ABERLE

Remarks: