



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (5-2011)

Chemical Test Operator Name (Print) <i>Newman, Dustin B #030924</i>		
Location <i>Wattford City</i>	Intoxilyzer® Serial Number <i>80-084938</i>	
Gas Lot Number <i>11611080A2</i>	Gas Cylinder Number <i>114</i>	Gas Expiration Date <i>06/01/13</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. *0.079* AC
 - B. *0.079* AC
 - C. *0.079* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Breath Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Dustin B Newman</i>	Date <i>9-9-11</i>
Reviewed By (Crime Laboratory Use Only) <i>Deb Kashner</i>	Date <i>03 Apr 12</i>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004938
Location = WATF 8164.13.00 06/09
09/09/2011 21:03

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	21:03
02 Std. Gas	0.079	21:04
03 Room Air	0.000	21:04
04 Std. Gas	0.079	21:05
05 Room Air	0.000	21:05
06 Std. Gas	0.079	21:06
07 Room Air	0.000	21:06

Lot No = 11611080A2
Cyl No = 114
Exp Date = 06/01/2013
County = 27 Oper No. = 030924



Operator Signature
DUSTIN B. NEWMAN

Remarks: