



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (5-2011)

Chemical Test Operator Name (Print) <i>Dustin Plummer</i>		
Location <i>Mohl</i>	Intoxilyzer® Serial Number <i>80-004947</i>	
Gas Lot Number <i>11611680A2</i>	Gas Cylinder Number <i>143</i>	Gas Expiration Date <i>6-1-13</i>

Check When Done:

1. Scan/Enter Gas Cylinder Information (Level 1, Function S).

2. Perform an ACA Test (Level 1, Function C).

Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).

A. *0.079* AC

B. *0.079* AC

C. *0.078* AC

3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.

4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.

5. Send the Following to the Crime Laboratory:

A. Ethanol Breath Standard Cylinder Report (SFN59282, Form 105-G).

B. ACA Test Record.

Chemical Test Operator Signature <i>Dustin Plummer</i>	Date <i>9-9-11</i>
Reviewed By (Crime Laboratory Use Only) <i>Deb Kashner</i>	Date <i>03 Apr 2012</i>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = MOHL 8164.13.00 06/09
09/09/2011 12:05

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:06
02 Std. Gas	0.079	12:07
03 Room Air	0.000	12:07
04 Std. Gas	0.079	12:07
05 Room Air	0.000	12:08
06 Std. Gas	0.078	12:08
07 Room Air	0.000	12:09

Lot No = 11611080A2
Cyl No = 143
Exp Date = 06/01/2013
County = 38 Oper No. = 100914



Operator Signature
DUSTIN J. PLUMMER

Remarks:

Form 106-I8000

I, the undersigned Sheriff in and for Benville
County, North Dakota, do certify that the
foregoing is a true copy of an original document
remaining on file and record in my office.

WITNESS my hand and seal of said Court this
09 day of Sept, 2011.
Dustin J. Plummer, Sheriff
