



**ETHANOL BREATH STANDARD CYLINDER REPORT**  
 OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION  
 Toxicology Section/Breath Alcohol Program  
 SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>RAY Weber</u>		
Location: <u>TRAIL CO S.O.</u>		
Intoxilyzer® Serial Number: <u>80-004200</u>		
Gas Lot Number: <u>659358</u>	Gas Cylinder Number: <u>#45</u>	Expiration Date: <u>09/03/11</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>079</u> %		
Test 2: 0. <u>080</u> %		
Test 3: 0. <u>079</u> %		
(ATTACH TEST RECORD)		

Ray Weber  
 Chemical Test Operator Signature

9/11/2010  
 Date

Deb Kashur  
 Reviewed By

15 Sept 2010  
 Date

80-004200  
 P. 9/11/10  
 RM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004200  
Location = HILL      8164.13.00 06/09  
09/11/2010      23:01

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	23:02
02 Std. Gas	0.079	23:02
03 Room Air	0.000	23:03
04 Std. Gas	0.080	23:03
05 Room Air	0.000	23:03
06 Std. Gas	0.079	23:04
07 Room Air	0.000	23:04

Lot No = 659358  
Cyl No = 45  
Exp Date = 09/03/2011  
County = 49      Oper No. = 058405

*Ray Weber*

Operator Signature  
RAY G. WEBER

Remarks:

Form 106-I8000

*ACA Test*

*80-004200  
P. 92 of 92  
JRM*