

INTOXILYZER® 8000 INITIAL INSPECTION

Serial Number: 80-005954

1. Power On and Warm-up:

- a. Power On
- b. 5 min. Countdown from Initial Warm-up
- c. 1 min. Countdown to Begin a Test
- d. Forms Load Complete

2. Setup and Checks:

a. Level 1:

- 1) Passwords: Field Inspector (Level 2) Lab Personnel (Level 3)
- 2) Gas Cylinder: Gas Cylinder Setup (Level 1,S)
- 3) Printer: Printer Setup (Level 3,S,P)
 - Internal Printer Paper Feed
 - Internal Printer Test (Level 1,P)
 - External Printer Test (Level 1,P)
- 4) Barcode Reader (Level 1,B): Operator ID
 ND License

b. Level 2:

- 1) Continuous Purge (Level 2,A)
- 2) Diagnostic (Level 2,D)
- 3) Prelim Data Entry (Level 2,E)
- 4) Ser Num/Version Info (Level 2,V): Ser Num 80-005954
Version 8164.13

c. Level 3:

- 1) DVM Monitor (Level 3,D,D): 3: 12831 s: 1 c: 47.00
9: 13616 s: 1 B: 45 0
(RFI indicator with keyed radio) 1

- 2) Function (F) Keys: F5 – Tone
 F6 – Green LED
 F7 – Red LED
 F9 – Fan
 F10 – Pump
 F11 – Solenoid CEE

- 3) Atmospheric Monitor (Level 3,D,F): Atmo Sensor, P: 948 CF: 1.06
- 4) Flow Monitor (Level 3,D,F): Flow Sensor, R: 222 F: 0.000 L/s

5) Temperature Monitor (Level 3,D,T): C: 47.0 A: 49.0 I: 0.0
B: 45.0 V: 0.0

6) Tank Monitor (Level 3,D,G): P (gauge): 625 P (display): 631
7) General Setup (Level 3,S,G)
8) Configure Start Test (Level 3,S,T)
9) Network (Level 3,S,N)

3. Tests:

a. Configure Standard for Wet Bath Simulator (Level 1,S):

- 1) Zero Alcohol Test (Level 1,C):
Sim Ser #: DR5188 Lot #: Milli QH₂O Exp Date: NA
Known Value: 0.000 AC
Results: 0.000 AC 0.000 AC 0.000 AC
- 2) Low AC Test (Level 1,C):
Sim Ser #: DR3455 Lot #: 12210 Exp Date: 11-2-14
Known Value: 0.020 AC
Results: 0.018 AC 0.019 AC 0.019 AC
- 3) Linearity Test (Level 1,C):
Sim Ser #: DR3847 Lot #: 13020 Exp Date: 1-15-15^{CEC}
Known Value: 0.300 AC
Results: 0.0298 AC 0.300 AC 0.301 AC

b. Configure Dry Gas Standard (Level 1,S):

- 1) Interferent Detected Tests (Level 1,B):
 1. 0.10 AC + 0.05% Acetone
Sim Ser #: DR3377 Lot #: ICS#1 Exp Date: 4-9-14
Known Value: 0.10 AC + 0.05% ACETONE
 Display reads "Interferent Detect"
 2. Acetone
Sim Ser #: G8919 Lot #: ACE1 Exp Date: 8-27-15
Known Value: 0.05% ACETONE
 Display reads "Interferent Detect"
 3. Methanol
Sim Ser #: G10573 Lot #: MET1 Exp Date: 8-27-15
Known Value: 0.10% METHANOL
 Display reads "Interferent Detect"

4. Isopropanol
Sim Ser #: G9998 Lot #: IPA1 Exp Date: 8-27-15
Known Value: 0.10% ISOPROPRANOL
 Display reads "Interferent Detect"
- 2) Ambient Fail (Level 1,B): Display and test record are correct
3) Invalid Sample (Level 1,B): Display and test record are correct
4) Subject Test Timeout (Level 1,B): Display and test record are correct.
5) Sequence Aborted (Level 1,B): Display and test record are correct
6) Control Outside Tolerance (CMS mode): Display and test record are correct
7) RFI Detected Test (CMS mode): Display and test record are correct
8) Tank Pressure Below Min (CMS mode): Display reads correct
9) Subject Refused (CMS mode): Display and test record are correct
10) Difference Too Great (CMS mode): Display and test record are correct
11) Stability Test (Level 3,T,S):
Average 0.0808 Std Dev 0.0007 Rel Std Dev 0.89%

Maintenance/Repairs/Notes: _____

THIS UNIT IS ACCEPTABLE TO BE USED IN THE FIELD.

Charles E. Edr
Field Inspector's Signature

10SEP13
Date

Deb Kashner
Reviewed by

11 Sept 2013 ✓
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005954
Location = TOXL 8164.13.00 06/09
09/09/2013 16:04

***** Printer Test *****

12 char/line
abcdefghijkl
ABCDEFGHIJKL

16 char per line
abcdefghijklmnop
ABCDEFGHIJKLMNPO

21 chars per line
abcdefghijklmnopqrstu
ABCDEFGHIJKLMNOPQRSTU

24 characters per line
abcdefghijklmnopqrstuwx
ABCDEFGHIJKLMNOPQRSTUWX

32 characters per line
abcdefghijklmnopqrstuwx123456
ABCDEFGHIJKLMNOPQRSTUWXYZ123456

42 characters per line
abcdefghijklmnopqrstuwx1234567890abcdef
ABCDEFGHIJKLMNOPQRSTUWXYZ1234567890ABCDEF

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 23411080A1
Standard Cyl #? 32
Standard Expiration? 10/01/2013
Oper No? 666666

Flow Cal. Date: 06/21/2013
Slope 688
Intercept -700230

IR Calibration Date: 06/21/2013
3um 9um

0th Coef(*100): -23841 -21159

1st Coef(*100): 286265 140038

2nd Coef(*100): 2119 1082

H2O adj(mg/l*10k): 533 468

***** Test End *****



Operator Signature
CHARLES EDER

Remarks: INTERNAL
PRINT TEST

Form 106-18000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005954
Location = TOXL 8164.13.00 06/09
09/09/2013 16:07

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 23411080A1
Standard Cyl #? 32
Standard Expiration? 10/01/2013
Oper No? 666666

Flow Cal. Date: 06/21/2013
Slope 688
Intercept -700230

IR Calibration Date: 06/21/2013

	3um	9um
0th Coef(*100):	-23841	-21159
1st Coef(*100):	286265	140038
2nd Coef(*100):	2119	1082
H2O adj(mg/l*10k):	533	468

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

EXTERNAL PRINT TEST


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005954
Location = TOXL 8164.13.00 06/09
09/09/2013 16:13

DIAGNOSTIC

Voltage/Current Test	Pass
RAM Test	Pass
EEPROM Checksum Test	Pass
Real Time Clock Test	Pass
DSP Test	Pass
Analytical Stability Test	Pass
Modem Test	Pass
Temp Regulation Test	Pass

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *DIAGNOSTIC CHECK*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005954
Location = TOXL 8164.13.00 06/09
09/09/2013 16:36

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:36
02 Std. Sol.	0.000	16:37
03 Room Air	0.000	16:38
04 Std. Sol.	0.000	16:38
05 Room Air	0.000	16:39
06 Std. Sol.	0.000	16:39
07 Room Air	0.000	16:40

08 Sim Temp = 34.0°C

Simul Ser No = DR5188
Std Sol No = MILLIQ H2O
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

ZERO ALCOHOL TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005954
Location = TOXL 8164.13.00 06/09
09/09/2013 16:41

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:42
02 Std. Sol.	0.018	16:43
03 Room Air	0.000	16:43
04 Std. Sol.	0.019	16:44
05 Room Air	0.000	16:44
06 Std. Sol.	0.019	16:45
07 Room Air	0.000	16:46

08 Sim Temp = 34.0°C

Simul Ser No = DR3455
Std Sol No = 12210
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Low AC*
 0.020 AC Sim. Sol.

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005954
Location = TOXL 8164.13.00 06/09
09/09/2013 16:47

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:48
02 Std. Sol.	0.298	16:49
03 Room Air	0.000	16:49
04 Std. Sol.	0.300	16:50
05 Room Air	0.000	16:51
06 Std. Sol.	0.301	16:51
07 Room Air	0.000	16:52

08 Sim Temp = 34.0°C

Simul Ser No = DR3847
Std Sol No = 13020
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *LINEARITY TEST*
0.300 AC Sim. Sol.

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005954
Location = TOXL 8164.13.00 06/09
09/09/2013 16:54

Test	AC	Time
01 Room Air	0.000	16:55
02 *Subject Test	INT*	16:56
03 Room Air	0.000	16:57

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = 0.10AC+0.05ACE
Dr. Lic. = ND/DIS821456
Lot No = 23411080A1
Cyl No = 32
Expiration Date = 10/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: 0.10AC + 0.05% ACETONE
INTERFERENT CHECK

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005954
Location = TOXL 8164.13.00 06/09
09/09/2013 16:59

Test	AC	Time
01 Room Air	0.000	17:00
02 *Subject Test	INT*	17:00
03 Room Air	0.000	17:01

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = 0.05 ACETONE
Dr. Lic. = ND/DIS821456
Lot No = 23411080A1
Cyl No = 32
Expiration Date = 10/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: 0.05% ACETONE
 INTERFERENT CHECK

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005954
Location = TOXL 8164.13.00 06/09
09/09/2013 17:03

Test	AC	Time
01 Room Air	0.000	17:04
02 *Subject Test	INT*	17:04
03 Room Air	0.000	17:05

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = 0.10 METHANOL
Dr. Lic. = ND/DIS821456
Lot No = 23411080A1
Cyl No = 32
Expiration Date = 10/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: 0.10% METHANOL
 INTERFERENT CHECK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005954
Location = TOXL 8164.13.00 06/09
09/09/2013 17:06

Test	AC	Time
01 Room Air	0.000	17:07
02 *Subject Test	INT*	17:08
03 Room Air	0.000	17:09

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = 0.10 IPA
Dr. Lic. = ND/DIS821456
Lot No = 23411080A1
Cyl No = 32
Expiration Date = 10/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: 0.10% ISOPROPANOL
 INTERFERENT CHECK

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005954
Location = TOXL 8164.13.00 06/09
09/10/2013 09:21

Test	AC	Time
01 Room Air	AMB*	09:21
02 Room Air	0.000	09:22

*Invalid Test
Check Ambient Conditions

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = AMBIENT FAIL
Dr. Lic. = ND/DIS821456
Lot No = 23411080A1
Cyl No = 32
Expiration Date = 10/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *AMBIENT FAIL*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005954
Location = TOXL 8164.13.00 06/09
09/10/2013 10:25

Test	AC	Time
01 Room Air	0.000	10:26
02 Invalid Sample	X.XXX	10:26
03 Room Air	0.000	10:27

*Invalid Test - Mouth Alcohol

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INVALID SAMPLE
Dr. Lic. = ND/DIS821456
Lot No = 23411080A1
Cyl No = 32
Expiration Date = 10/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *INVALID SAMPLE*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005954
Location = TOXL 8164.13.00 06/09
09/10/2013 10:30

Test	AC	Time
01 Room Air	0.000	10:31
02 *Subject Test	0.000*	10:34
03 Room Air	0.000	10:35
No RFI Detected		

*Deficient Sample - Value Printed was
Highest Obtained

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = TEST TIMEOUT
Dr. Lic. = ND/DIS821456
Lot No = 23411080A1
Cyl No = 32
Expiration Date = 10/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: SUBJECT TEST TIMEOUT


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005954
Location = TOXL 8164.13.00 06/09
09/10/2013 10:36

Test	AC	Time
01 Room Air	0.000	10:37
02 *Subject Test	ABT*	10:38
03 Room Air	0.000	10:38

*Sequence Aborted

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = SEQUENCE ABORT
Dr. Lic. = ND/DIS821456
Lot No = 23411080A1
Cyl No = 32
Expiration Date = 10/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

SEQUENCE ABORTED

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005954
Location = TOXL 8164.13.00 06/09
09/10/2013 10:39

Test	AC	Time
01 Diagnostic	OK	10:40
02 Room Air	0.000	10:41
03 Subject Test 1	0.000	10:41
04 Room Air	0.000	10:43
05 Std. Gas	0.000*	10:44
06 Room Air	0.000	10:45

*Cal Check Out of Tolerance

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = CONTROL OUT TOL
Dr. Lic. = ND/DIS821456
Lot No = 23411080A1
Cyl No = 32
Expiration Date = 10/01/2013
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: *CONTROL OUTSIDE TOLERANCE*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005954
Location = TOXL 8164.13.00 06/09
09/10/2013 10:46

Test	AC	Time
01 Diagnostic	OK	10:47
02 Room Air	RFI*	10:47
03 Room Air	0.000	10:48

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 23411080A1
Cyl No = 32
Expiration Date = 10/01/2013
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005954
Location = TOXL 8164.13.00 06/09
09/10/2013 14:30

Test	AC	Time
01 Diagnostic	OK	14:31
02 Room Air	0.000	14:32
03 *Subject Test	REF*	14:32
04 Room Air	0.000	14:32

*Subject Test Refused

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = SUBJECT REFUSE
Dr. Lic. = ND/DIS821456
Lot No = 23411080A1
Cyl No = 32
Expiration Date = 10/01/2013
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: *SUBJECT REFUSED*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005954
Location = TOXL 8164.13.00 06/09
09/10/2013 14:34

Test	AC	Time
01 Diagnostic	OK	14:35
02 Room Air	0.000	14:36
03 Subject Test 1	0.099*	14:36
04 Room Air	0.000	14:42
05 Std. Gas	0.081	14:43
06 Room Air	0.000	14:44
07 Subject Test 2	0.000*	14:45
08 Room Air	0.000	14:46

*Difference Too Great

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = DIFF TOO GREAT
Dr. Lic. = ND/DIS821456
Lot No = 23411080A1
Cyl No = 32
Expiration Date = 10/01/2013
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

DIFFERENCE TOO GREAT

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005954
Location = TOXL 8164.13.00 06/09
09/09/2013 17:13

Test	AC	Time
01 Room Air	0.000	17:14
02 Std. Gas	0.080	17:14
03 Room Air	0.000	17:14
04 Std. Gas	0.081	17:15
05 Room Air	0.000	17:15
06 Std. Gas	0.080	17:16
07 Room Air	0.000	17:16
08 Std. Gas	0.080	17:17
09 Room Air	0.000	17:17
10 Std. Gas	0.081	17:18
11 Room Air	0.000	17:18
12 Std. Gas	0.081	17:18
13 Room Air	0.000	17:19
14 Std. Gas	0.081	17:19
15 Room Air	0.000	17:20
16 Std. Gas	0.080	17:20
17 Room Air	0.000	17:21
18 Std. Gas	0.081	17:21
19 Room Air	0.000	17:22
20 Std. Gas	0.080	17:22
21 Room Air	0.000	17:23
22 Std. Gas	0.081	17:23
23 Room Air	0.000	17:24
24 Std. Gas	0.082	17:24
25 Room Air	0.000	17:25
26 Std. Gas	0.081	17:25
27 Room Air	0.000	17:25
28 Std. Gas	0.080	17:26
29 Room Air	0.000	17:26
30 Std. Gas	0.082	17:27
31 Room Air	0.000	17:27
32 Std. Gas	0.080	17:28
33 Room Air	0.000	17:28
34 Std. Gas	0.081	17:28
35 Room Air	0.000	17:29
36 Std. Gas	0.082	17:29
37 Room Air	0.000	17:30
38 Std. Gas	0.081	17:30
39 Room Air	0.000	17:31
40 Std. Gas	0.080	17:31
41 Room Air	0.000	17:32
42 Cal Check Stats:		
43 Average	0.0808	
44 Std Dev	0.0007	
45 Rel Std Dev	0.89%	

Charles E. Ed
STABILITY TEST

Operator Signature
CHARLES EDER

Remarks: *Charles E. Eder*
STABILITY TEST
Form 106-I8000

Lot #: 23411080A1

Cyl #: 032

Exp. DATE: 10/1/2013

Charles E. Eder
STABILITY TEST