

**INTOXILYZER® 8000 INITIAL INSPECTION**

Serial Number: 80-005957

1. Power On and Warm-up:

- a.  Power On
- b.  5 min. Countdown from Initial Warm-up
- c.  1 min. Countdown to Begin a Test
- d.  Forms Load Complete

2. Setup and Checks:

a. Level 1:

- 1) Passwords:  Field Inspector (Level 2)  Lab Personnel (Level 3)
- 2) Gas Cylinder:  Gas Cylinder Setup (Level 1,S)
- 3) Printer:  Printer Setup (Level 3,S,P)
  - Internal Printer Paper Feed
  - Internal Printer Test (Level 1,P)
  - External Printer Test (Level 1,P)
- 4) Barcode Reader (Level 1,B):  Operator ID  
 ND License

b. Level 2:

- 1)  Continuous Purge (Level 2,A)
- 2)  Diagnostic (Level 2,D)
- 3)  Prelim Data Entry (Level 2,E)
- 4)  Ser Num/Version Info (Level 2,V): Ser Num 80-005957  
Version 8164.13

c. Level 3:

- 1) DVM Monitor (Level 3,D,D): 3: 12898 s: 1 c: 47.00  
9: 13375 s: 1 B: 45 0  
(RFI indicator with keyed radio) 1
- 2) Function (F) Keys:  F5 – Tone  
 F6 – Green LED  
 F7 – Red LED  
 F9 – Fan  
 F10 – Pump  
 F11 – Solenoid *A CEE*
- 3) Atmospheric Monitor (Level 3,D,T): Atmo Sensor, P: 960 CF: 1.05
- 4) Flow Monitor (Level 3,D,F): Flow Sensor, R: 229 F: 0.000 L/s

5) Temperature Monitor (Level 3,D,T): C: 47.0 A: 46.0 I: 0.0  
B: 45.1 V: 0.0

6) Tank Monitor (Level 3,D,G): P (gauge): 510 P (display): 519  
7)  General Setup (Level 3,S,G)  
8)  Configure Start Test (Level 3,S,T)  
9)  Network (Level 3,S,N)

3. Tests:

a. Configure Standard for Wet Bath Simulator (Level 1,S):

1) Zero Alcohol Test (Level 1,C):  
Sim Ser #: DR5188 Lot #: MILLI QH<sub>2</sub>O Exp Date: NA  
Known Value: 0.000 AC  
Results: 0.000 AC 0.000 AC 0.000 AC

2) Low AC Test (Level 1,C):  
Sim Ser #: DR3455 Lot #: 12210 Exp Date: 11/2/14  
Known Value: 0.020 AC  
Results: 0.018 AC 0.019 AC 0.019 AC

3) Linearity Test (Level 1,C):  
Sim Ser #: DR3847 Lot #: 13020 Exp Date: 1-15-15  
Known Value: 0.300 AC  
Results: 0.294 AC 0.296 AC 0.296 AC

b. Configure Dry Gas Standard (Level 1,S):

1) Interferent Detected Tests (Level 1,B):

1. 0.10 AC + 0.05% Acetone CEE 4-9-14  
Sim Ser #: DR3377 Lot #: ICS#1 Exp Date: 09-14  
Known Value: 0.100 AC + 0.05% ACETONE  
 Display reads "Interferent Detect"

2. Acetone  
Sim Ser #: G8919 Lot #: ACE1 Exp Date: 8-27-15  
Known Value: 0.05% ACETONE  
 Display reads "Interferent Detect"

3. Methanol  
Sim Ser #: G10573 Lot #: MET1 Exp Date: 8-27-15  
Known Value: 0.10% METHANOL  
 Display reads "Interferent Detect"

4. Isopropanol  
Sim Ser #: 99998 Lot #: IPA1 Exp Date: 8-27-15  
Known Value: 0.10% ISOPROPANOL  
 Display reads "Interferent Detect"
- 2)  Ambient Fail (Level 1,B): Display and test record are correct  
3)  Invalid Sample (Level 1,B): Display and test record are correct  
4)  Subject Test Timeout (Level 1,B): Display and test record are correct.  
5)  Sequence Aborted (Level 1,B): Display and test record are correct  
6)  Control Outside Tolerance (CMS mode): Display and test record are correct  
7)  RFI Detected Test (CMS mode): Display and test record are correct  
8)  Tank Pressure Below Min (CMS mode): Display reads correct  
9)  Subject Refused (CMS mode): Display and test record are correct  
10)  Difference Too Great (CMS mode): Display and test record are correct  
11)  Stability Test (Level 3,T,S):  
Average 0.080 Std Dev 0.0006 Rel Std Dev 0.76%

Maintenance/Repairs/Notes: \_\_\_\_\_

THIS UNIT IS ACCEPTABLE TO BE USED IN THE FIELD

  
Field Inspector's Signature

11 SEP 13  
Date

  
Reviewed by

11 Sept 13  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer  
North Dakota Model 8000 SN 80-005957  
Location = TOXL 8164.13.00 06/09  
09/10/2013 16:40

\*\*\*\*\* Printer Test \*\*\*\*\*

12 char/line  
abcdefghijkl  
ABCDEFGHIJKL

16 char per line  
abcdefghijklmnop  
ABCDEFGHIJKLMN

21 chars per line  
abcdefghijklmnopqrstu  
ABCDEFGHIJKLMNQRSTU

24 characters per line  
abcdefghijklmnopqrstuwx  
ABCDEFGHIJKLMNQRSTUWX

32 characters per line  
abcdefghijklmnopqrstuwx123456  
ABCDEFGHIJKLMNQRSTUWXYZ123456

42 characters per line  
abcdefghijklmnopqrstuwx1234567890abcdef  
ABCDEFGHIJKLMNQRSTUWXYZ1234567890ABCDEF

Current Instrument Setup

Data Entry Mode: Enabled  
Start Test Sequence: DABACABA  
Display Prelim Rslt? Yes  
Display Third Digit? Yes  
Inhib Printer(Y/N)? No  
Display Volume? No  
Disable On Menfull? No  
# of Print Copies? 1  
Select Std (D/W/I)? Dry  
Standard Value? 0.080  
Standard Lot #? 23411080A1  
Standard Cyl #? 32  
Standard Expiration? 10/01/2013  
Oper No? 666666

Flow Cal. Date: 06/21/2013  
Slope 647  
Intercept -613317

IR Calibration Date: 06/21/2013

	3um	9um
0th Coef(*100):	-20907	-21174
1st Coef(*100):	287637	137056
2nd Coef(*100):	1788	1065
H2O adj(ng/10k):	508	418

\*\*\*\*\* Test End \*\*\*\*\*



Operator Signature  
CHARLES EDER

Remarks: INTERNAL  
PRINT TEST

Form 106-18000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005957  
Location = TOXL                    8164.13.00 06/09  
09/10/2013                            17:01

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^\*()\_+?

Current Instrument Setup


Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        No  
# of Print Copies?         1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            23411080A1  
Standard Cyl #?            32  
Standard Expiration?       10/01/2013  
Oper No?                    666666

Flow Cal. Date:            06/21/2013  
Slope                      647  
Intercept                  -613317

IR Calibration Date:        06/21/2013  
                                    3um                      9um

-----  
0th Coef(\*100):            -20907                    -21174  
1st Coef(\*100):            287637                    137056  
2nd Coef(\*100):            1788                      1065  
H2O adj(mg/l\*10k):        508                        418

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
CHARLES EDER

Remarks:    EXTERNAL PRINT  
                                    TEST


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005957  
Location = TOXL      8164.13.00 06/09  
09/10/2013      17:05

DIAGNOSTIC

Voltage/Current Test	Pass
RAM Test	Pass
EEPROM Checksum Test	Pass
Real Time Clock Test	Pass
DSP Test	Pass
Analytical Stability Test	Pass
Modem Test	Pass
Temp Regulation Test	Pass

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

DIAGNOSTIC CHECK

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005957  
Location = TOXL                8164.13.00 06/09  
09/11/2013                            10:45

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:46
02 Std. Sol.	0.000	10:47
03 Room Air	0.000	10:47
04 Std. Sol.	0.000	10:48
05 Room Air	0.000	10:48
06 Std. Sol.	0.000	10:49
07 Room Air	0.000	10:50

08 Sim Temp = 34.0°C

Simul Ser No = DR5188  
Std Sol No = MILLIQ H2O  
County = 08                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

ZERO ALCOHOL TEST  
0.000 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005957  
Location = TOXL                    8164.13.00 06/09  
09/11/2013                            11:50

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:51
02 Std. Sol.	0.018	11:52
03 Room Air	0.000	11:52
04 Std. Sol.	0.019	11:53
05 Room Air	0.000	11:53
06 Std. Sol.	0.019	11:54
07 Room Air	0.000	11:55

08 Sim Temp = 34.0°C

Simul Ser No = DR3455

Std Sol No = 12210

County = 08

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:            0.020 AC Sim. Sol.  
                          Low AC CHECK



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005957  
Location = TOXL      8164.13.00 06/09  
09/11/2013      12:07

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:08
02 Std. Sol.	0.294	12:08
03 Room Air	0.000	12:09
04 Std. Sol.	0.296	12:10
05 Room Air	0.000	12:10
06 Std. Sol.	0.296	12:11
07 Room Air	0.000	12:11

08 Sim Temp = 34.0°C

Simul Ser No = DR3847

Std Sol No = 13020

County = 08

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: 0.300 AC Sim. Sol.

LINEARITY TEST

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005957  
Location = TOXL                    8164.13.00 06/09  
09/11/2013                            13:11

Test	AC	Time
01 Room Air	0.000	13:12
02 *Subject Test	INT*	13:12
03 Room Air	0.000	13:13

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = 0.10AC+0.05ACE  
Dr. Lic. = ND/DIS821456  
Lot No = 23411080A1  
Cyl No = 32  
Expiration Date = 10/01/2013  
County = 08                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: 0.100AC + 0.05% ACETONE

INTERFERENT CHECK

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005957  
Location = TOXL                    8164.13.00 06/09  
09/11/2013                            13:18

Test	AC	Time
01 Room Air	0.000	13:19
02 *Subject Test	INT*	13:20
03 Room Air	0.000	13:20

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = 0.05 ACETONE  
Dr. Lic. = ND/DIS821456  
Lot No = 23411080A1  
Cyl No = 32  
Expiration Date = 10/01/2013  
County = 08                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: 0.05% ACETONE  
INTERFERENT CHECK

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005957  
Location = TOXL                    8164.13.00 06/09  
09/11/2013                            13:22

Test	AC	Time
01 Room Air	0.000	13:23
02 *Subject Test	INT*	13:24
03 Room Air	0.000	13:25

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = 0.10 METHANOL  
Dr. Lic. = ND/DIS821456  
Lot No = 23411080A1  
Cyl No = 32  
Expiration Date = 10/01/2013  
County = 08                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:    0.10% METHANOL  
                 INTERFERENT CHECK

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005957  
Location = TOXL      8164.13.00 06/09  
09/11/2013      13:26

Test	AC	Time
01 Room Air	0.000	13:28
02 *Subject Test	INT*	13:28
03 Room Air	0.000	13:29

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = 0.10 IPA  
Dr. Lic. = ND/DIS821456  
Lot No = 23411080A1  
Cyl No = 32  
Expiration Date = 10/01/2013  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:      0.10% ISOPROPANOL  
INTERFERENT CHECK


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer                      Alcohol Analyzer  
North Dakota Model 8000                      SN 80-005957  
Location = TOXL                                      8164.13.00 06/09  
09/11/2013    13:39

Test	AC	Time
01 Room Air	AMB*	13:40
02 Room Air	0.000	13:40

\*Invalid Test  
  Check Ambient Conditions

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                                      Weight = 150  
Test = DUI                                      Cit = AMBIENT FAIL  
Dr. Lic. = ND/DIS821456  
Lot No = 23411080A1  
Cyl No = 32  
Expiration Date = 10/01/2013  
County = 08                                      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *AMBIENT FAIL*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005957  
Location = TOXL                8164.13.00 06/09  
09/11/2013                        13:43

Test	AC	Time
01 Room Air	0.000	13:44
02 Invalid Sample	X.XXX	13:45
03 Room Air	0.000	13:45

\*Invalid Test - Mouth Alcohol

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = INVALID SAMPLE  
Dr. Lic. = ND/DIS821456  
Lot No = 23411080A1  
Cyl No = 32  
Expiration Date = 10/01/2013  
County = 08                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *INVALID SAMPLE*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005957  
Location = TOXL                    8164.13.00 06/09  
09/11/2013                            13:46

Test	AC	Time
01 Room Air	0.000	13:47
02 *Subject Test	0.000*	13:50
03 Room Air	0.000	13:51

No RFI Detected

\*Deficient Sample - Value Printed was  
Highest Obtained

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = SUBJECT TIMEOUT  
Dr. Lic. = ND/DIS821456  
Lot No = 23411080A1  
Cyl No = 32  
Expiration Date = 10/01/2013  
County = 08                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*SUBJECT TEST TIMEOUT*




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005957  
Location = TOXL                    8164.13.00 06/09  
09/11/2013                            13:54

Test	AC	Time
01 Room Air	0.000	13:55
02 *Subject Test	ABT*	13:55
03 Room Air	0.000	13:56

\*Sequence Aborted

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = SEQUENCE ABORT  
Dr. Lic. = ND/DIS821456  
Lot No = 23411080A1  
Cyl No = 32  
Expiration Date = 10/01/2013  
County = 08                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: SEQUENCE ABORTED

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005957  
Location = TOXL                    8164.13.00 06/09  
09/11/2013                            13:57

Test	AC	Time
01 Diagnostic	OK	13:58
02 Room Air	0.000	13:59
03 Subject Test 1	0.101	13:59
04 Room Air	0.000	14:02
05 Std. Gas	0.000*	14:03
06 Room Air	0.000	14:03

\*Cal Check Out of Tolerance

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = OUT TOLERANCE  
Dr. Lic. = ND/DIS821456  
Lot No = 23411080A1  
Cyl No = 32  
Expiration Date = 10/01/2013  
County = 08                            Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks: *Central outside Tolerance*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005957  
Location = TOXL                    8164.13.00 06/09  
09/11/2013                            14:05

Test	AC	Time
01 Diagnostic	OK	14:06
02 Room Air	RFI*	14:06
03 Room Air	0.000	14:06

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = RFI CHECK  
Dr. Lic. = ND/DIS821456  
Lot No = 23411080A1  
Cyl No = 32  
Expiration Date = 10/01/2013  
County = 08                            Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks:

RFI CHECK

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005957  
Location = TOXL                    8164.13.00 06/09  
09/11/2013                            14:09

Test	AC	Time
01 Diagnostic	OK	14:10
02 Room Air	0.000	14:10
03 *Subject Test	REF*	14:11
04 Room Air	0.000	14:11

\*Subject Test Refused

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = SUB REFUSED  
Dr. Lic. = ND/DIS821456  
Lot No = 23411080A1  
Cyl No = 32  
Expiration Date = 10/01/2013  
County = 08                            Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks:

*SUBJECT REFUSED*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005957  
Location = TOXL                    8164.13.00 06/09  
09/11/2013                            14:13

Test	AC	Time
01 Diagnostic	OK	14:14
02 Room Air	0.000	14:14
03 Subject Test 1	0.100*	14:15
04 Room Air	0.000	14:17
05 Std. Gas	0.080	14:18
06 Room Air	0.000	14:20
07 Subject Test 2	0.154*	14:20
08 Room Air	0.000	14:21

\*Difference Too Great

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = DIFF TOO GREAT  
Dr. Lic. = ND/DIS821456  
Lot No = 23411080A1  
Cyl No = 32  
Expiration Date = 10/01/2013  
County = 08                            Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks: *DIFFERENCE TOO GREAT*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005957  
Location = TOXL                8164.13.00 06/09  
09/11/2013                      14:23

Test	AC	Time
01 Room Air	0.000	14:24
02 Std. Gas	0.079	14:24
03 Room Air	0.000	14:25
04 Std. Gas	0.080	14:25
05 Room Air	0.000	14:26
06 Std. Gas	0.080	14:26
07 Room Air	0.000	14:27
08 Std. Gas	0.080	14:27
09 Room Air	0.000	14:28
10 Std. Gas	0.079	14:28
11 Room Air	0.000	14:28
12 Std. Gas	0.080	14:29
13 Room Air	0.000	14:29
14 Std. Gas	0.081	14:30
15 Room Air	0.000	14:30
16 Std. Gas	0.080	14:31
17 Room Air	0.000	14:31
18 Std. Gas	0.080	14:31
19 Room Air	0.000	14:32
20 Std. Gas	0.080	14:32
21 Room Air	0.000	14:33
22 Std. Gas	0.079	14:33
23 Room Air	0.000	14:34
24 Std. Gas	0.080	14:34
25 Room Air	0.000	14:35
26 Std. Gas	0.080	14:35
27 Room Air	0.000	14:35
28 Std. Gas	0.080	14:36
29 Room Air	0.000	14:36
30 Std. Gas	0.079	14:37
31 Room Air	0.000	14:37
32 Std. Gas	0.080	14:38
33 Room Air	0.000	14:38
34 Std. Gas	0.080	14:38
35 Room Air	0.000	14:39
36 Std. Gas	0.081	14:39
37 Room Air	0.000	14:40
38 Std. Gas	0.081	14:40
39 Room Air	0.000	14:41
40 Std. Gas	0.080	14:41
41 Room Air	0.000	14:42
42 Cal Check Stats:		
43 Average	0.0800	
44 Std Dev	0.0006	
45 Rel Std Dev	0.76%	

STABILITY TEST

*Charles E. Edin*

Operator Signature  
CHARLES EDER

Remarks:

STABILITY TEST

Form 106-I8000

0.080 AC E.B.S

Lot#: 234/11080A1

Cyl #: 032

Exp. DATE: 10/1/2013

