



ETHANOL BREATH STANDARD CYLINDER REPORT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Mark W. McDonald</u>		
Location: <u>Valley City Police Dept.</u>		
Intoxilyzer® Serial Number: <u>80-004201</u>		
Gas Lot Number: <u>65935K</u>	Gas Cylinder Number: <u>24</u>	Expiration Date: <u>09/03/11</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%) Test 1: 0. <u>074</u> % Test 2: 0. <u>074</u> % Test 3: 0. <u>077</u> %		
(ATTACH TEST RECORD)		

[Signature]
 Chemical Test Operator Signature

09/16/2010
09/16/2010
 Date

Deb Kashur
 Reviewed By

15 Sept 2010
 Date

80-004201
 P. 9/16/10
 km

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004201
Location = VALL 8164.13.00 06/09
09/12/2010 13:29

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:29
02 Std. Gas	0.078	13:30
03 Room Air	0.000	13:30
04 Std. Gas	0.078	13:30
05 Room Air	0.000	13:31
06 Std. Gas	0.077	13:31
07 Room Air	0.000	13:32


Lot No = 659358

Cyl No = 24

Exp Date = 09/03/2011

County = 02

Oper No. = 119502



Operator Signature
MARK W. MCDONALD

Remarks:

Form 106-18000

80-004201
p. 92 of 92
FM