

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004201  
Location = VALL      8164.13.00 06/09  
09/12/2010      13:29

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:29
02 Std. Gas	0.078	13:30
03 Room Air	0.000	13:30
04 Std. Gas	0.078	13:30
05 Room Air	0.000	13:31
06 Std. Gas	0.077	13:31
07 Room Air	0.000	13:32

Lot No = 659358  
Cyl No = 24  
Exp Date = 09/03/2011  
County = 02      Oper No. = 119502



Operator Signature  
MARK W. MCDONALD

Remarks:

Form 106-I8000

80-004201  
P. 95/95  
RHL