

Certificate of Calibration

Intoxilyzer® 8000

*This is to certify that the calibration of **Intoxilyzer**® serial number 80-002669, manufactured by CMI Inc., a subsidiary of MPD, Inc. of Owensboro, Kentucky, was tested and found to be in compliance with the National Highway Traffic Safety Administration Standard for Devices to Measure Breath Alcohol (F.R., No. 179 48705-48710 Sept. 17, 1993) and with International Recommendation OIML R126 Edition 1998 (E). Calibration solutions are traceable to NIST standard material 1828.*

Date Sept. 12, 2011

Signed Del Rohrer

Technician

EMI INC.

316 East 9th Street
Owensboro, KY 42303
USA

Part No. 650519 CMI 7/9/09



SERVICE WORK ORDER

DATE	7-13-11
RCVD VIA	FENX

316 E. 9th St., Owensboro KY 42303
 Phone 866-835-0690
 Fax: 270-685-6268

321032

BILL TO:		
ATTN:		
CITY	STATE	ZIP
CUSTOMER NO. 585013		
PHONE		

SHIP TO: ND Crime Laboratory		
ATTN: Deb Kachur		
2641 E. Main Ave.		
CITY	STATE	ZIP
Bismarck	ND	58501
PHONE 701-328-6159		
FAX 701-328-6185		

MODEL # 8000	SERIAL # 50-002669	BILL CODE 31
--------------	--------------------	--------------

EXTRA PARTS RCVD BT, 300 box, battery
DESCRIPTION OF PROBLEM Lightning - NO DISPLAY

Infrared / Fuel Cell PRELIMINARY TEST INFORMATION	BATTERY-EXT OK V; INT 3.00V
NO DISPLAY / orange light only - YES	
VERIFIED PROBLEM?	
WORK PERFORMED: EXT BTX CASE DAMAGED - Replaced parts listed (R) I O# 04296	

PARTS USED		
QTY	PART NO.	DESCRIPTION
1	441287	BTX CASE
1	220284	PROM-FLASH
1	013010	20 BTX
1	110211	.47cap

Miscellaneous PRELIMINARY TEST INFORMATION
WORK PERFORMED: PDS

REPAIR DU	CAL. ADJ. DR	CALIBRATION DR	FINAL DR
1/2 HOURS 8	HOURS 1	HOURS .5	HOURS .5
DATE 8/11/11	DATE 9/9/11	DATE 9/12/11	DATE 9/12/11

SHIPPING STICKER HERE
SPECIAL SHIPPING INSTRUCTIONS