

INTOXILYZER® 8000 INITIAL INSPECTION

Serial Number: 80-005956

1. Power On and Warm-up:

- a. Power On
- b. 5 min. Countdown from Initial Warm-up
- c. 1 min. Countdown to Begin a Test
- d. Forms Load Complete

2. Setup and Checks:

a. Level 1:

- 1) Passwords: Field Inspector (Level 2) Lab Personnel (Level 3)
- 2) Gas Cylinder: Gas Cylinder Setup (Level 1,S)
- 3) Printer: Printer Setup (Level 3,S,P)
 - Internal Printer Paper Feed
 - Internal Printer Test (Level 1,P)
 - External Printer Test (Level 1,P)
- 4) Barcode Reader (Level 1,B): Operator ID
 ND License

b. Level 2:

- 1) Continuous Purge (Level 2,A)
- 2) Diagnostic (Level 2,D)
- 3) Prelim Data Entry (Level 2,E)
- 4) Ser Num/Version Info (Level 2,V): Ser Num 80-005956
Version 8164.13

c. Level 3:

- 1) DVM Monitor (Level 3,D,D): 3: 12746 s: 1 C: 47.00
9: 13277 s: 1 B: 44 0
(RFI indicator with keyed radio) 1
- 2) Function (F) Keys: F5 – Tone
 F6 – Green LED
 F7 – Red LED
 F9 – Fan
 F10 – Pump
 F11 – Solenoid
- 3) Atmospheric Monitor (Level 3,D,A): Atmo Sensor, P: 966 CF: 1.04
- 4) Flow Monitor (Level 3,D,F): Flow Sensor, R: 226 F: 0.000 L/s

5) Temperature Monitor (Level 3,D,T): C: 47.0 A: 48.0 I: 0.0
B: 45.1 V: 0.0

- 6) Tank Monitor (Level 3,D,G): P (gauge): 2200 P (display): 209
7) General Setup (Level 3,S,G)
8) Configure Start Test (Level 3,S,T)
9) Network (Level 3,S,N)

3. Tests:

a. Configure Standard for Wet Bath Simulator (Level 1,S):

- 1) Zero Alcohol Test (Level 1,C):
Sim Ser #: G10329 Lot #: N/A Exp Date: N/A
Known Value: 0.000 AC
Results: 0.000 AC 0.000 AC 0.000 AC
- 2) Low AC Test (Level 1,C):
Sim Ser #: DR3803 Lot #: 13070 Exp Date: 03/06/2015
Known Value: 0.020 AC
Results: 0.018 AC 0.019 AC 0.019 AC
- 3) Linearity Test (Level 1,C):
Sim Ser #: DR5143 Lot #: 13080 Exp Date: 03/19/2015
Known Value: 0.300 AC
Results: 0.301 AC 0.302 AC 0.301 AC

b. Configure Dry Gas Standard (Level 1,S):

- 1) Interferent Detected Tests (Level 1,B):
- 0.10 AC + 0.05% Acetone
Sim Ser #: DR3846 Lot #: ICS2 Exp Date: 27 Aug 2015
Known Value: 0.10 AC + 0.05% acetone
 Display reads "Interferent Detect"
 - Acetone
Sim Ser #: DR 3375 Lot #: ACE1 Exp Date: 27 Aug 2015
Known Value: 0.05% acetone
 Display reads "Interferent Detect"
 - Methanol
Sim Ser #: DR 3454 Lot #: MET1 Exp Date: 27 Aug 2015
Known Value: 0.10% Methanol
 Display reads "Interferent Detect"

4. Isopropanol
Sim Ser #: DR5131 Lot #: IPA1 Exp Date: 27 Aug 2015
Known Value: 0.10% Isopropanol
 Display reads "Interferent Detect"
- 2) Ambient Fail (Level 1,B): Display and test record are correct
3) Invalid Sample (Level 1,B): Display and test record are correct
4) Subject Test Timeout (Level 1,B): Display and test record are correct.
5) Sequence Aborted (Level 1,B): Display and test record are correct
6) Control Outside Tolerance (CMS mode): Display and test record are correct
7) RFI Detected Test (CMS mode): Display and test record are correct
8) Tank Pressure Below Min (CMS mode): Display reads correct
9) Subject Refused (CMS mode): Display and test record are correct
10) Difference Too Great (CMS mode): Display and test record are correct
11) Stability Test (Level 3,T,S):
Average 0.0820 Std Dev 0.0006 Rel Std Dev 0.80%

Maintenance/Repairs/Notes: Initial inspection, recalibrated gas regulator.

This intox is ready to go out in the field.

Deb Kashner
Field Inspector's Signature

12 Sept 2013
Date

Charles E. Eden
Reviewed by

12 SEP 13
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005956
Location = 10XL 8164.13.00 06/09
09/12/2013 08:35

***** Printer Test *****

12 char/line
abcdefghijkl
ABCDEFGHIJKL

16 char per line
abcdefghijklmnop
ABCDEFGHIJKLMN

21 chars per line
abcdefghijklmnopqrstu
ABCDEFGHIJKLMNOPQR

24 characters per line
abcdefghijklmnopqrstu
ABCDEFGHIJKLMNOPQR

32 characters per line
abcdefghijklmnopqrstu
ABCDEFGHIJKLMNOPQR

42 characters per line
abcdefghijklmnopqrstu
ABCDEFGHIJKLMNOPQR

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Menfull? No
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 1234567890
Standard Cyl #? 12W3456
Standard Expiration? 11/11/2015
Oper No? 777777

Flow Cal. Date: 06/21/2013
Slope 706
Intercept -720901

IR Calibration Date: 06/21/2013
3um 9um

0th Coef(*100): -17737 -18829
1st Coef(*100): 287589 138428
2nd Coef(*100): 2236 1280
H2O adj(mg/l*10K): 645 488

***** Test End *****

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Print Test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005956
Location = TOXL 8164.13.00 06/09
09/12/2013 08:46

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMNopqrstuvwxyz!@#%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMNopqrstuvwxyz!@#%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? No
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 23411080A1
Standard Cyl #? 40
Standard Expiration? 10/01/2013
Oper No? 777777

Flow Cal. Date: 06/21/2013
Slope 706
Intercept -720901

IR Calibration Date: 06/21/2013

	3um	9um
0th Coef(*100):	-17737	-18829
1st Coef(*100):	287589	138428
2nd Coef(*100):	2236	1280
H2O adj(mg/l*10k):	645	488

***** Printer Test End *****

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Print test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005956
Location = TOXL 8164.13.00 06/09
09/12/2013 08:47

DIAGNOSTIC

Voltage/Current Test	Pass
RAM Test	Pass
EEPROM Checksum Test	Pass
Real Time Clock Test	Pass
DSP Test	Pass
Analytical Stability Test	Pass
Modem Test	Pass
Temp Regulation Test	Pass

Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Diagnostic test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005956
Location = TOXL 8164.13.00 06/09
09/12/2013 08:55

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:56
02 Std. Sol.	0.000	08:56
03 Room Air	0.000	08:57
04 Std. Sol.	0.000	08:58
05 Room Air	0.000	08:58
06 Std. Sol.	0.000	08:59
07 Room Air	0.000	08:59

08 Sim Temp = 34.0°C

Simul Ser No = G10329
Std Sol No = MILLI Q
County = 08

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

zero alcohol Milli-Q H₂O

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005956
Location = TOXL 8164.13.00 06/09
09/12/2013 09:37

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:38
02 Std. Sol.	0.018	09:39
03 Room Air	0.000	09:39
04 Std. Sol.	0.019	09:40
05 Room Air	0.000	09:40
06 Std. Sol.	0.019	09:41
07 Room Air	0.000	09:42

08 Sim Temp = 34.0°C

Simul Ser No = DR3803

Std Sol No = 13070

County = 08

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Low AC 0.020AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005956
Location = TOXL 8164.13.00 06/09
09/12/2013 09:46

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:46
02 Std. Sol.	0.301	09:49
03 Room Air	0.000	09:49
04 Std. Sol.	0.302	09:50
05 Room Air	0.000	09:51
06 Std. Sol.	0.301	09:51
07 Room Air	0.000	09:52

08 Sim Temp = 34.0°C

Simul Ser No = DR5143
Std Sol No = 13080
County = 08 Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:

Linearity test 0.300 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005956
Location = TOXL 8164.13.00 06/09
09/12/2013 09:54

Test	AC	Time
01 Room Air	0.000	09:55
02 *Subject Test	INT*	09:55
03 Room Air	0.000	09:56

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 23411080A1
Cyl No = 40
Expiration Date = 10/01/2013
County = 08 Oper No. = 777777

Deb Kasher
Operator Signature
DEB KASHUR

Remarks: *Interferent check*
0.10 AC + 0.05% acetone

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005956
Location = TOXL 8164.13.00 06/09
09/12/2013 09:57

Test	AC	Time
01 Room Air	0.000	09:58
02 *Subject Test	INT*	09:59
03 Room Air	0.000	09:59

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 23411080A1
Cyl No = 40
Expiration Date = 10/01/2013
County = 08 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

*Interferent check
0.05% acetone*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005956
Location = TOXL 8164.13.00 06/09
09/12/2013 10:01

Test	AC	Time
01 Room Air	0.000	10:02
02 *Subject Test	INT*	10:02
03 Room Air	0.000	10:03

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 23411080A1
Cyl No = 40
Expiration Date = 10/01/2013
County = 08 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks: *Interferent check*
0.10% Methanol

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005956
Location = TOXL 8164.13.00 06/09
09/12/2013 10:03

Test	AC	Time
01 Room Air	0.000	10:04
02 *Subject Test	INT*	10:04
03 Room Air	0.000	10:05

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 23411080A1
Cyl No = 40
Expiration Date = 10/01/2013
County = 08 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

*Interferent check
0.10% Isopropanol*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005956
Location = TOXL 8164.13.00 06/09
09/12/2013 10:05

Test	AC	Time
01 Room Air	AMB*	10:06
02 Room Air	0.000	10:06

*Invalid Test
Check Ambient Conditions

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 23411080A1
Cyl No = 40
Expiration Date = 10/01/2013
County = 08 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Ambient fail test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005956
Location = TOXL 8164.13.00 06/09
09/12/2013 10:08

Test	AC	Time
01 Room Air	0.000	10:08
02 Invalid Sample	X.XXX	10:09
03 Room Air	0.000	10:10

*Invalid Test - Mouth Alcohol

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 23411080A1
Cyl No = 40
Expiration Date = 10/01/2013
County = 08 Oper No. = 777777



Operator Signature
DEB KASHUR

Remarks:

Invalid sample test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005956
Location = TOXL 8164.13.00 06/09
09/12/2013 10:14

Test	AC	Time
01 Room Air	0.000	10:14
02 *Subject Test	0.000*	10:18
03 Room Air	0.000	10:18

No RFI Detected

*Deficient Sample - Value Printed was
Highest Obtained

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 23411080A1
Cyl No = 40
Expiration Date = 10/01/2013
County = 08 Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Subject test timeout

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005956
Location = TOXL 8164.13.00 06/09
09/12/2013 10:25

Test	AC	Time
01 Room Air	0.000	10:25
02 *Subject Test	ABT*	10:25
03 Room Air	0.000	10:26

*Sequence Aborted

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 23411080A1
Cyl No = 40
Expiration Date = 10/01/2013
County = 08 Oper No. = 777777



Operator Signature
DEB KASHUR

Remarks:

Sequence aborted test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005956
Location = TOXL 8164.13.00 06/09
09/12/2013 10:26

Test	AC	Time
01 Diagnostic	OK	10:27
02 Room Air	0.000	10:28
03 Subject Test 1	0.000	10:28
04 Room Air	0.000	10:30
05 Std. Gas	0.000*	10:31
06 Room Air	0.000	10:32

*Cal Check Out of Tolerance

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 23411080A1
Cyl No = 40
Expiration Date = 10/01/2013
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Control outside tolerance

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005956
Location = TOXL 8164.13.00 06/09
09/12/2013 14:26

Test	AC	Time
01 Diagnostic	OK	14:27
02 Room Air	RFI*	14:27
03 Room Air	0.000	14:27

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 23411080A1
Cyl No = 40
Expiration Date = 10/01/2013
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

RFI test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005956
Location = TOXL 8164.13.00 06/09
09/12/2013 14:31

Test	AC	Time
01 Diagnostic	OK	14:32
02 Room Air	0.000	14:32
03 *Subject Test	REF*	14:33
04 Room Air	0.000	14:33

*Subject Test Refused

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 23411080A1
Cyl No = 40
Expiration Date = 10/01/2013
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Subject Refused test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005956
Location = TOXL 8164.13.00 06/09
09/12/2013 14:34

Test	AC	Time
01 Diagnostic	OK	14:35
02 Room Air	0.000	14:35
03 Subject Test 1	0.000*	14:36
04 Room Air	0.000	14:38
05 Std. Gas	0.082	14:39
06 Room Air	0.000	14:40
07 Subject Test 2	0.052*	14:41
08 Room Air	0.000	14:42

*Difference Too Great

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 23411080A1
Cyl No = 40
Expiration Date = 10/01/2013
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
DEB KASHUR

Remarks:



Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005956
Location = TOXL 8164.13.00 06/09
09/12/2013 14:45

Test	AC	Time
01 Room Air	0.000	14:45
02 Std. Gas	0.082	14:46
03 Room Air	0.000	14:46
04 Std. Gas	0.082	14:47
05 Room Air	0.000	14:47
06 Std. Gas	0.081	14:47
07 Room Air	0.000	14:48
08 Std. Gas	0.081	14:48
09 Room Air	0.000	14:49
10 Std. Gas	0.082	14:49
11 Room Air	0.000	14:50
12 Std. Gas	0.081	14:50
13 Room Air	0.000	14:50
14 Std. Gas	0.081	14:51
15 Room Air	0.000	14:51
16 Std. Gas	0.082	14:52
17 Room Air	0.000	14:52
18 Std. Gas	0.082	14:53
19 Room Air	0.000	14:53
20 Std. Gas	0.082	14:53
21 Room Air	0.000	14:54
22 Std. Gas	0.083	14:54
23 Room Air	0.000	14:55
24 Std. Gas	0.082	14:55
25 Room Air	0.000	14:56
26 Std. Gas	0.082	14:56
27 Room Air	0.000	14:57
28 Std. Gas	0.083	14:57
29 Room Air	0.000	14:57
30 Std. Gas	0.083	14:58
31 Room Air	0.000	14:58
32 Std. Gas	0.082	14:59
33 Room Air	0.000	14:59
34 Std. Gas	0.082	15:00
35 Room Air	0.000	15:00
36 Std. Gas	0.082	15:00
37 Room Air	0.000	15:01
38 Std. Gas	0.083	15:01
39 Room Air	0.000	15:02
40 Std. Gas	0.082	15:02
41 Room Air	0.000	15:03
42 Cal Check Stats:		
43 Average	0.0820	
44 Std Dev	0.0006	
45 Rel Std Dev	0.80%	

*Stability test
Deb Kashner
p. 1/2*

Operator Signature
DEB KASHUR

Remarks:

Lot #: 23411080A1

Cyl. #: 40

Exp. date: 10/01/2013

Form 106-I8000

0.080 AC

Stability test
Deb Kashur
p. 2/2