



ETHANOL BREATH STANDARD CYLINDER REPORT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Richard D. Johnson</u>		
Location: <u>McLean L.E.C. Washburn ND</u>		
Intoxilyzer® Serial Number: <u>80-004196</u>		
Gas Lot Number: <u>659357</u>	Gas Cylinder Number: <u>47</u>	Expiration Date: <u>09/03/11</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%) Test 1: <u>0.081</u> % Test 2: <u>0.080</u> % Test 3: <u>0.081</u> %		

(ATTACH TEST RECORD)

Richard D. Johnson
 Chemical Test Operator Signature

9-13-10
 Date

Deb Kashur
 Reviewed By

15 Sept 2010
 Date

80-004196
 P. Lot # 93 of 94
 sm

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004196
Location = WASH 8164.13.00 06/09
09/13/2010 14:34

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:35
02 Std. Gas	0.081	14:36
03 Room Air	0.000	14:36
04 Std. Gas	0.080	14:37
05 Room Air	0.000	14:37
06 Std. Gas	0.081	14:38
07 Room Air	0.000	14:38

Lot No = 659357
Cyl No = 47
Exp Date = 09/03/2011
County = 28 Oper No. = 118707



Operator Signature
RICHARD D. JOHNSON

Remarks:

Form 106-I8000

80-004196
P. 2 of 2 9/13/10
JCM