



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

| | | |
|---|--|----------------------------------|
| Chemical Test Operator Name: <u>Krabbenhoft, Adam R.</u> | | |
| Location: <u>LaMoure Co. Courthouse</u> | | |
| Intoxilyzer® Serial Number: <u>80-004942</u> | | |
| Gas Lot Number: <u>709352</u> | Gas Cylinder Number: <u>14</u> | Expiration Date: <u>12-23-11</u> |
| <input type="checkbox"/> New Cylinder Installed | <input checked="" type="checkbox"/> Cylinder Not Changed | |
| Instrument Test Results (Report to 3 Digits; Example, 0.081%) | | |
| Test 1: <u>0.078</u> % | | |
| Test 2: <u>0.079</u> % | | |
| Test 3: <u>0.079</u> % | | |
| (ATTACH TEST RECORD) | | |

[Signature]
Chemical Test Operator Signature

9-13-10
Date

Deb Kashur
Reviewed By

16 Sept 2010
Date


Intoxilyzer Test Record and Checklist
NDODJG Crime Lab. Div., Bismarck, ND 58501

CVI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = LEMO 8164.13.00 06/09
09/13/2010 13:27

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 13:28 |
| 02 Std. Gas | 0.078 | 13:28 |
| 03 Room Air | 0.000 | 13:29 |
| 04 Std. Gas | 0.079 | 13:29 |
| 05 Room Air | 0.000 | 13:30 |
| 06 Std. Gas | 0.079 | 13:30 |
| 07 Room Air | 0.000 | 13:30 |

Lot No = 709352
Cyl No = 14
Exp Date = 12/23/2011
County = 23
Oper No. = 021828



Operator Signature
ADAM R. KRABBEHOFT

Remarks: ACA Test
Printer unplugged

Form 106-18000

80-004942
P. 50 of 50
Jan