

INTOXILYZER® 8000 INITIAL INSPECTION

Serial Number: 80-005941

1. Power On and Warm-up:
 - a. Power On
 - b. 5 min. Countdown from Initial Warm-up
 - c. 1 min. Countdown to Begin a Test
 - d. Forms Load Complete

2. Setup and Checks:
 - a. Level 1:
 - 1) Passwords: Field Inspector (Level 2) Lab Personnel (Level 3)
 - 2) Gas Cylinder: Gas Cylinder Setup (Level 1,S)
 - 3) Printer: Printer Setup (Level 3,S,P)
 - Internal Printer Paper Feed
 - Internal Printer Test (Level 1,P)
 - External Printer Test (Level 1,P)
 - 4) Barcode Reader (Level 1,B): Operator ID
 ND License

 - b. Level 2:
 - 1) Continuous Purge (Level 2,A)
 - 2) Diagnostic (Level 2,D)
 - 3) Prelim Data Entry (Level 2,E)
 - 4) Ser Num/Version Info (Level 2,V): Ser Num 80-005941
Version 8164.13

 - c. Level 3:
 - 1) DVM Monitor (Level 3,D,D): 3: 12604 s: 0 C: 47.00
g: 13617 s: 1 B: 44 0
(RFI indicator with keyed radio) 1

 - 2) Function (F) Keys: F5 – Tone
 F6 – Green LED
 F7 – Red LED
 F9 – Fan
 F10 – Pump
 F11 – Solenoid

 - 3) Atmospheric Monitor (Level 3,D,A): Atmo Sensor, P: 967 CF: 1.04
 - 4) Flow Monitor (Level 3,D,F): Flow Sensor, R: 207 F: 0.000 L/s

5) Temperature Monitor (Level 3,D,T): C: 47.0 A: 48.0 I: 0.0
B: 44.9 V: 0.0

6) Tank Monitor (Level 3,D,G): P (gauge): 100 P (display): 103
7) General Setup (Level 3,S,G)
8) Configure Start Test (Level 3,S,T)
9) Network (Level 3,S,N)

3. Tests:

a. Configure Standard for Wet Bath Simulator (Level 1,S):

1) Zero Alcohol Test (Level 1,C):

Sim Ser #: G-10329 Lot #: N/A Exp Date: N/A
Known Value: 0.000 AC
Results: 0.000 AC 0.000 AC 0.000 AC

2) Low AC Test (Level 1,C):

Sim Ser #: DR3803 Lot #: 13070 Exp Date: 03/06/2015
Known Value: 0.020 AC
Results: 0.016 AC 0.018 AC 0.018 AC

3) Linearity Test (Level 1,C):

Sim Ser #: DR5143 Lot #: 13080 Exp Date: 03/19/2015
Known Value: 0.300 AC
Results: 0.297 AC 0.297 AC 0.298 AC

b. Configure Dry Gas Standard (Level 1,S):

1) Interferent Detected Tests (Level 1,B):

1. 0.10 AC + 0.05% Acetone

Sim Ser #: DR3846 Lot #: ICS 2 Exp Date: 27 Aug 2013 ²⁰¹⁵ DK
Known Value: 0.10 AC + 0.05% acetone
 Display reads "Interferent Detect"

2. Acetone

Sim Ser #: DR3375 Lot #: ACE 1 Exp Date: 27 Aug 2013 ²⁰¹⁵ DK
Known Value: 0.05% acetone
 Display reads "Interferent Detect"

3. Methanol

Sim Ser #: DR3454 Lot #: MET 1 Exp Date: 27 Aug 2013 ²⁰¹⁵ DK
Known Value: 0.10% Methanol
 Display reads "Interferent Detect"

4. Isopropanol
Sim Ser #: DR5131 Lot #: IPA1 Exp Date: 27 Aug 2014 ²⁰¹⁵ ^{OK}
Known Value: 0.10% Isopropanol
 Display reads "Interferent Detect"
- 2) Ambient Fail (Level 1,B): Display and test record are correct
3) Invalid Sample (Level 1,B): Display and test record are correct
4) Subject Test Timeout (Level 1,B): Display and test record are correct.
5) Sequence Aborted (Level 1,B): Display and test record are correct
6) Control Outside Tolerance (CMS mode): Display and test record are correct
7) RFI Detected Test (CMS mode): Display and test record are correct
8) Tank Pressure Below Min (CMS mode): Display reads correct
9) Subject Refused (CMS mode): Display and test record are correct
10) Difference Too Great (CMS mode): Display and test record are correct
11) Stability Test (Level 3,T,S):
Average 0.0811 Std Dev 0.0006 Rel Std Dev 0.75%

Maintenance/Repairs/Notes: Initial inspection. Recalibrated gas regulator.
This intox passed inspection & is ready for use in the field.

Deb Kashner
Field Inspector's Signature

12 Sept 2013
Date

Charles E. Eds
Reviewed by

13 SEP 13 ✓
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005941
Location = TOXL 8164.13.00 06/09
09/12/2013 08:38

***** Printer Test *****

12 char/line
abcdefghijkl
ABCDEFGHIJKL

16 char per line
abcdefghijklmnop
ABCDEFGHIJKLMN

21 chars per line
abcdefghijklmnopqrstu
ABCDEFGHIJKLMNQRSTU

24 characters per line
abcdefghijklmnopqrstuwx
ABCDEFGHIJKLMNQRSTUWX

32 characters per line
abcdefghijklmnopqrstuwx123456
ABCDEFGHIJKLMNQRSTUWXYZ123456

42 characters per line
abcdefghijklmnopqrstuwx1234567890abcdef
ABCDEFGHIJKLMNQRSTUWXYZ1234567890ABCDEF

Current Instrument Setup
Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On MemFull? No
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 1234567
Standard Cyl #? 123456
Standard Expiration? 11/11/2014
Oper No? 777777

Flow Cal. Date: 06/18/2013
Slope 695
Intercept -675557

IR Calibration Date: 06/18/2013
3um 9um

0th Coef(*100): -20306 -25137
1st Coef(*100): 291687 138772
2nd Coef(*100): 2450 1058
H2O adj(ng/l*10k): 562 452

***** Test End *****

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:

Print test

Form 106-18000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005941
Location = TOXL 8164.13.00 06/09
09/12/2013 15:59

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 23411080A1
Standard Cyl #? 40
Standard Expiration? 10/01/2013
Oper No? 777777

Flow Cal. Date: 06/18/2013
Slope 695
Intercept -675557

IR Calibration Date: 06/18/2013
 3um 9um

0th Coef(*100): -20306 -25137
1st Coef(*100): 291687 138772
2nd Coef(*100): 2450 1058
H2O adj(mg/l*10k): 562 452

***** Printer Test End *****



Operator Signature
DEB KASHUR

Remarks:

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005941
Location = TOXL 8164.13.00 06/09
09/12/2013 16:00

DIAGNOSTIC

Voltage/Current Test	Pass
RAM Test	Pass
EEPROM Checksum Test	Pass
Real Time Clock Test	Pass
DSP Test	Pass
Analytical Stability Test	Pass
Modem Test	Pass
Temp Regulation Test	Pass

Oper No. = 77777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Diagnostic Test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005941
Location = TOXL 8164.13.00 06/09
09/12/2013 16:04

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:05
02 Std. Sol.	0.000	16:06
03 Room Air	0.000	16:06
04 Std. Sol.	0.000	16:07
05 Room Air	0.000	16:07
06 Std. Sol.	0.000	16:08
07 Room Air	0.000	16:08

08 Sim Temp = 34.0°C

Simul Ser No = G10329

Std Sol No = MILLI Q

County = 08

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

zero alcohol test 0.000 ac

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005941
Location = TOXL 8164.13.00 06/09
09/12/2013 16:09

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:10
02 Std. Sol.	0.016	16:10
03 Room Air	0.000	16:11
04 Std. Sol.	0.018	16:12
05 Room Air	0.000	16:12
06 Std. Sol.	0.018	16:13
07 Room Air	0.000	16:13

08 Sim Temp = 34.0°C

Simul Ser No = DR3803
Std Sol No = 13070
County = 08 Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Low AC 0.020AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005941
Location = TOXL 8164.13.00 06/09
09/12/2013 16:14

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:14
02 Std. Sol.	0.297	16:15
03 Room Air	0.000	16:16
04 Std. Sol.	0.297	16:17
05 Room Air	0.000	16:17
06 Std. Sol.	0.298	16:18
07 Room Air	0.000	16:18

08 Sim Temp = 34.0°C

Simul Ser No = DR5143

Std Sol No = 13080

County = 08

Oper No. = 777777

Deb Keshur
Operator Signature
DEB KASHUR

Remarks:

Linearity test 0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005941
Location = TOXL 8164.13.00 06/09
09/12/2013 16:20

Test	AC	Time
01 Room Air	0.000	16:21
02 *Subject Test	INT*	16:21
03 Room Air	0.000	16:22

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 23411080A1
Cyl No = 40
Expiration Date = 10/01/2013
County = 08 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

*Interferent check
0.10 AC + 0.05% acetone*

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005941
Location = TOXL 8164.13.00 06/09
09/12/2013 16:22

Test	AC	Time
01 Room Air	0.000	16:23
02 *Subject Test	INT*	16:23
03 Room Air	0.000	16:24

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 23411080A1
Cyl No = 40
Expiration Date = 10/01/2013
County = 08 Oper No. = 777777



Operator Signature
DEB KASHUR

Remarks:

*Interferent detected check
0.05% acetone*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005941
Location = TOXL 8164.13.00 06/09
09/12/2013 16:24

Test	AC	Time
01 Room Air	0.000	16:25
02 *Subject Test	INT*	16:25
03 Room Air	0.000	16:26

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 23411080A1
Cyl No = 40
Expiration Date = 10/01/2013
County = 08 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

*Interferent check
0.10% Methanol*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005941
Location = TOXL 8164.13.00 06/09
09/12/2013 16:26

Test	AC	Time
01 Room Air	0.000	16:27
02 *Subject Test	INT*	16:27
03 Room Air	0.000	16:28

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 23411080A1
Cyl No = 40
Expiration Date = 10/01/2013
County = 08 Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

*Interferent check
0.10% Isopropanol*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005941
Location = TOXL 8164.13.00 06/09
09/12/2013 16:28

Test	AC	Time
01 Room Air	AMB*	16:29
02 Room Air	0.000	16:30

*Invalid Test
Check Ambient Conditions

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 23411080A1
Cyl No = 40
Expiration Date = 10/01/2013
County = 08 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Ambient fail test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005941
Location = TOXL 8164.13.00 06/09
09/12/2013 16:30

Test	AC	Time
01 Room Air	0.000	16:30
02 Invalid Sample	X.XXX	16:31
03 Room Air	0.000	16:31

*Invalid Test - Mouth Alcohol

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 23411080A1
Cyl No = 40
Expiration Date = 10/01/2013
County = 08 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Invalid sample test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005941
Location = TOXL 8164.13.00 06/09
09/12/2013 16:32

Test	AC	Time
01 Room Air	0.000	16:32
02 *Subject Test	0.000*	16:35
03 Room Air	0.000	16:36
No RFI Detected		

*Deficient Sample - Value Printed was
Highest Obtained

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 23411080A1
Cyl No = 40
Expiration Date = 10/01/2013
County = 08 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Subject test incorrect

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005941
Location = TOXL 8164.13.00 06/09
09/12/2013 16:36

Test	AC	Time
01 Room Air	ABT*	16:37
02 Room Air	0.000	16:37

*Sequence Aborted

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 23411080A1
Cyl No = 40
Expiration Date = 10/01/2013
County = 08 Oper No. = 777777



Operator Signature
DEB KASHUR

Remarks:

DK
Seq Sequence aborted test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005941
Location = TOXL 8164.13.00 06/09
09/12/2013 16:37

Test	AC	Time
01 Diagnostic	OK	16:38
02 Room Air	0.000	16:39
03 Subject Test 1	0.000	16:39
04 Room Air	0.000	16:41
05 Std. Gas	0.000*	16:42
06 Room Air	0.000	16:43

*Cal Check Out of Tolerance

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 23411080A1
Cyl No = 40
Expiration Date = 10/01/2013
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
DEB KASHUR

Remarks:

Control outside tolerance

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005941
Location = TOXL 8164.13.00 06/09
09/12/2013 16:43

Test	AC	Time
01 Diagnostic	OK	16:44
02 Room Air	RFI*	16:44
03 Room Air	0.000	16:45

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 23411080A1
Cyl No = 40
Expiration Date = 10/01/2013
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
DEB KASHUR

Remarks:

RFI test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005941
Location = TOXL 8164.13.00 06/09
09/12/2013 16:45

Test	AC	Time
01 Diagnostic	OK	16:46
02 Room Air	0.000	16:46
03 *Subject Test	REF*	16:46
04 Room Air	0.000	16:47

*Subject Test Refused

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 23411080A1
Cyl No = 40
Expiration Date = 10/01/2013
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
DEB KASHUR

Remarks:



Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005941
Location = TOXL 8164.13.00 06/09
09/12/2013 16:47

Test	AC	Time
01 Diagnostic	OK	16:48
02 Room Air	0.000	16:48
03 Subject Test 1	0.051*	16:48
04 Room Air	0.000	16:51
05 Std. Gas	0.081	16:52
06 Room Air	0.000	16:53
07 Subject Test 2	0.000*	16:53
08 Room Air	0.000	16:54

*Difference Too Great

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 23411080A1
Cyl No = 40
Expiration Date = 10/01/2013
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
DEB KASHUR

Remarks:



Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005941
Location = TOXL 8164.13.00 06/09
09/12/2013 16:55

Test	AC	Time
01 Room Air	0.000	16:55
02 Std. Gas	0.081	16:56
03 Room Air	0.000	16:56
04 Std. Gas	0.082	16:56
05 Room Air	0.000	16:57
06 Std. Gas	0.081	16:57
07 Room Air	0.000	16:58
08 Std. Gas	0.081	16:58
09 Room Air	0.000	16:59
10 Std. Gas	0.081	16:59
11 Room Air	0.000	17:00
12 Std. Gas	0.082	17:00
13 Room Air	0.000	17:00
14 Std. Gas	0.081	17:01
15 Room Air	0.000	17:01
16 Std. Gas	0.080	17:02
17 Room Air	0.000	17:02
18 Std. Gas	0.082	17:03
19 Room Air	0.000	17:03
20 Std. Gas	0.081	17:03
21 Room Air	0.000	17:04
22 Std. Gas	0.081	17:04
23 Room Air	0.000	17:05
24 Std. Gas	0.081	17:05
25 Room Air	0.000	17:06
26 Std. Gas	0.080	17:06
27 Room Air	0.000	17:07
28 Std. Gas	0.081	17:07
29 Room Air	0.000	17:07
30 Std. Gas	0.081	17:08
31 Room Air	0.000	17:08
32 Std. Gas	0.081	17:09
33 Room Air	0.000	17:09
34 Std. Gas	0.081	17:09
35 Room Air	0.000	17:10
36 Std. Gas	0.082	17:10
37 Room Air	0.000	17:11
38 Std. Gas	0.081	17:11
39 Room Air	0.000	17:12
40 Std. Gas	0.080	17:12
41 Room Air	0.000	17:13
42 Cal Check Stats:		
43 Average	0.0811	
44 Std Dev	0.0006	
45 Rel Std Dev	0.75%	

*Stability test
Deb Kashner
p.1/2*

Operator Signature
DEB KASHUR

Remarks:

Lot #: 23411080A1 Cyl #: 40

Exp. date: 10/01/2013

Form 106-I8000

Stability test
Deb Kashur
p. 2/2