

### INTOXILYZER® 8000 INITIAL INSPECTION

Serial Number: 80-005945

1. Power On and Warm-up:
  - a.  Power On
  - b.  5 min. Countdown from Initial Warm-up
  - c.  1 min. Countdown to Begin a Test
  - d.  Forms Load Complete
  
2. Setup and Checks:
  - a. Level 1:
    - 1) Passwords:  Field Inspector (Level 2)  Lab Personnel (Level 3)
    - 2) Gas Cylinder:  Gas Cylinder Setup (Level 1,S)
    - 3) Printer:  Printer Setup (Level 3,S,P)  
 Internal Printer Paper Feed  
 Internal Printer Test (Level 1,P)  
 External Printer Test (Level 1,P)
    - 4) Barcode Reader (Level 1,B):  Operator ID  
 ND License
  
  - b. Level 2:
    - 1)  Continuous Purge (Level 2,A)
    - 2)  Diagnostic (Level 2,D)
    - 3)  Prelim Data Entry (Level 2,E)
    - 4)  Ser Num/Version Info (Level 2,V): Ser Num 80-005945  
Version 8164.13
  
  - c. Level 3:
    - 1) DVM Monitor (Level 3,D,D): 3: 12583 s: 1 C: 47.00  
9: 13441 s: 3 B: 45   
(RFI indicator with keyed radio) 1
  
    - 2) Function (F) Keys:  F5 – Tone  
 F6 – Green LED  
 F7 – Red LED  
 F9 – Fan  
 F10 – Pump  
 F11 – Solenoid
  
    - 3) Atmospheric Monitor (Level 3,D,A): Atmo Sensor, P: 960 CF: 1.05
    - 4) Flow Monitor (Level 3,D,F): Flow Sensor, R: 217 F: 0.000 L/s

5) Temperature Monitor (Level 3,D,T): C: 47.0 A: 47.5 I: 0.0

B: 44.9 V: 0.0

6) Tank Monitor (Level 3,D,G): P (gauge): ≈40 P (display): 38

7)  General Setup (Level 3,S,G)

8)  Configure Start Test (Level 3,S,T)

9)  Network (Level 3,S,N)

3. Tests:

a. Configure Standard for Wet Bath Simulator (Level 1,S):

1) Zero Alcohol Test (Level 1,C):

Sim Ser #: G 10329 Lot #: N/A Exp Date: N/A

Known Value: 0.000 AC

Results: 0.000 AC 0.000 AC 0.000 AC

2) Low AC Test (Level 1,C):

Sim Ser #: DR3803 Lot #: 13070 Exp Date: 03/06/2015

Known Value: 0.020 AC

Results: 0.017 AC 0.018 AC <sup>OK</sup> 0.018 AC

3) Linearity Test (Level 1,C):

Sim Ser #: DR5143 Lot #: 13080 Exp Date: 03/19/2015

Known Value: 0.300 AC

Results: 0.292 AC 0.295 AC 0.295 AC

b. Configure Dry Gas Standard (Level 1,S):

1) Interferent Detected Tests (Level 1,B):

1. 0.10 AC + 0.05% Acetone

Sim Ser #: DR3846 Lot #: FCS 2 Exp Date: 08/27/2015

Known Value: 0.10 AC + 0.05% acetone

Display reads "Interferent Detect"

2. Acetone

Sim Ser #: DR3375 Lot #: ACE 1 Exp Date: 08/27/2015

Known Value: 0.05% acetone

Display reads "Interferent Detect"

3. Methanol

Sim Ser #: DR3454 Lot #: MET 1 Exp Date: 08/27/2015

Known Value: 0.10% methanol

Display reads "Interferent Detect"

4. Isopropanol

Sim Ser #: DR5131 Lot #: IPA1 Exp Date: 08/27/2015

Known Value: 0.10<sup>g%</sup> Isopropanol

Display reads "Interferent Detect"

- 2)  Ambient Fail (Level 1,B): Display and test record are correct
- 3)  Invalid Sample (Level 1,B): Display and test record are correct
- 4)  Subject Test Timeout (Level 1,B): Display and test record are correct.
- 5)  Sequence Aborted (Level 1,B): Display and test record are correct
- 6)  Control Outside Tolerance (CMS mode): Display and test record are correct
- 7)  RFI Detected Test (CMS mode): Display and test record are correct
- 8)  Tank Pressure Below Min (CMS mode): Display reads correct
- 9)  Subject Refused (CMS mode): Display and test record are correct
- 10)  Difference Too Great (CMS mode): Display and test record are correct
- 11)  Stability Test (Level 3,T,S):  
Average <sup>0.0012<sup>bk</sup></sup> 0.0801 Std Dev 0.0011 Rel Std Dev <sup>1.01%</sup> 1.34%

Maintenance/Repairs/Notes: Initial inspection

This intox passed inspection and can be used in the field.

Deb Kashner  
Field Inspector's Signature

13 Sept 13  
Date

Charles E. [Signature]  
Reviewed by

13 SEP 13  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005945  
Location = TOXL      8164.13.00 06/09  
09/13/2013      08:46

\*\*\*\*\* Printer Test \*\*\*\*\*

12 char/line  
abcdefghijkl  
ABCDEFGHIJKL

16 char per line  
abcdefghijklmnop  
ABCDEFGHIJKLMN

21 chars per line  
abcdefghijklmnopqrstu  
ABCDEFGHIJKLMNQRSTU

24 characters per line  
abcdefghijklmnopqrstuwx  
ABCDEFGHIJKLMNQRSTUWX

32 characters per line  
abcdefghijklmnopqrstuwx123456  
ABCDEFGHIJKLMNQRSTUWXYZ123456

42 characters per line  
abcdefghijklmnopqrstuwx1234567890abcdef  
ABCDEFGHIJKLMNQRSTUWXYZ1234567890ABCDEF

Current Instrument Setup  
Data Entry Mode: Enabled  
Start Test Sequence: DABACABA  
Display Prelim Rslt? Yes  
Display Third Digit? Yes  
Inhib Printer(Y/N)? No  
Display Volume? No  
Disable On MemFull? No  
# of Print Copies? 1  
Select Std (D/W/I)? Dry  
Standard Value? 0.080  
Standard Lot #? 123456789  
Standard Cyl #? 1123456  
Standard Expiration? 11/11/2015  
Oper No? 77777

Flow Cal. Date: 06/20/2013  
Slope 637  
Intercept -566407

IR Calibration Date: 06/20/2013  
3um      9um  
-----  
0th Coef(\*100): -21655      -20408  
1st Coef(\*100): 276095      136803  
2nd Coef(\*100): 2542      1254  
H2O adj(mg/l\*10K): 708      490

\*\*\*\*\* Test End \*\*\*\*\*

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Print test*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005945  
Location = TOXL                    8164.13.00 06/09  
09/13/2013                            09:42

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&\*()\_+?

Current Instrument Setup

Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        No  
# of Print Copies?         1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            23411080A1  
Standard Cyl #?            40  
Standard Expiration?       10/01/2013  
Oper No?                    777777

Flow Cal. Date:            06/20/2013  
Slope                      637  
Intercept                  -566407

IR Calibration Date:        06/20/2013

	3um	9um
0th Coef(*100):	-21655	-20408
1st Coef(*100):	276095	136803
2nd Coef(*100):	2542	1254
H2O adj(mg/l*10k):	708	490

\*\*\*\*\* Printer Test End \*\*\*\*\*

Deb Kashur

Operator Signature  
DEB KASHUR

Remarks:

Print test

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005945  
Location = TOXL                8164.13.00 06/09  
09/13/2013                        09:44

DIAGNOSTIC

Voltage/Current Test	Pass
RAM Test	Pass
EEPROM Checksum Test	Pass
Real Time Clock Test	Pass
DSP Test	Pass
Analytical Stability Test	Pass
Modem Test	Pass
Temp Regulation Test	Pass

Oper No. = 777777

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Diagnostic test*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005945  
Location = TOXL      8164.13.00 06/09  
09/13/2013      10:08

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:08
02 Std. Sol.	0.000	10:09
03 Room Air	0.000	10:10
04 Std. Sol.	0.000	10:10
05 Room Air	0.000	10:11
06 Std. Sol.	0.000	10:11
07 Room Air	0.000	10:12

08 Sim Temp = 34.0°C

Simul Ser No = G10329

Std Sol No = MILLI Q

County = 08

Oper No. = 777777

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Zero alcohol test      0.000 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005945  
Location = TOXL      8164.13.00 06/09  
09/13/2013      10:14

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:14
02 Std. Sol.	0.017	10:15
03 Room Air	0.000	10:15
04 Std. Sol.	0.018	10:16
05 Room Air	0.000	10:17
06 Std. Sol.	0.018	10:17
07 Room Air	0.000	10:18

08 Sim Temp = 34.0°C

Simul Ser No = DR3803

Std Sol No = 13070

County = 08

Oper No. = 777777

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*low AC      0.020 AC*

Form 106-18000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005945  
Location = TOXL                8164.13.00 06/09  
09/13/2013                            10:20

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:20
02 Std. Sol.	0.292	10:21
03 Room Air	0.000	10:22
04 Std. Sol.	0.295	10:22
05 Room Air	0.000	10:23
06 Std. Sol.	0.295	10:24
07 Room Air	0.000	10:24

08 Sim Temp = 34.0°C

Simul Ser No = DR5143

Std Sol No = 13080

County = 08

Oper No. = 777777

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Linearity test 0.300 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005945  
Location = TOXL      8164.13.00 06/09  
09/13/2013      13:46

Test	AC	Time
01 Room Air	0.000	13:47
02 *Subject Test	INT*	13:47
03 Room Air	0.000	13:48

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 23411080A1  
Cyl No = 38  
Expiration Date = 10/01/2013  
County = 08      Oper No. = 777777

Deb Kashur

Operator Signature  
DEB KASHUR

Remarks:

*Interferent check*  
*0.10AC + 0.05% acetone*  
Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005945  
Location = TOXL                    8164.13.00 06/09  
09/13/2013                            13:49

Test	AC	Time
01 Room Air	0.000	13:50
02 *Subject Test	INT*	13:50
03 Room Air	0.000	13:51

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 23411080A1  
Cyl No = 38  
Expiration Date = 10/01/2013  
County = 08                            Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Interferent check  
0.05% acetone*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005945  
Location = TOXL                    8164.13.00 06/09  
09/13/2013                            13:51

Test	AC	Time
01 Room Air	0.000	13:52
02 *Subject Test	INT*	13:52
03 Room Air	0.000	13:53

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 23411080A1  
Cyl No = 38  
Expiration Date = 10/01/2013  
County = 08                            Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Interferent check  
0.1090 Methanol*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005945  
Location = TOXL                    8164.13.00 06/09  
09/13/2013                            13:54

Test	AC	Time
01 Room Air	0.000	13:55
02 *Subject Test	INT*	13:55
03 Room Air	0.000	13:56

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 23411080A1  
Cyl No = 38  
Expiration Date = 10/01/2013  
County = 08                            Oper No. = 777777

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Interferent check  
0.10% Isopropanol*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005945  
Location = TOXL                8164.13.00 06/09  
09/13/2013                        13:58

Test	AC	Time
01 Room Air	AMB*	13:59
02 Room Air	0.000	13:59

\*Invalid Test  
Check Ambient Conditions

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 23411080A1  
Cyl No = 38  
Expiration Date = 10/01/2013  
County = 08                          Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*ambient fail test*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005945  
Location = TOXL                    8164.13.00 06/09  
09/13/2013                            14:02

Test	AC	Time
01 Room Air	0.000	14:03
02 Invalid Sample	X.XXX	14:03
03 Room Air	0.000	14:04

\*Invalid Test - Mouth Alcohol

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 23411080A1  
Cyl No = 38  
Expiration Date = 10/01/2013  
County = 08                            Oper No. = 777777



Operator Signature  
DEB KASHUR

Remarks:

*Invalid sample test*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005945  
Location = TOXL                    8164.13.00 06/09  
09/13/2013                            14:04

Test	AC	Time
01 Room Air	0.000	14:05
02 *Subject Test	0.000*	14:09
03 Room Air	0.000	14:09

No RFI Detected

\*Deficient Sample - Value Printed was  
Highest Obtained

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 23411080A1  
Cyl No = 38  
Expiration Date = 10/01/2013  
County = 08                            Oper No. = 777777

Deb Kashur

Operator Signature  
DEB KASHUR

Remarks:

~~Interferent~~ DK  
Subject test timeout

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005945  
Location = TOXL                    8164.13.00 06/09  
09/13/2013                            14:15

Test	AC	Time
01 Room Air	0.000	14:15
02 *Subject Test	ABT*	14:16
03 Room Air	0.000	14:16

\*Sequence Aborted

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 23411080A1  
Cyl No = 38  
Expiration Date = 10/01/2013  
County = 08                            Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Sequence aborted test*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005945  
Location = TOXL                    8164.13.00 06/09  
09/13/2013                            14:16

Test	AC	Time
01 Diagnostic	OK	14:17
02 Room Air	0.000	14:17
03 Subject Test 1	0.000	14:18
04 Room Air	0.000	14:20
05 Std. Gas	0.000*	14:21
06 Room Air	0.000	14:22

\*Cal Check Out of Tolerance

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 23411080A1  
Cyl No = 38  
Expiration Date = 10/01/2013  
County = 08                          Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kashur

Operator Signature  
DEB KASHUR

Remarks:

*Control outside tolerance test*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005945  
Location = TOXL                    8164.13.00 06/09  
09/13/2013                            14:26

Test	AC	Time
01 Diagnostic	OK	14:27
02 Room Air	RFI*	14:27
03 Room Air	0.000	14:28

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 23411080A1  
Cyl No = 38  
Expiration Date = 10/01/2013  
County = 08                            Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

  
\_\_\_\_\_  
Operator Signature  
DEB KASHUR

Remarks:

*RFI test*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005945  
Location = TOXL      8164.13.00 06/09  
09/13/2013      14:28

Test	AC	Time
01 Diagnostic	OK	14:29
02 Room Air	0.000	14:30
03 *Subject Test	REF*	14:30
04 Room Air	0.000	14:30

\*Subject Test Refused

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 23411080A1  
Cyl No = 38  
Expiration Date = 10/01/2013  
County = 08      Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Subject test refused*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005945  
Location = TOXL                    8164.13.00 06/09  
09/13/2013                            14:31

Test	AC	Time
01 Diagnostic	OK	14:32
02 Room Air	0.000	14:32
03 Subject Test 1	0.000*	14:35
04 Room Air	0.000	14:37
05 Std. Gas	0.081	14:38
06 Room Air	0.000	14:40
07 Subject Test 2	0.050*	14:40
08 Room Air	0.000	14:41

\*Difference Too Great

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 23411080A1  
Cyl No = 38  
Expiration Date = 10/01/2013  
County = 08                    Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kashur  
Operator Signature  
DEB KASHUR

Remarks:

*Difference too Great test*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005945  
Location = TOXL                    8164.13.00 06/09  
09/13/2013                            14:44

Test	AC	Time
01 Room Air	0.000	14:44
02 Std. Gas	0.081	14:45
03 Room Air	0.000	14:45
04 Std. Gas	0.081	14:46
05 Room Air	0.000	14:46
06 Std. Gas	0.081	14:47
07 Room Air	0.000	14:47
08 Std. Gas	0.081	14:47
09 Room Air	0.000	14:48
10 Std. Gas	0.080	14:48
11 Room Air	0.000	14:49
12 Std. Gas	0.082	14:49
13 Room Air	0.000	14:50
14 Std. Gas	0.082	14:50
15 Room Air	0.000	14:51
16 Std. Gas	0.081	14:51
17 Room Air	0.000	14:52
18 Std. Gas	0.081	14:52
19 Room Air	0.000	14:52
20 Std. Gas	0.080	14:53
21 Room Air	0.000	14:53
22 Std. Gas	0.081	14:54
23 Room Air	0.000	14:54
24 Std. Gas	0.082	14:55
25 Room Air	0.000	14:55
26 Std. Gas	0.082	14:55
27 Room Air	0.000	14:56
28 Std. Gas	0.082	14:56
29 Room Air	0.000	14:57
30 Std. Gas	0.080	14:57
31 Room Air	0.000	14:58
32 Std. Gas	0.080	14:58
33 Room Air	0.000	14:59
34 Std. Gas	0.082	14:59
35 Room Air	0.000	14:59
36 Std. Gas	0.080	15:00
37 Room Air	0.000	15:00
38 Std. Gas	0.082	15:01
39 Room Air	0.000	15:01
40 Std. Gas	0.082	15:02
41 Room Air	0.000	15:02
42 Cal Check Stats:		
43 Average	0.0812	
44 Std Dev	0.0008	
45 Rel Std Dev	1.01%	

*Stability test  
Job Kasher  
p. 1/2*

Operator Signature  
DEB KASHUR

Remarks:

Lot #: 23411680A1

Cyl #: 38  
Form 106-I8000

Exp. date: 10/01/2013

Stability test  
Deb Kashur  
p. 2/2