



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59281 (03-2014)

Serial Number 80-005946	Instrument Location Williston - Williams County Jail
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Properly Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. RFI Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if Any): N/A
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer®8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature * 289 <i>Sgt. Chad Cook</i>	Date 9/16/2015 ✓
Reviewed By (Crime Laboratory Use Only) <i>Deb Shanever</i>	Date 18 Feb 2016

This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Certified By: *Charles E. Egan* Certified Date: 18 Nov. 2016

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005946
Location = WILL 8164.13.00 06/09
09/16/2015 08:34

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 29614080A2
Standard Cyl #? 9
Standard Expiration? 12/05/2016
Oper No? 010910

Flow Cal. Date: 08/10/2015
Slope 673
Intercept -638348

IR Calibration Date: 08/10/2015

	3um	9um
0th Coef(*100):	-22507	-13665
1st Coef(*100):	273651	130615
2nd Coef(*100):	5249	2073
H2O adj(mg/l*10k):	617	504

***** Printer Test End *****



Operator Signature
MICHAEL ROARK

Remarks:

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005946
Location = WILL 8164.13.00 06/09
09/16/2015 08:37

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:37
02 Std. Gas	0.080	08:38
03 Room Air	0.000	08:38
04 Std. Gas	0.080	08:39
05 Room Air	0.000	08:39
06 Std. Gas	0.081	08:39
07 Room Air	0.000	08:40

Lot No = 29614080A2
Cyl No = 9
Exp Date = 12/05/2016
County = 53 Oper No. = 010910



Operator Signature
MICHAEL ROARK

Remarks:

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005946
Location = WILL 8164.13.00 06/09
09/16/2015 08:35

DRY CAL CHECK

Test	AC	Time
01 Room Air	RFI*	08:36
02 Room Air	0.000	08:36

*Invalid Test
Inhibited - RFI

Lot No = 29614080A2
Cyl No = 9
Exp Date = 12/05/2016
County = 53 Oper No. = 010910



Operator Signature
MICHAEL ROARK

Remarks: