



ETHANOL BREATH STANDARD CYLINDER REPORT
OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Matthew Anderson</u>		
Location: <u>Richland County Jail</u>		
Intoxilyzer® Serial Number: <u>80-004206</u>		
Gas Lot Number: <u>659358</u>	Gas Cylinder Number: <u>52</u>	Expiration Date: <u>9-03-11</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>083</u> %		
Test 2: 0. <u>082</u> %		
Test 3: 0. <u>082</u> %		
(ATTACH TEST RECORD)		

[Signature]
Chemical Test Operator Signature

9-17-10
Date

Deb Kashner
Reviewed By

20 Sept 2010
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004206
Location = WAHP 8164.13.00 06/09
09/17/2010 13:55

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:55
02 Std. Gas	0.083	13:56
03 Room Air	0.000	13:56
04 Std. Gas	0.082	13:57
05 Room Air	0.000	13:57
06 Std. Gas	0.082	13:58
07 Room Air	0.000	13:58

Lot No = 659358
Cyl No = 52
Exp Date = 09/03/2011
County = 39 Oper No. = 020401



Operator Signature
MATTHEW W. ANDERSON

Remarks:

Form 106-I8000

80-004206
P. 107 of 107
pm