



# ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (5-2011)

Chemical Test Operator Name (Print) <i>Jonathan Lampi</i>		
Location <i>GFPD</i>	Intoxilyzer® Serial Number <i>80-004198</i>	
Gas Lot Number <i>16512080A2</i>	Gas Cylinder Number <i>24</i>	Gas Expiration Date <i>08-01-14</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).  
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
  - A. *0.081* AC
  - B. *0.081* AC
  - C. *0.080* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
  - A. Ethanol Breath Standard Cylinder Report (SFN59282, Form 105-G).
  - B. ACA Test Record.

Chemical Test Operator Signature <i>Jonathan Lampi</i>	Date <i>09-17-13</i>
Reviewed By (Crime Laboratory Use Only) <i>Deb Kashner</i>	Date <i>20 Sept 2013</i>

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004198  
Location = GFPD      8164.13.00 06/09  
09/17/2013      23:06

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	23:06
02 Std. Gas	0.081	23:07
03 Room Air	0.000	23:07
04 Std. Gas	0.081	23:08
05 Room Air	0.000	23:08
06 Std. Gas	0.080	23:09
07 Room Air	0.000	23:09

Lot No = 16512080A2  
Cyl No = 24  
Exp Date = 08/01/2014  
County = 18      Oper No. = 039947

  
Operator Signature  
JONATHAN LAMPI

Remarks:

Form 106-I8000

*cylinder change*