



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

| |
|---|
| Chemical Test Operator Name: <u>Chad L. Harmon</u> |
| Location: <u>FYPD</u> |
| Intoxilyzer® Serial Number: <u>80-004954</u> |
| Gas Lot Number: <u>709352</u> Gas Cylinder Number: <u>62-26 DK</u> Expiration Date: <u>12-23-11</u> |
| <input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed |
| Instrument Test Results (Report to 3 Digits; Example, 0.081%) Test 1: <u>0.079</u> % Test 2: <u>0.079</u> % Test 3: <u>0.080</u> % <p style="text-align: right;">(ATTACH TEST RECORD)</p> |

Chad L. Harmon
Chemical Test Operator Signature

09-19-10
Date

Deb Kashur
Reviewed By

21 Oct 2010
Date

80-004954
P. 59460
RM


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = FYPD 8164.13.00 06/09
09/19/2010 02:48

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 02:48 |
| 02 Std. Gas | 0.079 | 02:49 |
| 03 Room Air | 0.000 | 02:49 |
| 04 Std. Gas | 0.079 | 02:50 |
| 05 Room Air | 0.000 | 02:50 |
| 06 Std. Gas | 0.080 | 02:50 |
| 07 Room Air | 0.000 | 02:51 |

Lot No = 709352
Cyl No = 26
Exp Date = 12/23/2011
County = 61 Oper No. = 010606


Operator Signature
CHAD L. HARMON

Remarks:

Form 106-I8000

80-004954
P. 60460
JAH