

**INTOXILYZER® 8000 INITIAL INSPECTION**

Serial Number: 80-005939

1. Power On and Warm-up:
  - a.  Power On
  - b.  5 min. Countdown from Initial Warm-up
  - c.  1 min. Countdown to Begin a Test
  - d.  Forms Load Complete
  
2. Setup and Checks:
  - a. Level 1:
    - 1) Passwords:  Field Inspector (Level 2)  Lab Personnel (Level 3)
    - 2) Gas Cylinder:  Gas Cylinder Setup (Level 1,S)
    - 3) Printer:  Printer Setup (Level 3,S,P)  
 Internal Printer Paper Feed  
 Internal Printer Test (Level 1,P)  
 External Printer Test (Level 1,P)
    - 4) Barcode Reader (Level 1,B):  Operator ID  
 ND License
  
  - b. Level 2:
    - 1)  Continuous Purge (Level 2,A)
    - 2)  Diagnostic (Level 2,D)
    - 3)  Prelim Data Entry (Level 2,E)
    - 4)  Ser Num/Version Info (Level 2,V): Ser Num 80-005939  
Version 8164.13
  
  - c. Level 3:
    - 1) DVM Monitor (Level 3,D,D): 3: 12695 s: 2 C: 47.06  
9: 13228 s: 1 B: 44 0  
(RFI indicator with keyed radio) 1
  
    - 2) Function (F) Keys:  F5 – Tone  
 F6 – Green LED  
 F7 – Red LED  
 F9 – Fan  
 F10 – Pump  
 F11 – Solenoid
  
    - 3) Atmospheric Monitor (Level 3,D,A): Atmo Sensor, P: 951 CF: 1.06
    - 4) Flow Monitor (Level 3,D,F): Flow Sensor, R: 233 F: 0.000 L/s

- 5) Temperature Monitor (Level 3,D,T): C: 47.0 A: 48.0 I: 0.0  
B: 45.0 V: 0.0  
6) Tank Monitor (Level 3,D,G): P (gauge): 125 P (display): 120  
7)  General Setup (Level 3,S,G)  
8)  Configure Start Test (Level 3,S,T)  
9)  Network (Level 3,S,N)

3. Tests:

a. Configure Standard for Wet Bath Simulator (Level 1,S):

- 1) Zero Alcohol Test (Level 1,C):  
Sim Ser #: DR5188 Lot #: MILLI @ H<sub>2</sub>O Exp Date: NA  
Known Value: 0.000 AC  
Results: 0.000 AC 0.000 AC 0.000 AC
- 2) Low AC Test (Level 1,C):  
Sim Ser #: DR3455 Lot #: 12210 Exp Date: 11-2-14  
Known Value: 0.020 AC  
Results: 0.019 AC 0.018 AC 0.020 AC
- 3) Linearity Test (Level 1,C):  
Sim Ser #: DR3847 Lot #: 13020 Exp Date: 1-15-15  
Known Value: 0.300 AC  
Results: 0.300 AC 0.300 AC 0.301 AC

b. Configure Dry Gas Standard (Level 1,S):

- 1) Interferent Detected Tests (Level 1,B):
1. 0.10 AC + 0.05% Acetone  
Sim Ser #: DR3377 Lot #: ICS#1 Exp Date: 4-9-14  
Known Value: 0.10 AC + 0.05% ACETONE  
 Display reads "Interferent Detect"
2. Acetone  
Sim Ser #: G8919 Lot #: ACE1 Exp Date: 8-27-15  
Known Value: 0.05% ACETONE  
 Display reads "Interferent Detect"
3. Methanol  
Sim Ser #: G10573 Lot #: MET1 Exp Date: 8-27-15  
Known Value: 0.10% METHANOL  
 Display reads "Interferent Detect"

4. Isopropanol  
Sim Ser #: G9998 Lot #: IPA1 Exp Date: 8-27-15  
Known Value: 0.10% ISOPROPANOL  
 Display reads "Interferent Detect"
- 2)  Ambient Fail (Level 1,B): Display and test record are correct
  - 3)  Invalid Sample (Level 1,B): Display and test record are correct
  - 4)  Subject Test Timeout (Level 1,B): Display and test record are correct.
  - 5)  Sequence Aborted (Level 1,B): Display and test record are correct
  - 6)  Control Outside Tolerance (CMS mode): Display and test record are correct
  - 7)  RFI Detected Test (CMS mode): Display and test record are correct
  - 8)  Tank Pressure Below Min (CMS mode): Display reads correct
  - 9)  Subject Refused (CMS mode): Display and test record are correct
  - 10)  Difference Too Great (CMS mode): Display and test record are correct
  - 11)  Stability Test (Level 3,T,S):  
Average 0.0812 Std Dev 0.0008 Rel Std Dev 0.95%

Maintenance/Repairs/Notes: \_\_\_\_\_

THIS UNIT IS ACCEPTABLE TO BE USED IN THE FIELD.

Charles E. Eden

Field Inspector's Signature

19 SEP 13

Date

Deb Keshu

Reviewed by

19 Sept 2013

Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer  
North Dakota Model 8000 SN 80-005939  
Location = TOXL 8164.13.00 06/09  
09/19/2013 09:05

\*\*\*\*\* Printer Test \*\*\*\*\*

12 char/line  
abcdefghijkl  
ABCDEFGHIJKL

16 char per line  
abcdefghijklmnop  
ABCDEFGHIJKLMN

21 chars per line  
abcdefghijklmnopqrstu  
ABCDEFGHIJKLMN

24 characters per line  
abcdefghijklmnopqrstuwx  
ABCDEFGHIJKLMN

32 characters per line  
abcdefghijklmnopqrstuwx123456  
ABCDEFGHIJKLMN

42 characters per line  
abcdefghijklmnopqrstuwx1234567890abcdef  
ABCDEFGHIJKLMN

Current Instrument Setup

Data Entry Mode: Enabled  
Start Test Sequence: DABACABA  
Display Prelim Rslt? Yes  
Display Third Digit? Yes  
Inhib Printer(Y/N)? No  
Display Volume? No  
Disable On MemFull? No  
# of Print Copies? 1  
Select Std (D/W/I)? Dry  
Standard Value? 0.080  
Standard Lot #? 23411080A1  
Standard Cyl #? 71  
Standard Expiration? 10/01/2013  
Oper No? 666666

Flow Cal. Date: 06/18/2013  
Slope 676  
Intercept -672473

IR Calibration Date: 06/18/2013

	Sum	Sum
0th Coef(*100):	-16377	-17793
1st Coef(*100):	286199	137970
2nd Coef(*100):	3159	1297
H2O adj(mg/l*10K):	522	398

\*\*\*\*\* Test End \*\*\*\*\*



Operator Signature  
CHARLES EDER

Remarks: **INTERNAL  
PRINT  
TEST**

Form 106-18000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005939  
Location = TOXL                    8164.13.00 06/09  
09/19/2013    09:10

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&\*()\_+?

Current Instrument Setup

Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        No  
# of Print Copies?         1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            23411080A1  
Standard Cyl #?            71  
Standard Expiration?       10/01/2013  
Oper No?                    666666

Flow Cal. Date:            06/18/2013  
Slope                        676  
Intercept                  -672473

IR Calibration Date:        06/18/2013

	3um	9um
0th Coef(*100):	-16377	-17793
1st Coef(*100):	286199	137970
2nd Coef(*100):	3159	1297
H2O adj(mg/l*10k):	522	398

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
CHARLES EDER

Remarks: *EXTERNAL PRINT  
TEST*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer	Alcohol Analyzer
North Dakota Model 8000	SN 80-005939
Location = TOXL	8164.13.00 06/09
09/19/2013	09:12

DIAGNOSTIC

Voltage/Current Test	Pass
RAM Test	Pass
EEPROM Checksum Test	Pass
Real Time Clock Test	Pass
DSP Test	Pass
Analytical Stability Test	Pass
Modem Test	Pass
Temp Regulation Test	Pass

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *DIAGNOSTIC*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005939  
Location = TOXL                8164.13.00 06/09  
09/19/2013                        09:19

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:20
02 Std. Sol.	0.000	09:20
03 Room Air	0.000	09:21
04 Std. Sol.	0.000	09:21
05 Room Air	0.000	09:22
06 Std. Sol.	0.000	09:23
07 Room Air	0.000	09:23

08 Sim Temp = 34.0°C

Simul Ser No = DR5188  
Std Sol No = MILLIQ H2O  
County = 08                      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: ~~0.10AC + 0.05 CEE 0.000AC~~

ZERO ALCOHOL TEST

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005939  
Location = TOXL      8164.13.00 06/09  
09/19/2013      09:26

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:27
02 Std. Sol.	0.019	09:28
03 Room Air	0.000	09:29
04 Std. Sol.	0.018	09:29
05 Room Air	0.000	09:30
06 Std. Sol.	0.020	09:31
07 Room Air	0.000	09:31

08 Sim Temp = 34.0°C

Simul Ser No = DR3455  
Std Sol No = 12210  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *Low AC*  
*0.020 AC Sim. Sol.*



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005939  
Location = TOXL      8164.13.00 06/09  
09/19/2013      09:34

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:35
02 Std. Sol.	0.300	09:35
03 Room Air	0.000	09:36
04 Std. Sol.	0.300	09:37
05 Room Air	0.000	09:37
06 Std. Sol.	0.301	09:38
07 Room Air	0.000	09:39

08 Sim Temp = 34.0°C

Simul Ser No = DR3847

Std Sol No = 13020

County = 08

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

*LINEARITY TEST*

*0.300 AC Sim. Sol.*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005939  
Location = TOXL                    8164.13.00 06/09  
09/19/2013                            09:52

Test	AC	Time
01 Room Air	0.000	09:53
02 *Subject Test	INT*	09:53
03 Room Air	0.000	09:54

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = 0.10AC+0.05ACE  
Dr. Lic. = ND/DIS821456  
Lot No = 23411080A1  
Cyl No = 71  
Expiration Date = 10/01/2013  
County = 08                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: **INTERFERENT DETECTED CHECK**  
**0.10AC + 0.05% ACETONE**

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005939  
Location = TOXL                8164.13.00 06/09  
09/19/2013                        10:04

Test	AC	Time
01 Room Air	0.000	10:05
02 *Subject Test	INT*	10:05
03 Room Air	0.000	10:06

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = 0.05 ACETONE  
Dr. Lic. = ND/DIS821456  
Lot No = 23411080A1  
Cyl No = 71  
Expiration Date = 10/01/2013  
County = 08                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: **INTERFERENT DETECTED CHECK**  
**0.05% ACETONE**

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer  
North Dakota Model 8000 SN 80-005939  
Location = TOXL 8164.13.00 06/09  
09/19/2013 10:08

Test	AC	Time
01 Room Air	0.000	10:09
02 *Subject Test	INT*	10:10
03 Room Air	0.000	10:11

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female Weight = 150  
Test = DUI Cit = 0.10 METHANOL  
Dr. Lic. = ND/DIS821456  
Lot No = 23411080A1  
Cyl No = 71  
Expiration Date = 10/01/2013  
County = 08 Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: 0.10% METHANOL

*INTERFERENT DETECTED CHECK*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005939  
Location = TOXL                 8164.13.00 06/09  
09/19/2013                        10:12

Test	AC	Time
01 Room Air	0.000	10:12
02 *Subject Test	INT*	10:13
03 Room Air	0.000	10:14

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                 Weight = 150  
Test = DUI                         Cit = 0.10 IPA  
Dr. Lic. = ND/DIS821456  
Lot No = 23411080A1  
Cyl No = 71  
Expiration Date = 10/01/2013  
County = 08                        Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *INTERFERENT DETECTED CHECK*  
*0.10% ISOPROPANOL*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005939  
Location = TOXL                      8164.13.00 06/09  
09/19/2013                                  10:17

Test	AC	Time
01 Room Air	AMB*	10:18
02 Room Air	0.000	10:18

\*Invalid Test  
Check Ambient Conditions

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                      Weight = 150  
Test = DUI                              Cit = AMBIENT FAIL  
Dr. Lic. = ND/DIS821456  
Lot No = 23411080A1  
Cyl No = 71  
Expiration Date = 10/01/2013  
County = 08                              Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *AMBIENT FAIL CHECK*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005939  
Location = TOXL      8164.13.00 06/09  
09/19/2013      10:24

Test	AC	Time
01 Room Air	0.000	10:25
02 Invalid Sample	X.XXX	10:25
03 Room Air	0.000	10:26

\*Invalid Test - Mouth Alcohol

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = INVALID SAMPLE  
Dr. Lic. = ND/DIS821456  
Lot No = 23411080A1  
Cyl No = 71  
Expiration Date = 10/01/2013  
County = 08      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *INVALID SAMPLE CHECK*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005939  
Location = TOXL                8164.13.00 06/09  
09/19/2013                        10:31

Test	AC	Time
01 Room Air	0.000	10:32
02 *Subject Test	0.000*	10:35
03 Room Air	0.000	10:36

No RFI Detected

\*Deficient Sample - Value Printed was  
Highest Obtained

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = TEST TIMEOUT  
Dr. Lic. = ND/DIS821456  
Lot No = 23411080A1  
Cyl No = 71  
Expiration Date = 10/01/2013  
County = 08                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *SUBJECT TEST TIMEOUT*



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005939  
Location = TOXL                    8164.13.00 06/09  
09/19/2013                            10:38

Test	AC	Time
01 Room Air	0.000	10:38
02 *Subject Test	ABT*	10:39
03 Room Air	0.000	10:39

\*Sequence Aborted

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = SEQUENCE ABORT  
Dr. Lic. = ND/DIS821456  
Lot No = 23411080A1  
Cyl No = 71  
Expiration Date = 10/01/2013  
County = 08                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: **SEQUENCE ABORTED**

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005939  
Location = TOXL                8164.13.00 06/09  
09/19/2013                      10:40

Test	AC	Time
01 Diagnostic	OK	10:41
02 Room Air	0.000	10:41
03 Subject Test 1	0.098	10:42
04 Room Air	0.000	10:44
05 Std. Gas	0.087*	10:45
06 Room Air	0.000	10:46

\*Cal Check Out of Tolerance

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                      Weight = 150  
Test = DUI                              Cit = CON OUT TOL  
Dr. Lic. = ND/DIS821456  
Lot No = 23411080A1  
Cyl No = 71  
Expiration Date = 10/01/2013  
County = 08                              Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks: *CONTROL OUTSIDE  
TOLERANCE*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005939  
Location = TOXL                    8164.13.00 06/09  
09/19/2013                            10:51

Test	AC	Time
01 Diagnostic	OK	10:53
02 Room Air	0.000	10:53
03 *Subject Test	RFI*	10:54
04 Room Air	0.000	10:55

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = RFI DETECT  
Dr. Lic. = ND/DIS821456  
Lot No = 23411080A1  
Cyl No = 71  
Expiration Date = 10/01/2013  
County = 08                            Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks: *RFI DETECT*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005939  
Location = TOXL                    8164.13.00 06/09  
09/19/2013                            10:57

Test	AC	Time
01 Diagnostic	OK	10:58
02 Room Air	0.000	10:58
03 Subject Test 1	0.099	10:59
04 Room Air	0.000	11:01
05 Std. Gas	0.080	11:02
06 Room Air	0.000	11:04
07 *Subject Test	REF*	11:04
08 Room Air	0.000	11:05

\*Subject Test Refused

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = SUBJECT REFUSED  
Dr. Lic. = ND/DIS821456  
Lot No = 23411080A1  
Cyl No = 71  
Expiration Date = 10/01/2013  
County = 08                            Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks: **SUBJECT TEST REFUSED**

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005939  
Location = TOXL                    8164.13.00 06/09  
09/19/2013                            11:07

Test	AC	Time
01 Diagnostic	OK	11:08
02 Room Air	0.000	11:09
03 Subject Test 1	0.043*	11:09
04 Room Air	0.000	11:11
05 Std. Gas	0.080	11:13
06 Room Air	0.000	11:14
07 Subject Test 2	0.099*	11:14
08 Room Air	0.000	11:15

\*Difference Too Great

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = DIFF TOO GREAT  
Dr. Lic. = ND/DIS821456  
Lot No = 23411080A1  
Cyl No = 71  
Expiration Date = 10/01/2013  
County = 08                            Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks: *DIFFERENCE TOO GREAT*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005939  
Location = TOXL                8164.13.00 06/09  
09/19/2013                      11:17

Test	AC	Time
01 Room Air	0.000	11:17
02 Std. Gas	0.081	11:18
03 Room Air	0.000	11:18
04 Std. Gas	0.081	11:19
05 Room Air	0.000	11:19
06 Std. Gas	0.080	11:19
07 Room Air	0.000	11:20
08 Std. Gas	0.082	11:20
09 Room Air	0.000	11:21
10 Std. Gas	0.081	11:21
11 Room Air	0.000	11:22
12 Std. Gas	0.081	11:22
13 Room Air	0.000	11:23
14 Std. Gas	0.081	11:23
15 Room Air	0.000	11:24
16 Std. Gas	0.080	11:24
17 Room Air	0.000	11:24
18 Std. Gas	0.080	11:25
19 Room Air	0.000	11:25
20 Std. Gas	0.082	11:26
21 Room Air	0.000	11:26
22 Std. Gas	0.080	11:26
23 Room Air	0.000	11:27
24 Std. Gas	0.082	11:27
25 Room Air	0.000	11:28
26 Std. Gas	0.081	11:28
27 Room Air	0.000	11:29
28 Std. Gas	0.082	11:29
29 Room Air	0.000	11:30
30 Std. Gas	0.082	11:30
31 Room Air	0.000	11:31
32 Std. Gas	0.081	11:31
33 Room Air	0.000	11:31
34 Std. Gas	0.082	11:32
35 Room Air	0.000	11:32
36 Std. Gas	0.082	11:33
37 Room Air	0.000	11:33
38 Std. Gas	0.082	11:33
39 Room Air	0.000	11:34
40 Std. Gas	0.081	11:34
41 Room Air	0.000	11:35
42 Cal Check Stats:		
43 Average	0.0812	
44 Std Dev	0.0008	
45 Rel Std Dev	0.95%	

STABILITY TEST  
Charles E. Eide  
pg. 1 of 2

Operator Signature  
CHARLES EDER

Remarks:

STABILITY TEST

Form 106-I8000

LOT #: 23411080A1

Cyl #: 071

EXP. DATE: 10/1/2013

0.080 AC

STABILITY TEST  
Charles Eder  
pg. 2 of