

INTOXILYZER® 8000 INITIAL INSPECTION

Serial Number: 80-005940

1. Power On and Warm-up:

- a. Power On
- b. 5 min. Countdown from Initial Warm-up
- c. 1 min. Countdown to Begin a Test
- d. Forms Load Complete

2. Setup and Checks:

a. Level 1:

- 1) Passwords: Field Inspector (Level 2) Lab Personnel (Level 3)
- 2) Gas Cylinder: Gas Cylinder Setup (Level 1,S)
- 3) Printer: Printer Setup (Level 3,S,P)
 Internal Printer Paper Feed
 Internal Printer Test (Level 1,P)
 External Printer Test (Level 1,P)
- 4) Barcode Reader (Level 1,B): Operator ID
 ND License

b. Level 2:

- 1) Continuous Purge (Level 2,A)
- 2) Diagnostic (Level 2,D)
- 3) Prelim Data Entry (Level 2,E)
- 4) Ser Num/Version Info (Level 2,V): Ser Num 80-005940
Version 8164.13

c. Level 3:

- 1) DVM Monitor (Level 3,D,D): 3: 12750 s: 0 c: 47.00
9: 13245 s: 3 B: 45 0
(RFI indicator with keyed radio) 1

- 2) Function (F) Keys: F5 – Tone
 F6 – Green LED
 F7 – Red LED
 F9 – Fan
 F10 – Pump
 F11 – Solenoid

- 3) Atmospheric Monitor (Level 3,D,A): Atmo Sensor, P: 953 CF: 1.06
- 4) Flow Monitor (Level 3,D,F): Flow Sensor, R: 220 F: 0.000 L/s

5) Temperature Monitor (Level 3,D,T): C: 47.0 A: 47.5 I: 0.0
B: 45.0 V: 0.0

6) Tank Monitor (Level 3,D,G): P (gauge): ~50 P (display): 42
7) General Setup (Level 3,S,G)
8) Configure Start Test (Level 3,S,T)
9) Network (Level 3,S,N)

3. Tests:

a. Configure Standard for Wet Bath Simulator (Level 1,S):

1) Zero Alcohol Test (Level 1,C):
Sim Ser #: DR5188 Lot #: MILLI Q H₂O Exp Date: NA
Known Value: 0.000 AC
Results: 0.000 AC 0.000 AC 0.000 AC

2) Low AC Test (Level 1,C):
Sim Ser #: DR3455 Lot #: 12210 Exp Date: 11-2-14
Known Value: 0.020 AC
Results: 0.018 AC 0.019 AC 0.019 AC

3) Linearity Test (Level 1,C):
Sim Ser #: DR3847 Lot #: 13020 Exp Date: 1-15-15
Known Value: 0.300 AC
Results: 0.297 AC 0.299 AC 0.298 AC

b. Configure Dry Gas Standard (Level 1,S):

1) Interferent Detected Tests (Level 1,B):

1. 0.10 AC + 0.05% Acetone
Sim Ser #: DR3377 Lot #: ICS#1 Exp Date: 4-9-14
Known Value: 0.10AC + 0.05% ACETONE
 Display reads "Interferent Detect"

2. Acetone
Sim Ser #: G8919 Lot #: ACE1 Exp Date: 8-27-15
Known Value: 0.05% ACETONE
 Display reads "Interferent Detect"

3. Methanol
Sim Ser #: G10573 Lot #: MET1 Exp Date: 8-27-15
Known Value: 0.10% METHANOL
 Display reads "Interferent Detect"

4. Isopropanol
Sim Ser #: G9998 Lot #: IPA1 Exp Date: 8-27-15
Known Value: 0.10% ISOPROPRANOL
 Display reads "Interferent Detect"
- 2) Ambient Fail (Level 1,B): Display and test record are correct
3) Invalid Sample (Level 1,B): Display and test record are correct
4) Subject Test Timeout (Level 1,B): Display and test record are correct.
5) Sequence Aborted (Level 1,B): Display and test record are correct
6) Control Outside Tolerance (CMS mode): Display and test record are correct
7) RFI Detected Test (CMS mode): Display and test record are correct
8) Tank Pressure Below Min (CMS mode): Display reads correct
9) Subject Refused (CMS mode): Display and test record are correct
10) Difference Too Great (CMS mode): Display and test record are correct
11) Stability Test (Level 3,T,S):
Average 0.0801 Std Dev 0.0004 Rel Std Dev 0.56%

Maintenance/Repairs/Notes: _____

THIS UNIT IS ACCEPTABLE TO BE USED IN THE FIELD.


Field Inspector's Signature

19 SEP 13
Date

Deb Kashner
Reviewed by

19 Sept 2013
Date

Intoxilyzer Test Record and Checklist
NDDAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005940
Location = TOXL 8164.13.00 06/09
09/19/2013 12:46

***** Printer Test *****

12 char/line
abcdefghijkl
ABCDEFGHIJKL

16 char per line
abcdefghijklmnop
ABCDEFGHIJKLMN

21 chars per line
abcdefghijklmnopqrstu
ABCDEFGHIJKLMNQRSTU

24 characters per line
abcdefghijklmnopqrstuvw
ABCDEFGHIJKLMNQRSTUW

32 characters per line
abcdefghijklmnopqrstuvwxy123456
ABCDEFGHIJKLMNQRSTUWXY123456

42 characters per line
abcdefghijklmnopqrstuvwxy1234567890abcdef
ABCDEFGHIJKLMNQRSTUWXY1234567890ABCDEF

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? No
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 23411080A1
Standard Cyl #? 71
Standard Expiration? 10/01/2013
Oper No? 666666

Flow Cal. Date: 06/21/2013
Slope 688
Intercept -659053

IR Calibration Date: 06/22/2013
3um 9um

0th Coef(*100):	-26055	-23570
1st Coef(*100):	280441	140486
2nd Coef(*100):	1298	1124
H2O adj(mg/l*10K):	422	422

***** Test End *****



Operator Signature
CHARLES EDER

Remarks: **INTERNAL
PRINT
TEST**
Form 106-18000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005940
Location = TOXL 8164.13.00 06/09
09/19/2013 12:47

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? No
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 23411080A1
Standard Cyl #? 71
Standard Expiration? 10/01/2013
Oper No? 666666

Flow Cal. Date: 06/21/2013
Slope 688
Intercept -659053

IR Calibration Date: 06/22/2013
 3um 9um

0th Coef(*100): -26055 -23570
1st Coef(*100): 280441 140486
2nd Coef(*100): 1298 1124
H2O adj(mg/l*10k): 422 422

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

EXTERNAL PRINT
TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005940
Location = TOXL 8164.13.00 06/09
09/19/2013 13:01

DIAGNOSTIC

Voltage/Current Test	Pass
RAM Test	Pass
EEPROM Checksum Test	Pass
Real Time Clock Test	Pass
DSP Test	Pass
Analytical Stability Test	Pass
Modem Test	Pass
Temp Regulation Test	Pass

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *DIAGNOSTIC CHECK*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005940
Location = TOXL 8164.13.00 06/09
09/19/2013 13:05

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:05
02 Std. Sol.	0.000	13:06
03 Room Air	0.000	13:07
04 Std. Sol.	0.000	13:07
05 Room Air	0.000	13:08
06 Std. Sol.	0.000	13:09
07 Room Air	0.000	13:09

08 Sim Temp = 34.0°C

Simul Ser No = DR5188
Std Sol No = MILLIQ H2O
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: **ZERO ALCOHOL TEST**

0.000AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005940
Location = TOXL 8164.13.00 06/09
09/19/2013 13:10

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:11
02 Std. Sol.	0.018	13:12
03 Room Air	0.000	13:12
04 Std. Sol.	0.019	13:13
05 Room Air	0.000	13:14
06 Std. Sol.	0.019	13:14
07 Room Air	0.000	13:15

08 Sim Temp = 34.0°C

Simul Ser No = DR3455

Std Sol No = 12210

County = 08

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

Low AC TEST

0.020 AC Sim. Sol.

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005940
Location = TOXL 8164.13.00 06/09
09/19/2013 13:18

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:19
02 Std. Sol.	0.297	13:19
03 Room Air	0.000	13:20
04 Std. Sol.	0.299	13:21
05 Room Air	0.000	13:21
06 Std. Sol.	0.298	13:22
07 Room Air	0.000	13:22

08 Sim Temp = 34.0°C

Simul Ser No = DR3847
Std Sol No = 13020
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *LINEARITY TEST*
0.300AC Sim. Sol.

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005940
Location = TOXL 8164.13.00 06/09
09/19/2013 13:30

Test	AC	Time
01 Room Air	0.000	13:31
02 *Subject Test	INT*	13:31
03 Room Air	0.000	13:32

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = 0.10AC+0.05ACE
Dr. Lic. = ND/DIS821456
Lot No = 23411080A1
Cyl No = 71
Expiration Date = 10/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *INTERFERENT DETECTED CHECK*
0.10AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005940
Location = TOXL 8164.13.00 06/09
09/19/2013 13:32

Test	AC	Time
01 Room Air	0.000	13:33
02 *Subject Test	INT*	13:34
03 Room Air	0.000	13:34

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = 0.05 ACETONE
Dr. Lic. = ND/DIS821456
Lot No = 23411080A1
Cyl No = 71
Expiration Date = 10/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: **INTERFERENT DETECTED CHECK
0.05% ACETONE**

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005940
Location = TOXL 8164.13.00 06/09
09/19/2013 13:34

Test	AC	Time
01 Room Air	0.000	13:35
02 *Subject Test	INT*	13:36
03 Room Air	0.000	13:37

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = 0.10 METHANOL
Dr. Lic. = ND/DIS821456
Lot No = 23411080A1
Cyl No = 71
Expiration Date = 10/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *INTERFERENT DETECTED CHECK*
0.10% METHANOL

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005940
Location = TOXL 8164.13.00 06/09
09/19/2013 13:37

Test	AC	Time
01 Room Air	0.000	13:38
02 *Subject Test	INT*	13:38
03 Room Air	0.000	13:39

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = 0.10 IPA
Dr. Lic. = ND/DIS821456
Lot No = 23411080A1
Cyl No = 71
Expiration Date = 10/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *INTERFERENT DETECTED CHECK*
0.10% ISOPROPANOL

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005940
Location = TOXL 8164.13.00 06/09
09/19/2013 13:40

Test	AC	Time
01 Room Air	AMB*	13:41
02 Room Air	0.000	13:41

*Invalid Test
Check Ambient Conditions

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = AMBIENT FAIL
Dr. Lic. = ND/DIS821456
Lot No = 23411080A1
Cyl No = 71
Expiration Date = 10/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: **AMBIENT FAIL CHECK**

Intoxilyzer Test Record and Checklist
 NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
 North Dakota Model 8000 SN 80-005940
 Location = TOXL 8164.13.00 06/09
 09/19/2013 13:42

Test	AC	Time
01 Room Air	0.000	13:43
02 Invalid Sample	X.XXX	13:43
03 Room Air	0.000	13:44

*Invalid Test - Mouth Alcohol

Sub Name = DISCOVER, THE SPIRIT
 Sub DOB = 01/01/1982
 Sub Sex = Female Weight = 150
 Test = DUI Cit = INVALID SAMPLE
 Dr. Lic. = ND/DIS821456
 Lot No = 23411080A1
 Cyl No = 71
 Expiration Date = 10/01/2013
 County = 08 Oper No. = 666666



Operator Signature
 CHARLES EDER

Remarks: **INVALID SAMPLE**

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005940
Location = TOXL 8164.13.00 06/09
09/19/2013 13:45

Test	AC	Time
01 Room Air	0.000	13:46
02 *Subject Test	0.000*	13:49
03 Room Air	0.000	13:50
No RFI Detected		

*Deficient Sample - Value Printed was
Highest Obtained

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = TEST TIMEOUT
Dr. Lic. = ND/DIS821456
Lot No = 23411080A1
Cyl No = 71
Expiration Date = 10/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *SUBJECT TEST TIMEOUT*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005940
Location = TOXL 8164.13.00 06/09
09/19/2013 13:51

Test	AC	Time
01 Room Air	0.000	13:51
02 *Subject Test	ABT*	13:52
03 Room Air	0.000	13:52

*Sequence Aborted

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = SEQUENCE ABORT
Dr. Lic. = ND/DIS821456
Lot No = 23411080A1
Cyl No = 71
Expiration Date = 10/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *SEQUENCE ABORTED*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005940
Location = TOXL 8164.13.00 06/09
09/19/2013 13:53

Test	AC	Time
01 Diagnostic	OK	13:54
02 Room Air	0.000	13:54
03 Subject Test 1	0.042	13:55
04 Room Air	0.000	13:57
05 Std. Gas	0.086*	13:58
06 Room Air	0.000	13:59

*Cal Check Out of Tolerance

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = CON OUT TOL
Dr. Lic. = ND/DIS821456
Lot No = 23411080A1
Cyl No = 71
Expiration Date = 10/01/2013
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: **CONTROL OUTSIDE TOLERANCE**

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005940
Location = TOXL 8164.13.00 06/09
09/19/2013 14:00

Test	AC	Time
01 Diagnostic	OK	14:01
02 Room Air	0.000	14:01
03 *Subject Test	RFI*	14:01
04 Room Air	0.000	14:02

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 23411080A1
Cyl No = 71
Expiration Date = 10/01/2013
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: **RFI CHECK**

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005940
Location = TOXL 8164.13.00 06/09
09/19/2013 14:05

Test	AC	Time
01 Diagnostic	OK	14:06
02 Room Air	0.000	14:07
03 *Subject Test	REF*	14:07
04 Room Air	0.000	14:08

*Subject Test Refused

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = SUBJECT REFUSED
Dr. Lic. = ND/DIS821456
Lot No = 23411080A1
Cyl No = 71
Expiration Date = 10/01/2013
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: *SUBJECT REFUSED*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005940
Location = TOXL 8164.13.00 06/09
09/19/2013 14:09

Test	AC	Time
01 Diagnostic	OK	14:10
02 Room Air	0.000	14:10
03 Subject Test 1	0.042*	14:11
04 Room Air	0.000	14:13
05 Std. Gas	0.080	14:14
06 Room Air	0.000	14:15
07 Subject Test 2	0.099*	14:16
08 Room Air	0.000	14:17

*Difference Too Great

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = DIFF TOO GREAT
Dr. Lic. = ND/DIS821456
Lot No = 23411080A1
Cyl No = 71
Expiration Date = 10/01/2013
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

DIFFERENCE TOO GREAT CHECK

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005940
Location = TOXL 8164.13.00 06/09
09/19/2013 14:41

Test	AC	Time
01 Room Air	0.000	14:42
02 Std. Gas	0.080	14:42
03 Room Air	0.000	14:43
04 Std. Gas	0.080	14:43
05 Room Air	0.000	14:44
06 Std. Gas	0.080	14:44
07 Room Air	0.000	14:44
08 Std. Gas	0.080	14:45
09 Room Air	0.000	14:45
10 Std. Gas	0.081	14:46
11 Room Air	0.000	14:46
12 Std. Gas	0.080	14:46
13 Room Air	0.000	14:47
14 Std. Gas	0.080	14:47
15 Room Air	0.000	14:48
16 Std. Gas	0.080	14:48
17 Room Air	0.000	14:49
18 Std. Gas	0.080	14:49
19 Room Air	0.000	14:50
20 Std. Gas	0.080	14:50
21 Room Air	0.000	14:50
22 Std. Gas	0.081	14:51
23 Room Air	0.000	14:51
24 Std. Gas	0.080	14:52
25 Room Air	0.000	14:52
26 Std. Gas	0.081	14:53
27 Room Air	0.000	14:53
28 Std. Gas	0.080	14:53
29 Room Air	0.000	14:54
30 Std. Gas	0.080	14:54
31 Room Air	0.000	14:55
32 Std. Gas	0.080	14:55
33 Room Air	0.000	14:56
34 Std. Gas	0.080	14:56
35 Room Air	0.000	14:56
36 Std. Gas	0.079	14:57
37 Room Air	0.000	14:57
38 Std. Gas	0.080	14:58
39 Room Air	0.000	14:58
40 Std. Gas	0.080	14:59
41 Room Air	0.000	14:59
42 Cal Check Stats:		
43 Average	0.0801	
44 Std Dev	0.0004	
45 Rel Std Dev	0.56%	

STABILITY
TEST
Charles E. Edwards
pg 1 of 2

Operator Signature
CHARLES EDER

Remarks:

STABILITY TEST

Form 106-I8000

LOT#: 23411080A1

Cyl #: 047

EXP. DATE: 10/1/2013

0.080AC

STABILITY TEST
Charles Eder
pg 2 of 2