

**INTOXILYZER® 8000 INITIAL INSPECTION**

Serial Number: 80-005952

1. Power On and Warm-up:

- a.  Power On
- b.  5 min. Countdown from Initial Warm-up
- c.  1 min. Countdown to Begin a Test
- d.  Forms Load Complete

2. Setup and Checks:

- a. Level 1:
  - 1) Passwords:  Field Inspector (Level 2)  Lab Personnel (Level 3)
  - 2) Gas Cylinder:  Gas Cylinder Setup (Level 1,S)
  - 3) Printer:  Printer Setup (Level 3,S,P)
    - Internal Printer Paper Feed
    - Internal Printer Test (Level 1,P)
    - External Printer Test (Level 1,P)
  - 4) Barcode Reader (Level 1,B):  Operator ID  
 ND License

b. Level 2:

- 1)  Continuous Purge (Level 2,A)
- 2)  Diagnostic (Level 2,D)
- 3)  Prelim Data Entry (Level 2,E)
- 4)  Ser Num/Version Info (Level 2,V): Ser Num 80-005952  
Version 8164.13

c. Level 3:

- 1) DVM Monitor (Level 3,D,D): 3: 12458 s: 1 C: 47.00  
g: 13398 s: 1 B: 44 0  
(RFI indicator with keyed radio) 1
- 2) Function (F) Keys:  F5 – Tone  
 F6 – Green LED  
 F7 – Red LED  
 F9 – Fan  
 F10 – Pump  
 F11 – Solenoid
- 3) Atmospheric Monitor (Level 3,D,A): Atmo Sensor, P: 953 CF: 1.06
- 4) Flow Monitor (Level 3,D,F): Flow Sensor, R: 219 F: 0.000 L/s

5) Temperature Monitor (Level 3,D,T): C: 47.0 A: 43.0 I: 0.0  
B: 45.0 V: 0.0

6) Tank Monitor (Level 3,D,G): P (gauge): ~900 P (display): 885  
7)  General Setup (Level 3,S,G)  
8)  Configure Start Test (Level 3,S,T)  
9)  Network (Level 3,S,N)

3. Tests:

a. Configure Standard for Wet Bath Simulator (Level 1,S):

- 1) Zero Alcohol Test (Level 1,C):  
Sim Ser #: DR5188 Lot #: MILLI Q H<sub>2</sub>O Exp Date: NA  
Known Value: 0.000 AC  
Results: 0.000 AC 0.000 AC 0.000 AC
- 2) Low AC Test (Level 1,C):  
Sim Ser #: DR3455 Lot #: 12210 Exp Date: 11-2-14  
Known Value: 0.020 AC  
Results: 0.019 AC 0.019 AC 0.020 AC
- 3) Linearity Test (Level 1,C):  
Sim Ser #: DR3847 Lot #: 13020 Exp Date: 1-15-15  
Known Value: 0.300 AC  
Results: 0.298 AC 0.300 AC 0.299 AC

b. Configure Dry Gas Standard (Level 1,S):

- 1) Interferent Detected Tests (Level 1,B):
1. 0.10 AC + 0.05% Acetone  
Sim Ser #: DR3377 Lot #: ICS#1 Exp Date: 4-9-14  
Known Value: 0.10AC + 0.05% ACETONE  
 Display reads "Interferent Detect"
2. Acetone  
Sim Ser #: G8919 Lot #: ACE1 Exp Date: 8-27-15  
Known Value: 0.05% ACETONE  
 Display reads "Interferent Detect"
3. Methanol  
Sim Ser #: G10573 Lot #: MET1 Exp Date: 8-27-15  
Known Value: 0.10% METHANOL  
 Display reads "Interferent Detect"

4. Isopropanol  
Sim Ser #: 69998 Lot #: IPA1 Exp Date: 8-27-15  
Known Value: 0.10% ISOPROPANOL  
 Display reads "Interferent Detect"
- 2)  Ambient Fail (Level 1,B): Display and test record are correct  
3)  Invalid Sample (Level 1,B): Display and test record are correct  
4)  Subject Test Timeout (Level 1,B): Display and test record are correct.  
5)  Sequence Aborted (Level 1,B): Display and test record are correct  
6)  Control Outside Tolerance (CMS mode): Display and test record are correct  
7)  RFI Detected Test (CMS mode): Display and test record are correct  
8)  Tank Pressure Below Min (CMS mode): Display reads correct  
9)  Subject Refused (CMS mode): Display and test record are correct  
10)  Difference Too Great (CMS mode): Display and test record are correct  
11)  Stability Test (Level 3,T,S):  
Average 0.0812 Std Dev 0.0007 Rel Std Dev 0.83%

Maintenance/Repairs/Notes: \_\_\_\_\_

THIS UNIT IS ACCEPTABLE TO BE USED IN THE FIELD.

Charles E. Eden  
Field Inspector's Signature

19 SEP 13  
Date

Deb Kashum  
Reviewed by

19 Sept 2013  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer  
North Dakota Model 8000 SN 80-005952  
Location = TOXL 8164.13.00 06/09  
09/19/2013 14:14

\*\*\*\*\* Printer Test \*\*\*\*\*

12 char/line  
abcdefghijkl  
ABCDEFGHIJKL

16 char per line  
abcdefghijklmnop  
ABCDEFGHIJKLMN

21 chars per line  
abcdefghijklmnopqrstu  
ABCDEFGHIJKLMNQRSTU

24 characters per line  
abcdefghijklmnopqrstuwx  
ABCDEFGHIJKLMNQRSTUWX

32 characters per line  
abcdefghijklmnopqrstuwx123456  
ABCDEFGHIJKLMNQRSTUWXYZ123456

42 characters per line  
abcdefghijklmnopqrstuwx1234567890abcdef  
ABCDEFGHIJKLMNQRSTUWXYZ1234567890ABCDEF

Current Instrument Setup

Data Entry Mode: Enabled  
Start Test Sequence: DABACABA  
Display Prelim Rslt? Yes  
Display Third Digit? Yes  
Inhib Printer(Y/N)? No  
Display Volume? No  
Disable On MemFull? Yes  
# of Print Copies? 1  
Select Std (D/W/L)? Dry  
Standard Value? 0.080  
Standard Lot #? 23411080A1  
Standard Cyl #? 47  
Standard Expiration? 10/01/2013  
Oper No? 666666

Flow Cal. Date: 06/21/2013  
Slope: 734  
Intercept: -865305

IR Calibration Date: 06/21/2013  
3um 9um

0th Coef(*100):	-21281	-23057
1st Coef(*100):	292132	137092
2nd Coef(*100):	2334	1231
H2O adj(mg/l*10K):	497	482

\*\*\*\*\* Test End \*\*\*\*\*



Operator Signature  
CHARLES EDER

Remarks: **INTERNAL  
PRINT  
TEST**

Form 106-18000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005952  
Location = TOXL                    8164.13.00 06/09  
09/19/2013                            14:23

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&\*()\_+?

Current Instrument Setup

Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?          1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            23411080A1  
Standard Cyl #?            47  
Standard Expiration?       10/01/2013  
Oper No?                    666666

Flow Cal. Date:            06/21/2013  
Slope                        734  
Intercept                   -865305

IR Calibration Date:        06/21/2013

	3um	9um
0th Coef(*100):	-21281	-23057
1st Coef(*100):	292132	137092
2nd Coef(*100):	2334	1231
H2O adj(mg/l*10k):	497	482

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
CHARLES EDER

Remarks:    **EXTERNAL PRINT  
TEST**

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005952  
Location = TOXL                    8164.13.00 06/09  
09/19/2013                            14:24

DIAGNOSTIC

Voltage/Current Test	Pass
RAM Test	Pass
EEPROM Checksum Test	Pass
Real Time Clock Test	Pass
DSP Test	Pass
Analytical Stability Test	Pass
Modem Test	Pass
Temp Regulation Test	Pass

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *DIAGNOSTIC CHECK*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005952  
Location = TOXL                    8164.13.00 06/09  
09/19/2013                            15:15

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:15
02 Std. Sol.	0.000	15:16
03 Room Air	0.000	15:16
04 Std. Sol.	0.000	15:17
05 Room Air	0.000	15:18
06 Std. Sol.	0.000	15:18
07 Room Air	0.000	15:19

08 Sim Temp = 34.0°C

Simul Ser No = DR5188  
Std Sol No = MILLIQ H2O  
County = 08                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: **ZERO ALCOHOL TEST**  
**0.000 AC**

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005952  
Location = TOXL                    8164.13.00 06/09  
09/19/2013                            15:21

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:21
02 Std. Sol.	0.019	15:22
03 Room Air	0.000	15:23
04 Std. Sol.	0.019	15:23
05 Room Air	0.000	15:24
06 Std. Sol.	0.020	15:25
07 Room Air	0.000	15:25

08 Sim Temp = 34.0°C

Simul Ser No = DR3455  
Std Sol No = 12210  
County = 08                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *Low AC TEST*  
*0.020 AC Sim. Sol.*  
Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005952  
Location = TOXL                 8164.13.00 06/09  
09/19/2013                        15:28

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:29
02 Std. Sol.	0.298	15:29
03 Room Air	0.000	15:30
04 Std. Sol.	0.300	15:31
05 Room Air	0.000	15:31
06 Std. Sol.	0.299	15:32
07 Room Air	0.000	15:33

08 Sim Temp = 34.0°C

Simul Ser No = DR3847  
Std Sol No = 13020  
County = 08    Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *LINEARITY TEST*  
*0.300AC Sim. Sol.*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005952  
Location = TOXL                    8164.13.00 06/09  
09/19/2013                            14:31

Test	AC	Time
01 Room Air	0.000	14:32
02 *Subject Test	INT*	14:33
03 Room Air	0.000	14:33

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = 0.10AC+0.05ACE  
Dr. Lic. = ND/DIS821456  
Lot No = 23411080A1  
Cyl No = 47  
Expiration Date = 10/01/2013  
County = 08                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *INTERFERENT DETECTED  
CHECK  
0.10AC + 0.05% ACETONE*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer                      Alcohol Analyzer  
North Dakota Model 8000                      SN 80-005952  
Location = TOXL                              8164.13.00 06/09  
09/19/2013    14:36

Test	AC	Time
01 Room Air	0.000	14:37
02 *Subject Test	INT*	14:38
03 Room Air	0.000	14:38

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                              Weight = 150  
Test = DUI                                      Cit = ~~0.10~~ ACETONE  
Dr. Lic. = ND/DIS821456                      0.05% CEE  
Lot No = 23411080A1  
Cyl No = 47  
Expiration Date = 10/01/2013  
County = 08                                      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: INTERFERENT DETECTED CHECK  
CEE  
0.05% ~~0.10%~~ ACETONE

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005952  
Location = TOXL                8164.13.00 06/09  
09/19/2013                        14:47

Test	AC	Time
01 Room Air	0.000	14:48
02 *Subject Test	INT*	14:49
03 Room Air	0.000	14:49

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = 0.10 METHANOL  
Dr. Lic. = ND/DIS821456  
Lot No = 23411080A1  
Cyl No = 47  
Expiration Date = 10/01/2013  
County = 08                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *INTERFERENT DETECTED CHECK*  
*0.10% METHANOL*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005952  
Location = TOXL                8164.13.00 06/09  
09/19/2013                        14:51

Test	AC	Time
01 Room Air	0.000	14:51
02 *Subject Test	INT*	14:52
03 Room Air	0.000	14:53

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = 0.10 IPA  
Dr. Lic. = ND/DIS821456  
Lot No = 23411080A1  
Cyl No = 47  
Expiration Date = 10/01/2013  
County = 08                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: **INTERFERENT DETECTED CHECK**  
**0.10% ISOPROPANOL**

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer          Alcohol Analyzer  
North Dakota Model 8000          SN 80-005952  
Location = TOXL                  8164.13.00 06/09  
09/19/2013                                  15:36

Test	AC	Time
01 Room Air	AMB*	15:37
02 Room Air	0.000	15:38

\*Invalid Test  
Check Ambient Conditions

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                  Weight = 150  
Test = DUI                          Cit = AMBIENT FAIL  
Dr. Lic. = ND/DIS821456  
Lot No = 23411080A1  
Cyl No = 47  
Expiration Date = 10/01/2013  
County = 08                          Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *AMBIENT FAIL CHECK*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005952  
Location = TOXL                    8164.13.00 06/09  
09/19/2013                            15:39

Test	AC	Time
01 Room Air	0.000	15:40
02 Invalid Sample	X.XXX	15:40
03 Room Air	0.000	15:41

\*Invalid Test - Mouth Alcohol

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = INVALID SAMPLE  
Dr. Lic. = ND/DIS821456  
Lot No = 23411080A1  
Cyl No = 47  
Expiration Date = 10/01/2013  
County = 08                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *INVALID SAMPLE CHECK*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005952  
Location = TOXL                8164.13.00 06/09  
09/19/2013                        15:41

Test	AC	Time
01 Room Air	0.000	15:42
02 *Subject Test	0.000*	15:45
03 Room Air	0.000	15:45

No RFI Detected

\*Deficient Sample - Value Printed was  
Highest Obtained

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = TEST TIMEOUT  
Dr. Lic. = ND/DIS821456  
Lot No = 23411080A1  
Cyl No = 47  
Expiration Date = 10/01/2013  
County = 08                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *SUBJECT TEST TIMEOUT*



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005952  
Location = TOXL                    8164.13.00 06/09  
09/19/2013                            15:46

Test	AC	Time
01 Room Air	0.000	15:46
02 *Subject Test	ABT*	15:47
03 Room Air	0.000	15:47

\*Sequence Aborted

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = SEQUENCE ABORT  
Dr. Lic. = ND/DIS821456  
Lot No = 23411080A1  
Cyl No = 47  
Expiration Date = 10/01/2013  
County = 08                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: **SEQUENCE ABORTED**  
**CHECK**

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer           Alcohol Analyzer  
North Dakota Model 8000       SN 80-005952  
Location = TOXL                8164.13.00 06/09  
09/19/2013                      15:48

Test	AC	Time
01 Diagnostic	OK	15:49
02 Room Air	0.000	15:49
03 Subject Test 1	0.042	15:50
04 Room Air	0.000	15:52
05 Std. Gas	0.088*	15:53
06 Room Air	0.000	15:54

\*Cal Check Out of Tolerance

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                   Weight = 150  
Test = DUI                        Cit = CON OUT TOL  
Dr. Lic. = ND/DIS821456  
Lot No = 23411080A1  
Cyl No = 47  
Expiration Date = 10/01/2013  
County = 08                      Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks: *CONTROL OUTSIDE TOLERANCE*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer                      Alcohol Analyzer  
North Dakota Model 8000                      SN 80-005952  
Location = TOXL                              8164.13.00 06/09  
09/19/2013    15:56

Test	AC	Time
01 Diagnostic	OK	15:57
02 Room Air	RFI*	15:57
03 Room Air	0.000	15:58

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                              Weight = 150  
Test = DUI                                      Cit = RFI CHECK  
Dr. Lic. = ND/DIS821456  
Lot No = 23411080A1  
Cyl No = 47  
Expiration Date = 10/01/2013  
County = 08                                      Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks: *RFI CHECK*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005952  
Location = TOXL                    8164.13.00 06/09  
09/19/2013                            16:00

Test	AC	Time
01 Diagnostic	OK	16:01
02 Room Air	0.000	16:01
03 Subject Test 1	0.098	16:02
04 Room Air	0.000	16:04
05 Std. Gas	0.082	16:05
06 Room Air	0.000	16:06
07 *Subject Test	REF*	16:06
08 Room Air	0.000	16:07

\*Subject Test Refused

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = SUBJECT REFUSED  
Dr. Lic. = ND/DIS821456  
Lot No = 23411080A1  
Cyl No = 47  
Expiration Date = 10/01/2013  
County = 08                            Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks: **SUBJECT REFUSED  
CHECK**

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005952  
Location = TOXL                    8164.13.00 06/09  
09/19/2013                            16:08

Test	AC	Time
01 Diagnostic	OK	16:09
02 Room Air	0.000	16:09
03 Subject Test 1	0.043*	16:10
04 Room Air	0.000	16:12
05 Std. Gas	0.082	16:13
06 Room Air	0.000	16:14
07 Subject Test 2	0.097*	16:15
08 Room Air	0.000	16:16

\*Difference Too Great

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = DIFF TOO GREAT  
Dr. Lic. = ND/DIS821456  
Lot No = 23411080A1  
Cyl No = 47  
Expiration Date = 10/01/2013  
County = 08                            Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks: *DIFFERENCE TOO GREAT*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005952  
Location = TOXL                 8164.13.00 06/09  
09/19/2013                        16:18

Test	AC	Time
01 Room Air	0.000	16:19
02 Std. Gas	0.081	16:19
03 Room Air	0.000	16:20
04 Std. Gas	0.080	16:20
05 Room Air	0.000	16:21
06 Std. Gas	0.081	16:21
07 Room Air	0.000	16:21
08 Std. Gas	0.081	16:22
09 Room Air	0.000	16:22
10 Std. Gas	0.081	16:23
11 Room Air	0.000	16:23
12 Std. Gas	0.081	16:23
13 Room Air	0.000	16:24
14 Std. Gas	0.081	16:24
15 Room Air	0.000	16:25
16 Std. Gas	0.083	16:25
17 Room Air	0.000	16:26
18 Std. Gas	0.081	16:26
19 Room Air	0.000	16:27
20 Std. Gas	0.082	16:27
21 Room Air	0.000	16:27
22 Std. Gas	0.081	16:28
23 Room Air	0.000	16:28
24 Std. Gas	0.081	16:29
25 Room Air	0.000	16:29
26 Std. Gas	0.082	16:30
27 Room Air	0.000	16:30
28 Std. Gas	0.081	16:30
29 Room Air	0.000	16:31
30 Std. Gas	0.082	16:31
31 Room Air	0.000	16:32
32 Std. Gas	0.081	16:32
33 Room Air	0.000	16:33
34 Std. Gas	0.081	16:33
35 Room Air	0.000	16:34
36 Std. Gas	0.081	16:34
37 Room Air	0.000	16:34
38 Std. Gas	0.081	16:35
39 Room Air	0.000	16:35
40 Std. Gas	0.080	16:36
41 Room Air	0.000	16:36
42 Cal Check Stats:		
43 Average	0.0812	
44 Std Dev	0.0007	
45 Rel Std Dev	0.83%	

STABILITY TEST  
Charles E. Ehr  
pg 1 of 2

Operator Signature  
CHARLES EDER

Remarks:

STABILITY TEST

Form 106-I8000

LOT#: 2341108041

Cyl #: 047

Exp DATE: 10/1/2013

0.080 AC

STABILITY TEST  
Charles Eder  
pg 2 of 2