

INTOXILYZER® 8000 INITIAL INSPECTION

Serial Number: 80-005937

1. Power On and Warm-up:

- a. Power On
- b. 5 min. Countdown from Initial Warm-up
- c. 1 min. Countdown to Begin a Test
- d. Forms Load Complete

2. Setup and Checks:

a. Level 1:

- 1) Passwords: Field Inspector (Level 2) Lab Personnel (Level 3)
- 2) Gas Cylinder: Gas Cylinder Setup (Level 1,S)
- 3) Printer: Printer Setup (Level 3,S,P)
 Internal Printer Paper Feed
 Internal Printer Test (Level 1,P)
 External Printer Test (Level 1,P)
- 4) Barcode Reader (Level 1,B): Operator ID
 ND License

b. Level 2:

- 1) Continuous Purge (Level 2,A)
- 2) Diagnostic (Level 2,D)
- 3) Prelim Data Entry (Level 2,E)
- 4) Ser Num/Version Info (Level 2,V): Ser Num 80-005937
Version 8164.13

c. Level 3:

- 1) DVM Monitor (Level 3,D,D): 3: 12508 s: 1 c: 47.00
g: 13562 s: 2 B: 45 0
(RFI indicator with keyed radio) 1

- 2) Function (F) Keys: F5 – Tone
 F6 – Green LED
 F7 – Red LED
 F9 – Fan
 F10 – Pump
 F11 – Solenoid

- 3) Atmospheric Monitor (Level 3,D,A): Atmo Sensor, P: 957 CF: 1.05
- 4) Flow Monitor (Level 3,D,F): Flow Sensor, R: 207 F: 0.000 L/s

5) Temperature Monitor (Level 3,D,T): C: 47.0 A: 49.0 I: 0.0

B: 45.0 V: 0.0

6) Tank Monitor (Level 3,D,G): P (gauge): ~700 P (display): 691

7) General Setup (Level 3,S,G)

8) Configure Start Test (Level 3,S,T)

9) Network (Level 3,S,N)

3. Tests:

a. Configure Standard for Wet Bath Simulator (Level 1,S):

1) Zero Alcohol Test (Level 1,C):

Sim Ser #: DR5188 Lot #: MILLIQH₂O Exp Date: NA

Known Value: 0.000 AC

Results: 0.000 AC 0.000 AC 0.000 AC

2) Low AC Test (Level 1,C):

Sim Ser #: DR3455 Lot #: 12210 Exp Date: 11-2-14

Known Value: 0.020 AC

Results: 0.019 AC 0.020 AC 0.020 AC

3) Linearity Test (Level 1,C):

Sim Ser #: DR3847 Lot #: 13020 Exp Date: 1-15-15

Known Value: 0.300 AC

Results: ~~0.299~~ AC 0.299 AC 0.299 AC
0.295

b. Configure Dry Gas Standard (Level 1,S):

1) Interferent Detected Tests (Level 1,B):

1. 0.10 AC + 0.05% Acetone

Sim Ser #: DR3377 Lot #: ICS#1 Exp Date: 4-9-14

Known Value: 0.10 AC + 0.05% ACETONE

Display reads "Interferent Detect"

2. Acetone

Sim Ser #: G8919 Lot #: ACE1 Exp Date: 8-27-15

Known Value: 0.05% ACETONE

Display reads "Interferent Detect"

3. Methanol

Sim Ser #: G10573 Lot #: MET1 Exp Date: 8-27-15

Known Value: 0.10% METHANOL

Display reads "Interferent Detect"

4. Isopropanol
Sim Ser #: G9998 Lot #: IPA1 Exp Date: 8-27-15
Known Value: 0.10% ISOPROPANOL
 Display reads "Interferent Detect"
- 2) Ambient Fail (Level 1,B): Display and test record are correct
3) Invalid Sample (Level 1,B): Display and test record are correct
4) Subject Test Timeout (Level 1,B): Display and test record are correct.
5) Sequence Aborted (Level 1,B): Display and test record are correct
6) Control Outside Tolerance (CMS mode): Display and test record are correct
7) RFI Detected Test (CMS mode): Display and test record are correct
8) Tank Pressure Below Min (CMS mode): Display reads correct
9) Subject Refused (CMS mode): Display and test record are correct
10) Difference Too Great (CMS mode): Display and test record are correct
11) Stability Test (Level 3,T,S):
Average 0.0812 Std Dev 0.0006 Rel Std Dev 0.73%

Maintenance/Repairs/Notes: _____

THIS UNIT IS ACCEPTABLE TO BE USED IN THE FIELD.

Charles E. Eder

Field Inspector's Signature

20 SEP 13

Date

Deb Keshu

Reviewed by

20 Sept 2013

Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CM1, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005937
Location = TOXL 8164.13.00 06/09
09/20/2013 09:39

***** Printer Test *****

12 char/line
abcdefghijkl
ABCDEFGHIJKL

16 char per line
abcdefghijklmnop
ABCDEFGHIJKLMN

21 chars per line
abcdefghijklmnopqrstu
ABCDEFGHIJKLMN

24 characters per line
abcdefghijklmnopqrstuwx
ABCDEFGHIJKLMN

32 characters per line
abcdefghijklmnopqrstuwx123456
ABCDEFGHIJKLMN

42 characters per line
abcdefghijklmnopqrstuwx1234567890abcdef
ABCDEFGHIJKLMN

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On MemFull? No
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 23411080A1
Standard Cyl #? 47
Standard Expiration? 10/01/2013
Oper No? 666666

Flow Cal. Date: 07/19/2013
Slope 683
Intercept -742419

IR Calibration Date: 07/19/2013
3um 9um

0th Coef(*100): -21480 -18589
1st Coef(*100): 274353 136859
2nd Coef(*100): 2162 1226
H2O adj(mg/l*10K): 552 490

***** Test End *****



Operator Signature
CHARLES EDER

Remarks: **INTERNAL
PRINT
TEST**
Form 106-18000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005937
Location = TOXL 8164.13.00 06/09
09/20/2013 09:43

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? No
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 23411080A1
Standard Cyl #? 47
Standard Expiration? 10/01/2013
Oper No? 666666

Flow Cal. Date: 07/19/2013
Slope 683
Intercept -742419

IR Calibration Date: 07/19/2013
 3um 9um

0th Coef(*100): -21480 -18589
1st Coef(*100): 274353 136859
2nd Coef(*100): 2162 1226
H2O adj(mg/l*10k): 552 490

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks: **EXTERNAL
PRINT TEST**

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005937
Location = TOXL 8164.13.00 06/09
09/20/2013 09:43

DIAGNOSTIC

Voltage/Current Test	Pass
RAM Test	Pass
EEPROM Checksum Test	Pass
Real Time Clock Test	Pass
DSP Test	Pass
Analytical Stability Test	Pass
Modem Test	Pass
Temp Regulation Test	Pass

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *DIAGNOSTIC*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005937
Location = TOXL 8164.13.00 06/09
09/20/2013 09:52

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:53
02 Std. Sol.	0.000	09:54
03 Room Air	0.000	09:54
04 Std. Sol.	0.000	09:55
05 Room Air	0.000	09:55
06 Std. Sol.	0.000	09:56
07 Room Air	0.000	09:57

08 Sim Temp = 34.0°C

Simul Ser No = DR5188
Std Sol No = MILLIQ H2O
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

ZERO ALCOHOL CHECK

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005937
Location = TOXL 8164.13.00 06/09
09/20/2013 09:57

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:58
02 Std. Sol.	0.019	09:59
03 Room Air	0.000	09:59
04 Std. Sol.	0.020	10:00
05 Room Air	0.000	10:00
06 Std. Sol.	0.020	10:01
07 Room Air	0.000	10:02

08 Sim Temp = 34.0°C

Simul Ser No = DR3455

Std Sol No = 12210

County = 08

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

LOW AC CHECK

0.020AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005937
Location = TOXL 8164.13.00 06/09
09/20/2013 10:03

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:03
02 Std. Sol.	0.295	10:04
03 Room Air	0.000	10:05
04 Std. Sol.	0.299	10:05
05 Room Air	0.000	10:06
06 Std. Sol.	0.299	10:07
07 Room Air	0.000	10:07

08 Sim Temp = 34.0°C

Simul Ser No = DR3847
Std Sol No = 13020
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *LINEARITY CHECK*
0.300 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005937
Location = TOXL 8164.13.00 06/09
09/20/2013 10:09

Test	AC	Time
01 Room Air	0.000	10:10
02 *Subject Test	INT*	10:11
03 Room Air	0.000	10:12

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = 0.10AC+0.05ACE
Dr. Lic. = ND/DIS821456
Lot No = 23411080A1
Cyl No = 47
Expiration Date = 10/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *INTERFERENT DETECTED CHECK*
0.10AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005937
Location = TOXL 8164.13.00 06/09
09/20/2013 10:13

Test	AC	Time
01 Room Air	0.000	10:14
02 *Subject Test	INT*	10:15
03 Room Air	0.000	10:16

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = 0.05 ACETONE
Dr. Lic. = ND/DIS821456
Lot No = 23411080A1
Cyl No = 47
Expiration Date = 10/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *INTERFERENT DETECTED CHECK
0.05% ACETONE*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005937
Location = TOXL 8164.13.00 06/09
09/20/2013 10:17

Test	AC	Time
01 Room Air	0.000	10:18
02 *Subject Test	INT*	10:18
03 Room Air	0.000	10:19

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = 0.10 METHANOL
Dr. Lic. = ND/DIS821456
Lot No = 23411080A1
Cyl No = 47
Expiration Date = 10/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *INTERFERENT DETECTED CHECK*
0.10% METHANOL

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005937
Location = TOXL 8164.13.00 06/09
09/20/2013 10:20

Test	AC	Time
01 Room Air	0.000	10:21
02 *Subject Test	INT*	10:22
03 Room Air	0.000	10:22

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = 0.10 IPA
Dr. Lic. = ND/DIS821456
Lot No = 23411080A1
Cyl No = 47
Expiration Date = 10/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

INTERFERENT DETECTED CHECK
0.10% ISOPROPANOL

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005937
Location = TOXL 8164.13.00 06/09
09/20/2013 10:25

Test	AC	Time
01 Room Air	AMB*	10:26
02 Room Air	0.000	10:26

*Invalid Test
Check Ambient Conditions

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = AMBIENT FAIL
Dr. Lic. = ND/DIS821456
Lot No = 23411080A1
Cyl No = 47
Expiration Date = 10/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *AMBIENT FAIL CHECK*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005937
Location = TOXL 8164.13.00 06/09
09/20/2013 10:28

Test	AC	Time
01 Room Air	0.000	10:29
02 Invalid Sample	X.XXX	10:29
03 Room Air	0.000	10:30

*Invalid Test - Mouth Alcohol

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INVALID SAMPLE
Dr. Lic. = ND/DIS821456
Lot No = 23411080A1
Cyl No = 47
Expiration Date = 10/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *INVALID SAMPLE CHECK*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005937
Location = TOXL 8164.13.00 06/09
09/20/2013 10:31

Test	AC	Time
01 Room Air	0.000	10:31
02 *Subject Test	0.000*	10:34
03 Room Air	0.000	10:35

No RFI Detected

*Deficient Sample - Value Printed was
Highest Obtained

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = TEST TIMEOUT
Dr. Lic. = ND/DIS821456
Lot No = 23411080A1
Cyl No = 47
Expiration Date = 10/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *SUBJECT TEST TIMEOUT CHECK*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005937
Location = TOXL 8164.13.00 06/09
09/20/2013 10:39

Test	AC	Time
01 Room Air	0.000	10:40
02 *Subject Test	ABT*	10:40
03 Room Air	0.000	10:40

*Sequence Aborted

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = SEQUENCE ABORT
Dr. Lic. = ND/DIS821456
Lot No = 23411080A1
Cyl No = 47
Expiration Date = 10/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *SEQUENCE ABORTED CHECK*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005937
Location = TOXL 8164.13.00 06/09
09/20/2013 10:41

Test	AC	Time
01 Diagnostic	OK	10:42
02 Room Air	0.000	10:43
03 Subject Test 1	0.022	10:43
04 Room Air	0.000	10:46
05 Std. Gas	0.089*	10:47
06 Room Air	0.000	10:47

*Cal Check Out of Tolerance

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = CON OUT TOL
Dr. Lic. = ND/DIS821456
Lot No = 23411080A1
Cyl No = 47
Expiration Date = 10/01/2013
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: *CONTROL OUTSIDE TOLERANCE*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005937
Location = TOXL 8164.13.00 06/09
09/20/2013 10:49

Test	AC	Time
01 Diagnostic	OK	10:50
02 Room Air	0.000	10:51
03 *Subject Test	RFI*	10:51
04 Room Air	0.000	10:52

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 23411080A1
Cyl No = 47
Expiration Date = 10/01/2013
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005937
Location = TOXL 8164.13.00 06/09
09/20/2013 10:54

Test	AC	Time
01 Diagnostic	OK	10:55
02 Room Air	0.000	10:56
03 Subject Test 1	0.097	10:56
04 Room Air	0.000	10:58
05 Std. Gas	0.081	10:59
06 Room Air	0.000	11:01
07 *Subject Test	REF*	11:01
08 Room Air	0.000	11:01

*Subject Test Refused

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = SUBJECT REFUSED
Dr. Lic. = ND/DIS821456
Lot No = 23411080A1
Cyl No = 47
Expiration Date = 10/01/2013
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: **SUBJECT REFUSED CHECK**

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005937
Location = TOXL 8164.13.00 06/09
09/20/2013 11:01

Test	AC	Time
01 Diagnostic	OK	11:02
02 Room Air	0.000	11:03
03 Subject Test 1	0.043*	11:04
04 Room Air	0.000	11:07
05 Std. Gas	0.081	11:08
06 Room Air	0.000	11:09
07 Subject Test 2	0.097*	11:10
08 Room Air	0.000	11:12

*Difference Too Great

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = DIFF TOO GREAT
Dr. Lic. = ND/DIS821456
Lot No = 23411080A1
Cyl No = 47
Expiration Date = 10/01/2013
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

*DIFFERENCE TOO GREAT
CHECK*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005937
Location = TOXL 8164.13.00 06/09
09/20/2013 11:14

Test	AC	Time
01 Room Air	0.000	11:14
02 Std. Gas	0.080	11:14
03 Room Air	0.000	11:15
04 Std. Gas	0.081	11:15
05 Room Air	0.000	11:16
06 Std. Gas	0.080	11:16
07 Room Air	0.000	11:17
08 Std. Gas	0.081	11:17
09 Room Air	0.000	11:18
10 Std. Gas	0.081	11:18
11 Room Air	0.000	11:18
12 Std. Gas	0.082	11:19
13 Room Air	0.000	11:19
14 Std. Gas	0.081	11:20
15 Room Air	0.000	11:20
16 Std. Gas	0.081	11:21
17 Room Air	0.000	11:21
18 Std. Gas	0.081	11:21
19 Room Air	0.000	11:22
20 Std. Gas	0.081	11:22
21 Room Air	0.000	11:23
22 Std. Gas	0.081	11:23
23 Room Air	0.000	11:24
24 Std. Gas	0.081	11:24
25 Room Air	0.000	11:25
26 Std. Gas	0.081	11:25
27 Room Air	0.000	11:25
28 Std. Gas	0.082	11:26
29 Room Air	0.000	11:26
30 Std. Gas	0.081	11:27
31 Room Air	0.000	11:27
32 Std. Gas	0.081	11:28
33 Room Air	0.000	11:28
34 Std. Gas	0.081	11:28
35 Room Air	0.000	11:29
36 Std. Gas	0.082	11:29
37 Room Air	0.000	11:30
38 Std. Gas	0.082	11:30
39 Room Air	0.000	11:31
40 Std. Gas	0.082	11:31
41 Room Air	0.000	11:32
42 Cal Check Stats:		
43 Average	0.0812	
44 Std Dev	0.0006	
45 Rel Std Dev	0.73%	

STABILITY TEST
Charles E. Ede
pg 1 of 2

Operator Signature
CHARLES EDER

Remarks:

STABILITY TEST

Form 106-I8000

LOT#: 23411080 A1

Cyl #: 047

Exp. DATE: 10/1/2013

0.080 AC

STABILITY TEST
Charles Eder
pg 2 of 2