

**INTOXILYZER® 8000 INITIAL INSPECTION**

Serial Number: 80-005953

1. Power On and Warm-up:
  - a.  Power On
  - b.  5 min. Countdown from Initial Warm-up
  - c.  1 min. Countdown to Begin a Test
  - d.  Forms Load Complete
  
2. Setup and Checks:
  - a. Level 1:
    - 1) Passwords:  Field Inspector (Level 2)  Lab Personnel (Level 3)
    - 2) Gas Cylinder:  Gas Cylinder Setup (Level 1,S)
    - 3) Printer:  Printer Setup (Level 3,S,P)
      - Internal Printer Paper Feed
      - Internal Printer Test (Level 1,P)
      - External Printer Test (Level 1,P)
    - 4) Barcode Reader (Level 1,B):  Operator ID  
 ND License
  
  - b. Level 2:
    - 1)  Continuous Purge (Level 2,A)
    - 2)  Diagnostic (Level 2,D)
    - 3)  Prelim Data Entry (Level 2,E)
    - 4)  Ser Num/Version Info (Level 2,V): Ser Num 80-005953  
Version 8164.13
  
  - c. Level 3:
    - 1) DVM Monitor (Level 3,D,D): 3: 12698 s: 1 C: 47.00  
9: 13182 s: 2 B: 45 0  
(RFI indicator with keyed radio) 1
  
    - 2) Function (F) Keys:  F5 – Tone  
 F6 – Green LED  
 F7 – Red LED  
 F9 – Fan  
 F10 – Pump  
 F11 – Solenoid
  
    - 3) Atmospheric Monitor (Level 3,D,A): Atmo Sensor, P: 958 CF: 1.05
    - 4) Flow Monitor (Level 3,D,F): Flow Sensor, R: 213 F: 0.000 L/s

5) Temperature Monitor (Level 3,D,T): C: 47.0 A: 48.5 I: 0.0  
B: 45.0 V: 0.0

6) Tank Monitor (Level 3,D,G): P (gauge): 2485 P (display): 476  
7)  General Setup (Level 3,S,G)  
8)  Configure Start Test (Level 3,S,T)  
9)  Network (Level 3,S,N)

3. Tests:

a. Configure Standard for Wet Bath Simulator (Level 1,S):

1) Zero Alcohol Test (Level 1,C):

Sim Ser #: 610329 Lot #: N/A Exp Date: N/A  
Known Value: 0.000 AC  
Results: 0.000 AC 0.000 AC 0.000 AC

2) Low AC Test (Level 1,C):

Sim Ser #: DR3803 Lot #: 13070 Exp Date: 03/06/2015  
Known Value: 0.020 AC  
Results: 0.020 AC 0.020 AC 0.021 AC

3) Linearity Test (Level 1,C):

Sim Ser #: DR5143 Lot #: 13080 Exp Date: 03/19/2015  
Known Value: 0.300 AC  
Results: 0.297 AC 0.298 AC 0.299 AC

b. Configure Dry Gas Standard (Level 1,S):

1) Interferent Detected Tests (Level 1,B):

1. 0.10 AC + 0.05% Acetone

Sim Ser #: DR3846 Lot #: ICS2 Exp Date: 08/27/2015  
Known Value: 0.10% AC + 0.05% acetone  
 Display reads "Interferent Detect"

2. Acetone

Sim Ser #: DR3375 Lot #: ACE1 Exp Date: 08/27/2015  
Known Value: 0.05% acetone  
 Display reads "Interferent Detect"

3. Methanol

Sim Ser #: DR3454 Lot #: MET1 Exp Date: 08/27/2015  
Known Value: 0.10% Methanol  
 Display reads "Interferent Detect"

4. Isopropanol

Sim Ser #: DR5131 Lot #: IPA1 Exp Date: 08/27/2015

Known Value: 0.10% Methanol OK Isopropanol

Display reads "Interferent Detect"

- 2)  Ambient Fail (Level 1,B): Display and test record are correct
- 3)  Invalid Sample (Level 1,B): Display and test record are correct
- 4)  Subject Test Timeout (Level 1,B): Display and test record are correct.
- 5)  Sequence Aborted (Level 1,B): Display and test record are correct
- 6)  Control Outside Tolerance (CMS mode): Display and test record are correct
- 7)  RFI Detected Test (CMS mode): Display and test record are correct
- 8)  Tank Pressure Below Min (CMS mode): Display reads correct
- 9)  Subject Refused (CMS mode): Display and test record are correct
- 10)  Difference Too Great (CMS mode): Display and test record are correct
- 11)  Stability Test (Level 3,T,S):

Average 0.0809 Std Dev 0.0004 Rel Std Dev 0.5690

Maintenance/Repairs/Notes: Initial inspection. Recalibrated gas regulator.

Deb Kashner  
Field Inspector's Signature

20 Sept 2013  
Date

Charles E. Edler  
Reviewed by

20 SEP 13  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer  
North Dakota Model 8000 SN 80-005953  
Location = TOXL 8164.13.00 06/09  
09/20/2013 09:17

\*\*\*\*\* Printer Test \*\*\*\*\*

12 char/line  
abcdefghijkl  
ABCDEFGHIJKL

16 char per line  
abcdefghijklmnop  
ABCDEFGHIJKLMN

21 chars per line  
abcdefghijklmnopqrstu  
ABCDEFGHIJKLMN

24 characters per line  
abcdefghijklmnopqrstuwx  
ABCDEFGHIJKLMN

32 characters per line  
abcdefghijklmnopqrstuwx123456  
ABCDEFGHIJKLMN

42 characters per line  
abcdefghijklmnopqrstuwx1234567890abcdef  
ABCDEFGHIJKLMN

Current Instrument Setup

Data Entry Mode: Enabled  
Start Test Sequence: DABACABA  
Display Prelim Rslt? Yes  
Display Third Digit? Yes  
Inhib Printer(Y/N)? No  
Display Volume? No  
Disable On MenFull? No  
# of Print Copies? 1  
Select Std (D/W/I)? Dry  
Standard Value? 0.080  
Standard Lot #? 123456789  
Standard Cyl #? 120W3ER  
Standard Expiration? 11/11/2014  
Oper No? 777777

Flow Cal. Date: 06/21/2013  
Slope 674  
Intercept -679248

IR Calibration Date: 06/21/2013  
3um 9um

0th Coef(\*100): -14815 -18665  
1st Coef(\*100): 263530 132860  
2nd Coef(\*100): 3000 1324  
H2O adj(mg/l\*10K): 578 483

\*\*\*\*\* Test End \*\*\*\*\*

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Print test*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer           Alcohol Analyzer  
North Dakota Model 8000           SN 80-005953  
Location = TOXL                   8164.13.00 06/09  
09/20/2013                               09:21

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&\*()\_+?

Current Instrument Setup

Data Entry Mode:           Enabled  
Start Test Sequence:       DABACABA  
Display Prelim Rslt?       Yes  
Display Third Digit?       Yes  
Inhib Printer(Y/N)?       No  
Display Volume?           No  
Disable On Memfull?       No  
# of Print Copies?         1  
Select Std (D/W/I)?       Dry  
Standard Value?           0.080  
Standard Lot #?            123456789  
Standard Cyl #?           12QW3ER  
Standard Expiration?       11/11/2014  
Oper No?                   777777

Flow Cal. Date:            06/21/2013  
Slope                      674  
Intercept                  -679248

IR Calibration Date:       06/21/2013  
                              3um                    9um

	3um	9um
0th Coef(*100):	-14815	-18665
1st Coef(*100):	263530	132860
2nd Coef(*100):	3000	1324
H2O adj(mg/l*10k):	578	483

\*\*\*\*\* Printer Test End \*\*\*\*\*

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Print test*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005953  
Location = TOXL                    8164.13.00 06/09  
09/20/2013                            09:25

DIAGNOSTIC

Voltage/Current Test	Pass
RAM Test	Pass
EEPROM Checksum Test	Pass
Real Time Clock Test	Pass
DSP Test	Pass
Analytical Stability Test	Pass
Modem Test	Pass
Temp Regulation Test	Pass

Oper No. = 777777

*Deb Keshur*

Operator Signature  
DEB KASHUR

Remarks:

*Diagnostic test*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005953  
Location = TOXL      8164.13.00 06/09  
09/20/2013      09:33

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:33
02 Std. Sol.	0.000	09:34
03 Room Air	0.000	09:35
04 Std. Sol.	0.000	09:36
05 Room Air	0.000	09:36
06 Std. Sol.	0.000	09:37
07 Room Air	0.000	09:37

08 Sim Temp = 34.0°C

Simul Ser No = G10329  
Std Sol No = MILLI Q  
County = 08      Oper No. = 777777

Deb Kasher  
Operator Signature  
DEB KASHUR

Remarks:

Zero alcohol 0.000AE

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005953  
Location = TOXL      8164.13.00 06/09  
09/20/2013      09:38

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:39
02 Std. Sol.	0.020	09:39
03 Room Air	0.000	09:40
04 Std. Sol.	0.020	09:41
05 Room Air	0.000	09:41
06 Std. Sol.	0.021	09:42
07 Room Air	0.000	09:42

08 Sim Temp = 34.0°C

Simul Ser No = DR3803

Std Sol No = 13070

County = 08

Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Low AC      0.020 AC*

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005953  
Location = TOXL      8164.13.00 06/09  
09/20/2013      09:45

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:46
02 Std. Sol.	0.297	09:47
03 Room Air	0.000	09:47
04 Std. Sol.	0.298	09:48
05 Room Air	0.000	09:49
06 Std. Sol.	0.299	09:49
07 Room Air	0.000	09:50

08 Sim Temp = 34.0°C

Simul Ser No = DR5143

Std Sol No = 13080

County = 08

Oper No. = 777777

Deb Keshur

Operator Signature  
DEB KASHUR

Remarks:

Linearity test 0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005953  
Location = TOXL                8164.13.00 06/09  
09/20/2013                        10:46

Test	AC	Time
01 Room Air	0.000	10:47
02 *Subject Test	INT*	10:47
03 Room Air	0.000	10:48

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                Weight = 150  
Test = DUI                        Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 23411080A1  
Cyl No = 38  
Expiration Date = 10/01/2013  
County = 08                        Oper No. = 777777

Deb Kashur  
Operator Signature  
DEB KASHUR

Remarks:

*Interferent check*            *0.10AC + 0.05% acetone*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005953  
Location = TOXL                    8164.13.00 06/09  
09/20/2013                            10:48

Test	AC	Time
01 Room Air	0.000	10:49
02 *Subject Test	INT*	10:50
03 Room Air	0.000	10:50

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 23411080A1  
Cyl No = 38  
Expiration Date = 10/01/2013  
County = 08                            Oper No. = 777777

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Interferent check            0.05% acetone*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005953  
Location = TOXL                    8164.13.00 06/09  
09/20/2013                            10:50

Test	AC	Time
01 Room Air	0.000	10:51
02 *Subject Test	INT*	10:51
03 Room Air	0.000	10:52

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 23411080A1  
Cyl No = 38  
Expiration Date = 10/01/2013  
County = 08                            Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Interferent check 0.10% Methanol*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005953  
Location = TOXL                    8164.13.00 06/09  
09/20/2013                            10:53

Test	AC	Time
01 Room Air	0.000	10:54
02 *Subject Test	INT*	10:54
03 Room Air	0.000	10:55

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 23411080A1  
Cyl No = 38  
Expiration Date = 10/01/2013  
County = 08                            Oper No. = 777777

Deb Kashur  
Operator Signature  
DEB KASHUR

Remarks:  
*Interferent check 0.10% Isopropanol*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005953  
Location = TOXL      8164.13.00 06/09  
09/20/2013      10:56

Test	AC	Time
01 Room Air	AMB*	10:57
02 Room Air	0.000	10:57

\*Invalid Test  
Check Ambient Conditions

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 23411080A1  
Cyl No = 38  
Expiration Date = 10/01/2013  
County = 08      Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Ambient fail test*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005953  
Location = TOXL                    8164.13.00 06/09  
09/20/2013                            10:57

Test	AC	Time
01 Room Air	0.000	10:58
02 Invalid Sample	X.XXX	10:58
03 Room Air	0.000	10:59

\*Invalid Test - Mouth Alcohol

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 23411080A1  
Cyl No = 38  
Expiration Date = 10/01/2013  
County = 08                            Oper No. = 777777

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Invalid sample test*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005953  
Location = TOXL                8164.13.00 06/09  
09/20/2013                        11:00

Test	AC	Time
01 Room Air	0.000	11:01
02 *Subject Test	0.000*	11:04
03 Room Air	0.000	11:04

No RFI Detected

\*Deficient Sample - Value Printed was  
Highest Obtained

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 23411080A1  
Cyl No = 38  
Expiration Date = 10/01/2013  
County = 08                          Oper No. = 777777

Deb Kasher

Operator Signature  
DEB KASHUR

Remarks:

*Subject test timeout*

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005953  
Location = TOXL      8164.13.00 06/09  
09/20/2013      11:04

Test	AC	Time
01 Room Air	0.000	11:05
02 *Subject Test	ABT*	11:05
03 Room Air	0.000	11:06

\*Sequence Aborted

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 23411080A1  
Cyl No = 38  
Expiration Date = 10/01/2013  
County = 08      Oper No. = 777777



Operator Signature  
DEB KASHUR

Remarks:

*Sequence aborted test*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005953  
Location = TOXL                    8164.13.00 06/09  
09/20/2013                            11:07

Test	AC	Time
01 Diagnostic	OK	11:07
02 Room Air	0.000	11:08
03 Subject Test 1	0.000	11:09
04 Room Air	0.000	11:12
05 Std. Gas	0.000*	11:13
06 Room Air	0.000	11:13

\*Cal Check Out of Tolerance

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 23411080A1  
Cyl No = 38  
Expiration Date = 10/01/2013  
County = 08                            Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kasher

Operator Signature  
DEB KASHUR

Remarks:

*Control ~~outs~~ outside of tolerance*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005953  
Location = TOXL      8164.13.00 06/09  
09/20/2013      11:18

Test	AC	Time
01 Diagnostic	OK	11:19
02 Room Air	RFI*	11:19
03 Room Air	0.000	11:19

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 23411080A1  
Cyl No = 38  
Expiration Date = 10/01/2013  
County = 08      Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kashur

Operator Signature  
DEB KASHUR

Remarks:

RFI check

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005953  
Location = TOXL                    8164.13.00 06/09  
09/20/2013                            11:19

Test	AC	Time
01 Diagnostic	OK	11:20
02 Room Air	0.000	11:20
03 *Subject Test	REF*	11:21
04 Room Air	0.000	11:21

\*Subject Test Refused

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 23411080A1  
Cyl No = 38  
Expiration Date = 10/01/2013  
County = 08                            Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Subject refused test*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005953  
Location = TOXL                    8164.13.00 06/09  
09/20/2013                            11:30

Test	AC	Time
01 Diagnostic	OK	11:30
02 Room Air	0.000	11:31
03 Subject Test 1	0.051*	11:31
04 Room Air	0.000	11:34
05 Std. Gas	0.080	11:35
06 Room Air	0.000	11:36
07 Subject Test 2	0.000*	11:36
08 Room Air	0.000	11:38

\*Difference Too Great

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 23411080A1  
Cyl No = 38  
Expiration Date = 10/01/2013  
County = 08                            Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
DEB KASHUR

Remarks:

*Difference too great*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005953  
Location = TOXL                8164.13.00 06/09  
09/20/2013                      11:40

Test	AC	Time
01 Room Air	0.000	11:40
02 Std. Gas	0.081	11:41
03 Room Air	0.000	11:41
04 Std. Gas	0.081	11:41
05 Room Air	0.000	11:42
06 Std. Gas	0.080	11:42
07 Room Air	0.000	11:43
08 Std. Gas	0.081	11:43
09 Room Air	0.000	11:44
10 Std. Gas	0.081	11:44
11 Room Air	0.000	11:45
12 Std. Gas	0.081	11:45
13 Room Air	0.000	11:45
14 Std. Gas	0.081	11:46
15 Room Air	0.000	11:46
16 Std. Gas	0.080	11:47
17 Room Air	0.000	11:47
18 Std. Gas	0.081	11:48
19 Room Air	0.000	11:48
20 Std. Gas	0.081	11:48
21 Room Air	0.000	11:49
22 Std. Gas	0.080	11:49
23 Room Air	0.000	11:50
24 Std. Gas	0.081	11:50
25 Room Air	0.000	11:51
26 Std. Gas	0.081	11:51
27 Room Air	0.000	11:52
28 Std. Gas	0.081	11:52
29 Room Air	0.000	11:52
30 Std. Gas	0.081	11:53
31 Room Air	0.000	11:53
32 Std. Gas	0.081	11:54
33 Room Air	0.000	11:54
34 Std. Gas	0.081	11:54
35 Room Air	0.000	11:55
36 Std. Gas	0.081	11:55
37 Room Air	0.000	11:56
38 Std. Gas	0.081	11:56
39 Room Air	0.000	11:57
40 Std. Gas	0.082	11:57
41 Room Air	0.000	11:58
42 Cal Check Stats:		
43 Average	0.0809	
44 Std Dev	0.0004	
45 Rel Std Dev	0.56%	

Stability test  
Deb Keshner  
p. 1/2

Operator Signature  
DEB KASHUR

Remarks:

Lot #: ~~+234~~<sup>PK</sup> 23411080A1

Cyl #: 38

Form 106-I8000

Exp. date: 10/01/2013

Stability test  
Deb Kashur  
p. 2/2