



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name:	William Petty				
Location:	GFAE				
Intoxilyzer® Serial Number:	80-002667				
Gas Lot Number:	659358	Gas Cylinder Number:	39	Expiration Date:	09/03/11
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed				
Instrument Test Results (Report to 3 Digits; Example, 0.081%)					
Test 1: 0.079%					
Test 2: 0.080%					
Test 3: 0.080%					
(ATTACH TEST RECORD)					

Will Petty
Chemical Test Operator Signature

22 Sept 10
Date

Deb Kashner
Reviewed By

18 Oct 2010
Date

80-002667
P. 72 of 73
JCM

Intoxilyzer Test Record and Checklist
NDDAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8010 SN 60-002667
Location = TOXL 8164.13.00 05/09
09/22/2010 10:29

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:31
02 Std. Gas	0.079	10:32
03 Room Air	0.000	10:32
04 Std. Gas	0.080	10:32
05 Room Air	0.000	10:33
06 Std. Gas	0.080	10:33
07 Room Air	0.000	10:34

Lot No = 15410080A
Cyl No = 28
Exp Date = 07/01/2012
County = 18 Oper No. = 120011

William R. Petty

Operator Signature
WILLIAM R. PETTY

Remarks:

Form 106-18000

80-002667
P. 73 of 73
JRM