

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8111 SN 80-002667
Location = ~~TOX-6FAF~~ 8164.13.00 06/09
09/22/2010 ~~DK~~ *DK* 29 April 10:29

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:31
02 Std. Gas	0.079	10:32
03 Room Air	0.000	10:32
04 Std. Gas	0.080	10:32
05 Room Air	0.000	10:33
06 Std. Gas	0.080	10:33
07 Room Air	0.000	10:34

DK 29 April

Lot No = ~~45410000A1~~ *659358*
Cyl No = ~~28~~ *39*
Exp Date = ~~07-01-2012~~ *09/03/11*
County = 18 Oper No. = 120011

William R. Petty

Operator Signature
WILLIAM R. PETTY

Remarks:

Form 106-18000

*80-002667
P. 96 of 96
WMP*

*80-002667
P. 73 of 73
WMP*