



# INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59281 (03-2014)

Serial Number 80-005363	Instrument Location Dickinson LEC
Reason for Install/Repair <input type="checkbox"/> Install After Receiving From Crime Laboratory <input checked="" type="checkbox"/> Install After Location Change <input checked="" type="checkbox"/> Other (Specify) <u>moved to different room in LEC</u>	

Check When Done:

- 1. Surge Protector Installed/Properly Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E). *-all checked ok*
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S). *-already done*
- 6. Run Tests:
  - A. Print Test (Level 1, Function P).
  - B. ACA Test (Level 1, Function C).
  - C. RFI Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if Any): None
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer®8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature 	Date 9-22-15 ✓
Reviewed By (Crime Laboratory Use Only) Deb Stanaver	Date 13 Oct 15

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005363  
Location = DKPD                8164.13.00 06/09  
09/22/2015                      11:34

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^&\*()\_+?

Current Instrument Setup

Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?         1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            29614080A2  
Standard Cyl #?            10  
Standard Expiration?       12/05/2016  
Oper No?                    010107

Flow Cal. Date:            08/19/2011  
Slope                       660  
Intercept                  -560671

IR Calibration Date:       04/08/2015  
                                    3um                      9um

	3um	9um
0th Coef(*100):	-12875	-23244
1st Coef(*100):	269300	140223
2nd Coef(*100):	4133	1487
H2O adj(mg/l*10k):	341	408

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
CHRISTOPHER MESSER

Remarks:


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005363  
Location = DKPD      8164.13.00 06/09  
09/22/2015      11:35

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:36
02 Std. Gas	0.080	11:36
03 Room Air	0.000	11:37
04 Std. Gas	0.079	11:37
05 Room Air	0.000	11:37
06 Std. Gas	0.079	11:38
07 Room Air	0.000	11:38

Lot No = 29614080A2  
Cyl No = 10  
Exp Date = 12/05/2016  
County = 45      Oper No. = 010107



Operator Signature  
CHRISTOPHER MESSER

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005363  
Location = DKPD                8164.13.00 06/09  
09/22/2015                        11:39

Test	AC	Time
01 Room Air	0.000	11:41
02 *Subject Test	RFI*	11:41
03 Room Air	0.000	11:41

\*Invalid Test  
  Inhibited - RFI

Sub Name = TEST, RADIO NA  
Sub DOB = 01/01/1900  
Sub Sex = Male                    Weight = 120  
Test = DUI                        Cit = 111111  
Dr. Lic. = ND/1111111  
Lot No = 29614080A2  
Cyl No = 10  
Expiration Date = 12/05/2016  
County = 45                        Oper No. = 010107



Operator Signature  
CHRISTOPHER MESSER

Remarks:

*RFI Test*

Form 106-I8000