



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59281 (03-2014)

Serial Number <u>80-005363</u>	Instrument Location <u>Dickinson LEC</u>
Reason for Install/Repair <input type="checkbox"/> Install After Receiving From Crime Laboratory <input checked="" type="checkbox"/> Install After Location Change <input checked="" type="checkbox"/> Other (Specify) <u>Moved to different room in LEC</u>	

Check When Done:

- 1. Surge Protector Installed/Properly Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E). *-all checked ok*
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S). *-already done*
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. RFI Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if Any): None
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer®8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature 	Date <u>9-22-15</u> ✓
Reviewed By (Crime Laboratory Use Only) <u>Deb Stanaver</u>	Date <u>13 Oct 15</u>

This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Certified By: Certified Date: 22 Nov. 2016

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005363
Location = DKPD 8164.13.00 06/09
09/22/2015 11:34

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 29614080A2
Standard Cyl #? 10
Standard Expiration? 12/05/2016
Oper No? 010107

Flow Cal. Date: 08/19/2011
Slope 660
Intercept -560671

IR Calibration Date: 04/08/2015

	3um	9um
0th Coef(*100):	-12875	-23244
1st Coef(*100):	269300	140223
2nd Coef(*100):	4133	1487
H2O adj(mg/l*10k):	341	408

***** Printer Test End *****



Operator Signature
CHRISTOPHER MESSER

Remarks:

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005363
Location = DKPD 8164.13.00 06/09
09/22/2015 11:35

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:36
02 Std. Gas	0.080	11:36
03 Room Air	0.000	11:37
04 Std. Gas	0.079	11:37
05 Room Air	0.000	11:37
06 Std. Gas	0.079	11:38
07 Room Air	0.000	11:38


Lot No = 29614080A2

Cyl No = 10

Exp Date = 12/05/2016

County = 45

Oper No. = 010107



Operator Signature
CHRISTOPHER MESSER

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005363
Location = DKPD 8164.13.00 06/09
09/22/2015 11:39

Test	AC	Time
01 Room Air	0.000	11:41
02 *Subject Test	RFI*	11:41
03 Room Air	0.000	11:41

*Invalid Test
Inhibited - RFI

Sub Name = TEST, RADIO NA
Sub DOB = 01/01/1900
Sub Sex = Male Weight = 120
Test = DUI Cit = 111111
Dr. Lic. = ND/1111111
Lot No = 29614080A2
Cyl No = 10
Expiration Date = 12/05/2016
County = 45 Oper No. = 010107



Operator Signature
CHRISTOPHER MESSER

Remarks:

RFI Test

Form 106-I8000