



# ETHANOL GAS STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION

Toxicology Section/Breath Alcohol Program

SFN 59282 (10-2013)

Chemical Test Operator Name (Print) <i>Thompson, Chad A</i>		
Location <i>Cass Jail</i>	Intoxilyzer® Serial Number <i>80-005939</i>	
Gas Lot Number <i>29614080A2</i>	Gas Cylinder Number <i>8</i>	Gas Expiration Date <i>12/5/16</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).  
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
  - A. *0.081* AC
  - B. *0.082* AC
  - C. *0.081* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
  - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
  - B. ACA Test Record.

Chemical Test Operator Signature <i>[Signature]</i>	Date <i>9-23-15</i> ✓
Reviewed By (Crime Laboratory Use Only) <i>Deb (Kashur) Shanaver</i>	Date <i>28 Sept 15</i>

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005939  
Location = CASS      8164.13.00 06/09  
09/23/2015      01:16

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	01:17
02 Std. Gas	0.081	01:17
03 Room Air	0.000	01:18
04 Std. Gas	0.082	01:18
05 Room Air	0.000	01:18
06 Std. Gas	0.081	01:19
07 Room Air	0.000	01:19

Lot No = 29614080A2  
Cyl No = 8  
Exp Date = 12/05/2016  
County = 09      Oper No. = 050910

 3524

Operator Signature  
CHAD THOMPSON

Remarks:

*Cyl. Exchange*