

Database ✓

**INTOXILYZER 5000 G and 8000
INSTALLATION AND REPAIR CHECKOUT**

Serial No. 80-004186 Instrument Location BSPD

Reason for install/repair: _____ install after lab repair/ inspection
x Other (specify) Installation

Check when done

- ✓ 1. Surge protector installed/properly grounded.
- ✓ 2. Telephone line connected to Intoxilyzer KB.
- _____ 3. Simulator heated / recirculation hoses installed/heating.
- ✓ 4. Breath tube heated.
- ✓ 5. Review/enter Preliminary Data Entry (Level 2 Function E).
- ✓ 6. Ethanol Breath Standard: x Connected _____ Installed.
(Include Form 105-G if new cylinder is Installed.)
- ✓ 7. Review/enter Gas Set-up (Level 1 Function S). (Intox 8000 only)
- ✓ 8. Print Test (Level 1 Function P). (Attach Test Record.)
- ✓ 9. ACA Test (Level 1 Function C). (Attach Test Record.)
- ✓ 10. RFI Test (use CMS mode- 1st Room Air). (Attach Test Record.)
- 11. Remarks: _____
- 12. Repair and /or Maintenance Performed: _____

9-24-09

Date

10-2-09

Date

Pat Renz

Field Inspector's Signature

MA Deam

Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004186
Location = BSPD 8164.12.00 06/09
09/24/2009 12:52

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:53
02 Std. Gas	0.081	12:54
03 Room Air	0.000	12:54
04 Std. Gas	0.080	12:54
05 Room Air	0.000	12:55
06 Std. Gas	0.080	12:55
07 Room Air	0.000	12:56

Lot No = 659358
Cyl No = 13
Exp Date = 09/03/2011
County = 08 Oper No. = 029710

Pat Renz

Operator Signature
PATRICK M. RENZ

Remarks:

ACA

Form 106-I8000

80-004186
P. 3 of 4
JRM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004186
Location = BSPD 8164.12.00 06/09
09/24/2009 13:00

Test	AC	Time
01 Room Air	RFI*	13:02
02 Room Air	0.000	13:02

*Invalid Test
Inhibited - RFI

Sub Name = DOE, JON M
Sub DOB = 01/20/1959
Sub Sex = Male Weight = 180
Test = DUI Cit = 009900
Dr. Lic. = ND/2220022
Lot No = 659358
Cyl No = 13
Expiration Date = 09/03/2011
County = 08 Oper No. = 029710

Pat Renz

Operator Signature
PATRICK M. RENZ

Remarks:

RFI Test

Form 106-I8000

80-004186
p. 4 of 4 xlm