



OFFICE OF ATTORNEY GENERAL
 Crime Laboratory Division
 2641 East Main Avenue
 P.O. Box 937
 Bismarck, North Dakota 58502-0937

TEL 701-328-6159

ETHANOL BREATH STANDARD CYLINDER INSTALLATION FORM

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Pat Renz

Location: Bismarck PD

Intoxilyzer Serial No.: 80-004186

Lot Number: 659358

Cylinder Number: 13

Expiration Date: 9-3-2011

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0.080 %

Test 2: 0.080 %

Test 3: 0.080 %

(ATTACH TEST RECORD)

Pat Renz
 Chemical Test Operator Signature

9-24-09
 Date Cylinder Changed

Revid MAP

10-2-09

Form 105-G
 Toxicology Laboratory
 Crime Laboratory Division

80-004186
 P. 1 of 3
 KR


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004186
Location = BSPD 8164.12.00 06/09
09/24/2009 11:20

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:21
02 Std. Gas	0.080	11:21
03 Room Air	0.000	11:22
04 Std. Gas	0.080	11:22
05 Room Air	0.000	11:23
06 Std. Gas	0.080	11:23
07 Room Air	0.000	11:24

Lot No = 659358
Cyl No = 13
Exp Date = 09/03/2011
County = 08 Oper No. = 029710



Operator Signature
PATRICK M. RENZ

Remarks:

ACA

Form 106-I8000

80-004186
p. 2 of 3 FL


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004186
Location = BSPD 8164.12.00 06/09
09/24/2009 11:20

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:21
02 Std. Gas	0.080	11:21
03 Room Air	0.000	11:22
04 Std. Gas	0.080	11:22
05 Room Air	0.000	11:23
06 Std. Gas	0.080	11:23
07 Room Air	0.000	11:24

Lot No = 659358
Cyl No = 13
Exp Date = 09/03/2011
County = 08 Oper No. = 029710



Operator Signature
PATRICK M. RENZ

Remarks:

ACA

Form 106-I8000

80-004186
P. 3 of 3 PM