



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Steven R Fischer</u>		
Location: <u>Rolette County Sheriff's Office, Rolla, ND</u>		
Intoxilyzer® Serial Number: <u>80-004199</u>		
Gas Lot Number: <u>659358</u>	Gas Cylinder Number: <u>11</u>	Expiration Date: <u>9-3-11</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>081</u> %		
Test 2: 0. <u>081</u> %		
Test 3: 0. <u>080</u> %		
(ATTACH TEST RECORD)		

[Signature]
Chemical Test Operator Signature

9-26-10
Date

Deb Kasher
Reviewed By

18 Oct 2010
Date

80-004199
P. 914/92
Jan


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004199
Location = ROLL 8164.13.00 06/09
09/26/2010 17:16

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:16
02 Std. Gas	0.081	17:17
03 Room Air	0.000	17:17
04 Std. Gas	0.081	17:17
05 Room Air	0.000	17:18
06 Std. Gas	0.080	17:18
07 Room Air	0.000	17:19

Lot No = 659358
Cyl No = 11
Exp Date = 09/03/2011
County = 40 Oper No. = 049904



Operator Signature
STEVEN R. FISCHER

Remarks:

Form 106-18000

80-004199
p. 92 of 92
rhm