



ETHANOL BREATH STANDARD CYLINDER REPORT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>David J. Zeis</u>		
Location: <u>CAVAHER County SO</u>		<u>Langdon, ND</u>
Intoxilyzer® Serial Number: <u>80-004937</u>		
Gas Lot Number: <u>659358</u>	Gas Cylinder Number: <u>22</u>	Expiration Date: <u>09-03-2011</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%) Test 1: <u>0.077</u> % Test 2: <u>0.078</u> % Test 3: <u>0.077</u> %		
(ATTACH TEST RECORD)		

David J. Zeis
 Chemical Test Operator Signature

09/27/10
 Date

Deb Keshner
 Reviewed By

18 Oct 2010
 Date

80-004937
 P. 50 of 51
 JKM


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004937
Location = LANG 8164.13.00 06/09
09/27/2010 14:27

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:28
02 Std. Gas	0.077	14:29
03 Room Air	0.000	14:29
04 Std. Gas	0.078	14:29
05 Room Air	0.000	14:30
06 Std. Gas	0.077	14:30
07 Room Air	0.000	14:31

Lot No = 659358
Cyl No = 22
Exp Date = 09/03/2011
County = 10 Oper No. = 067518


Operator Signature
DAVID J. ZELS

Remarks:

Form 106-I8000

80-004937
p. 51 of 51
YJM