

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-002667  
Location = GFAP                    8164.13.00 06/09  
09/23/2010                            16:28

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxy1234567890-=  
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&\*()\_+?

abcdefghijklmnopqrstuvwxy1234567890-=  
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&\*()\_+?

Current Instrument Setup

Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?          1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            ~~15410080A1~~ 659358  
Standard Cyl #?            ~~28~~ 39  
Standard Expiration?       ~~07/01/2012~~ 09/03/11  
Oper No?                    120011


DK  
29 Apr 11

Flow Cal. Date:            09/10/2007  
Slope                        693  
Intercept                   -570251

IR Calibration Date:       09/10/2007  
                                  3um                    9um

-----  
0th Coef(\*100):            -21400                -13991  
1st Coef(\*100):            268177                135590  
2nd Coef(\*100):            1202                    658  
H2O adj(mg/l\*10k):        369                    165

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
WILLIAM PETTY

Remarks:

80-002667  
P. 97 of 97  
FM

80-002667  
P. 75 of 77  
FM