



# ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (5-2011)

Chemical Test Operator Name (Print) <i>Ray Weber</i>		
Location <i>TRAIL SO (HILL)</i>	Intoxilyzer® Serial Number <i>80-003067</i>	
Gas Lot Number <i>16512080A2</i>	Gas Cylinder Number <i>12</i>	Gas Expiration Date <i>08-01-2014</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).  
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
  - A. *0.080* AC
  - B. *0.079* AC
  - C. *0.080* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
  - A. Ethanol Breath Standard Cylinder Report (SFN59282, Form 105-G).
  - B. ACA Test Record.

Chemical Test Operator Signature <i>Ray Weber</i>	Date <del>09/28/2014</del> 09/28/2013 <i>DK</i>
Reviewed By (Crime Laboratory Use Only) <i>Deb Keshner</i>	Date <i>11 Oct 2013</i>

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003067  
Location = HILL      8164.13.00 06/09  
09/28/2013      15:40

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:41
02 Std. Gas	0.080	15:41
03 Room Air	0.000	15:42
04 Std. Gas	0.079	15:42
05 Room Air	0.000	15:43
06 Std. Gas	0.080	15:43
07 Room Air	0.000	15:43

Lot No = 16512080A2  
Cyl No = 12  
Exp Date = 08/01/2014  
County = 49      Oper No. = 058405

  
\_\_\_\_\_  
Operator Signature  
RAY WEBER

Remarks: