



OFFICE OF ATTORNEY GENERAL  
 Crime Laboratory Division  
 2641 East Main Avenue  
 P.O. Box 937  
 Bismarck, North Dakota 58502-0937

TEL 701-328-6159

**ETHANOL BREATH STANDARD CYLINDER INSTALLATION FORM**

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Mark W. McDonald

Location: Valley City Police Dept.

Intoxilyzer Serial No.: 80-003058

Lot Number: 659358

Cylinder Number: 47

Expiration Date: 09/03/2011

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0. 080 %

Test 2: 0. 081 %

Test 3: 0. 080 %

(ATTACH TEST RECORD)

*Mark W. McDonald*  
 Chemical Test Operator Signature

09/29/2009  
 Date Cylinder Changed

Deb Kashur  
 Form 105-G  
 Toxicology Laboratory  
 Crime Laboratory Division

13 Oct 09

80-003058  
 P. 1 of 2

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003058  
Location = VALL      8164.12.00 06/09  
09/29/2009      17:00

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:01
02 Std. Gas	0.080	17:01
03 Room Air	0.000	17:02
04 Std. Gas	0.081	17:02
05 Room Air	0.000	17:02
06 Std. Gas	0.080	17:03
07 Room Air	0.000	17:03


Lot No = 659358

Cyl No = 47

Exp Date = 09/03/2011

County = 02

Oper No. = 119502



Operator Signature  
MARK W. MCDONALD

Remarks:

Form 106-I8000

80-003058  
P. 2 of 2 FLW