



OFFICE OF ATTORNEY GENERAL  
Crime Laboratory Division  
2641 East Main Avenue  
P.O. Box 937  
Bismarck, North Dakota 58502-0937

TEL 701-328-6159

ETHANOL BREATH STANDARD CYLINDER FORM

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Jacquie DeMarce

Location: Fort Totten PD

Intoxilyzer Serial No.: 86-004201

Lot Number: 659358

Cylinder Number: 56

Expiration Date: 09/03/2011

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0. 083 %

Test 2: 0. 083 %

Test 3: 0. 084 %

(ATTACH TEST RECORD)

Jacquie DeMarce  
Chemical Test Operator Signature

9-29-09  
Date Cylinder Changed

Deb Kashner  
Reviewed By

30 Dec 2009  
Date

Form 105-G  
Toxicology Laboratory  
Crime Laboratory Division

[Type text]

80-004201  
P. 1 of 2 rx

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004201  
Location = FTOT      8164.12.00 06/09  
09/29/2009      16:31

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:31
02 Std. Gas	0.083	16:32
03 Room Air	0.000	16:32
04 Std. Gas	0.083	16:33
05 Room Air	0.000	16:33
06 Std. Gas	0.084	16:33
07 Room Air	0.000	16:34


Lot No = 659358

Cyl No = 56

Exp Date = 09/03/2011

County = 03

Oper No. = 100103

  
Operator Signature  
JACQUIE L. DEMARCE

Remarks:

*Cal test*

Form 106-18000

80-004201  
1-24-09 FM