



OFFICE OF ATTORNEY GENERAL  
 Crime Laboratory Division  
 2641 East Main Avenue  
 P.O. Box 937  
 Bismarck, North Dakota 58502-0937

TEL 701-328-6159

**ETHANOL BREATH STANDARD CYLINDER INSTALLATION FORM**

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Sean Murphy

Location: Bottineau County Sheriff's Dept.

Intoxilyzer Serial No.: 80-003667

Lot Number: 659358

Cylinder Number: 55

Expiration Date: 9-3-11

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0. 081 %

Test 2: 0. 081 %

Test 3: 0. 081 %

(ATTACH TEST RECORD)

Deb Keshur  
 Chemical Test Operator Signature

9-29-09  
 Date Cylinder Changed  
16 NOV 2009

Form 105-G  
 Toxicology Laboratory  
 Crime Laboratory Division

80-003667  
 P. 1 of 2 JKL


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003067  
Location = BOTT      8164.12.00 06/09  
09/29/2009      17:45

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:46
02 Std. Gas	0.081	17:47
03 Room Air	0.000	17:47
04 Std. Gas	0.081	17:47
05 Room Air	0.000	17:48
06 Std. Gas	0.081	17:48
07 Room Air	0.000	17:49

Lot No = 659358  
Cyl No = 55  
Exp Date = 09/03/2011  
County = 05      Oper No. = 039117

  
Operator Signature  
SEAN J. MURPHY

Remarks:

Form 106-I8000

80-003067  
p. 2 of 2 rhl