



OFFICE OF ATTORNEY GENERAL
 Crime Laboratory Division
 2641 East Main Avenue
 P.O. Box 937
 Bismarck, North Dakota 58502-0937

TEL 701-328-6159

ETHANOL BREATH STANDARD CYLINDER INSTALLATION FORM

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Jacque DeMarce

Location: Fort Totten PD

Intoxilyzer Serial No.: 80-004201

Lot Number: 659358

Cylinder Number: 56

Expiration Date: 9-3-2011

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0. 083 %

Test 2: 0. 084 %

Test 3: 0. 083 %

(ATTACH TEST RECORD)

Jacque DeMarce
 Chemical Test Operator Signature

9-29-09
 Date Cylinder Changed

Form 105-G
 Toxicology Laboratory
 Crime Laboratory Division

80-004201
 P.1 of 2 JKL


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004201
Location = FTOT 8164.12.00 06/09
09/29/2009 16:22

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:24
02 Std. Gas	0.083	16:24
03 Room Air	0.000	16:25
04 Std. Gas	0.084	16:25
05 Room Air	0.000	16:26
06 Std. Gas	0.083	16:26
07 Room Air	0.000	16:26

Lot No = 659358
Cyl No = 56
Exp Date = 09/03/2011
County = 03 Oper No. = 100103


Operator Signature
JACQUIE L. DEMARCE

Remarks:

Form 106-I8000

80-004201
P. 2-12-11