



OFFICE OF ATTORNEY GENERAL  
 Crime Laboratory Division  
 2641 East Main Avenue  
 P.O. Box 937  
 Bismarck, North Dakota 58502-0937

TEL 701-328-6159

**ETHANOL BREATH STANDARD CYLINDER INSTALLATION FORM**

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Shawn Getz

Location: Cass County Jail

Intoxilyzer Serial No.: 80-002669

Lot Number: 659357

Cylinder Number: 40

Expiration Date: 09/03/11

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0. 082 %

Test 2: 0. 082 %

Test 3: 0. 082 %

(ATTACH TEST RECORD)

S Getz  
 Chemical Test Operator Signature

10/01/09  
 Date Cylinder Changed

Deb Kashner

06 Oct 09

Form 105-G  
 Toxicology Laboratory  
 Crime Laboratory Division

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-002669  
Location = CASS      8164.12.00 06/09  
10/01/2009      16:37

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:38
02 Std. Gas	0.082	16:38
03 Room Air	0.000	16:39
04 Std. Gas	0.082	16:39
05 Room Air	0.000	16:40
06 Std. Gas	0.082	16:40
07 Room Air	0.000	16:41

Lot No = 659357  
Cyl No = 40  
Exp Date = 09/03/2011  
County = 09      Oper No. = 020535



Operator Signature  
SHAWN P. GETZ

Remarks:

Form 106-I8000

ACA - Initial Setup and Install


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-002669  
Location = CASS      8164.12.00 06/09  
10/01/2009      16:37

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:38
02 Std. Gas	0.082	16:38
03 Room Air	0.000	16:39
04 Std. Gas	0.082	16:39
05 Room Air	0.000	16:40
06 Std. Gas	0.082	16:40
07 Room Air	0.000	16:41

Lot No = 659357  
Cyl No = 40  
Exp Date = 09/03/2011  
County = 09      Oper No. = 020535

  
\_\_\_\_\_  
Operator Signature  
SHAWN P. GETZ

Remarks: