

✓ Database  
cm

INTOXILYZER® 5000 G AND 8000  
INSTALLATION AND REPAIR CHECKOUT

CM

Kidder Co So

Serial No.: 80-003059 Instrument Location: STEE

Reason for Install/Repair: \_\_\_\_\_ Install After Lab Repair/Inspection  
 Other (specify) New install

Check When Done:

- 7 A. Surge Protector Installed/Properly Grounded.
- \_\_\_\_\_ B. Telephone Line Connected to Intoxilyzer® KB.
- 7 C. Breath Tube Heated.
- 7 D. Review/Enter Preliminary Data Entry (Level 2, Function E).
- 7 E. Ethanol Breath Standard: \_\_\_\_\_ Connected \_\_\_\_\_ Installed.  
(Include Form 105-G if New Cylinder Installed.)
- 7 F. Review/Enter Gas Setup (Level 1, Function S). (Intoxilyzer® 8000 Only.)
- 7 G. Print Test (Level 1, Function P). (Attach Test Record.)
- 7 H. ACA Test (Level 1, Function C). (Attach Test Record.)
- 7 I. RFI Test (Use CMS Mode-First Room Air). (Attach Test Record.)
- J. Repair and/or Maintenance Performed: \_\_\_\_\_

01 oct 09

Date

02 oct 09 14 oct 09

Date

Debra Felber

Field Inspector's Signature

Deb Kashur

Reviewed By

80-003059  
P.1 of 10  
FM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-003059  
Location = STEE                    8164.12.00 06/09  
10/01/2009                            13:53

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

Current Instrument Setup


Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?         1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            659357  
Standard Cyl #?            37  
Standard Expiration?       09/03/2011  
Oper No?                    119404

Flow Cal. Date:            04/07/2008  
Slope                       713  
Intercept                  -692349

IR Calibration Date:       06/04/2008  
                                  3um                    9um

-----  
0th Coef(\*100):        -15808                -12583  
1st Coef(\*100):        279401                139580  
2nd Coef(\*100):        1046                    760  
H2O adj(mg/l\*10k):    478                    208

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
DAVID R. FELAND

Remarks:

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003059  
Location = STEE      8164.12.00 06/09  
10/02/2009      11:28

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:29
02 Std. Gas	0.079	11:29
03 Room Air	0.000	11:30
04 Std. Gas	0.079	11:30
05 Room Air	0.000	11:31
06 Std. Gas	0.080	11:31
07 Room Air	0.000	11:32


Lot No = 659357

Cyl No = 37

Exp Date = 09/03/2011

County = 22

Oper No. = 119404



Operator Signature  
DAVID FELAND

Remarks:

Form 106-18000

80-003059  
P. 3 of 10  
SAH


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003059  
Location = STEE      8164.12.00 06/09  
10/01/2009      14:03

Test	AC	Time
01 Room Air	RFI*	14:05
02 Room Air	0.000	14:05

\*Invalid Test  
Inhibited - RFI

Sub Name = TEST, SUBJECT NA  
Sub DOB = 06/15/1957  
Sub Sex = Male      Weight = NA  
Test = DUI      Cit = 0001  
Dr. Lic. = ND/NA  
Lot No = 659357  
Cyl No = 37  
Expiration Date = 09/03/2011  
County = 22      Oper No. = 119404



Operator Signature  
DAVID R. FELAND

Remarks:

Form 106-I8000

80-003059  
P. 4 of 10  
FM

INTOXILYZER  
INSTALLATION

Cindy -  
to Dave  
Feland  
Steel PD

OUT CML  
Kidder Co So  
STEE


Serial No.: 80-003059

Reason for Install/Repair: \_\_\_\_\_ Install After Lab Repair/Inspection  
 Other (specify) New install

Check When Done:

- 7 A. Surge Protector Installed/Properly Grounded.
- 7 B. Telephone Line Connected to Intoxilyzer® KB.
- 7 C. Breath Tube Heated.
- 7 D. Review/Enter Preliminary Data Entry (Level 2, Function E).
- 7 E. Ethanol Breath Standard: \_\_\_\_\_ Connected \_\_\_\_\_ Installed.  
(Include Form 105-G if New Cylinder Installed.)
- 7 F. Rev Need: \_\_\_\_\_ || 1, Function S). (Intoxilyzer® 8000  
Only
- 7 G. Prior ACA for \_\_\_\_\_ ). (Attach Test Record.)
- 7 H. AC/ 8000 install \_\_\_\_\_ ). (Attach Test Record.)  
+ new date
- 7 I. RFI Test (Use CMS Mode-First Room Air). (Attach Test Record.)
- J. Repair and/or Maintenance Performed: \_\_\_\_\_

01 Oct 09  
Date



**NOTE**

Dave Feland  
Field Inspector's Signature

Date

Reviewed By

INTOXILYZER® 5000 G AND 8000  
INSTALLATION AND REPAIR CHECKOUT

CML

Kidder Co So

Serial No.: 80-003059 Instrument Location: STEE

Reason for Install/Repair: \_\_\_\_\_ Install After Lab Repair/Inspection  
X Other (specify) New install

Check When Done:

- 7 A. Surge Protector Installed/Properly Grounded.
- 7 B. Telephone Line Connected to Intoxilyzer® KB.
- 7 C. Breath Tube Heated.
- 7 D. Review/Enter Preliminary Data Entry (Level 2, Function E).
- 7 E. Ethanol Breath Standard: \_\_\_\_\_ Connected \_\_\_\_\_ Installed.  
(Include Form 105-G if New Cylinder Installed.)
- 7 F. Review/Enter Gas Setup (Level 1, Function S). (Intoxilyzer® 8000 Only.)
- 7 G. Print Test (Level 1, Function P). (Attach Test Record.)
- 7 H. ACA Test (Level 1, Function C). (Attach Test Record.)
- 7 I. RFI Test (Use CMS Mode-First Room Air). (Attach Test Record.)
- J. Repair and/or Maintenance Performed: \_\_\_\_\_

01 oct 09  
Date

Dore Felber  
Field Inspector's Signature

\_\_\_\_\_  
Date

\_\_\_\_\_  
Reviewed By

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-003059  
Location = STEE                    8164.12.00 06/09  
10/01/2009                            13:53

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

Current Instrument Setup


Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?         1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            659357  
Standard Cyl #?            37  
Standard Expiration?       09/03/2011  
Oper No?                    119404

Flow Cal. Date:            04/07/2008  
Slope                       713  
Intercept                  -692349

IR Calibration Date:       06/04/2008  
                                    3um                            9um

0th Coef(*100):	-15808	-12583
1st Coef(*100):	279401	139580
2nd Coef(*100):	1046	760
H2O adj(mg/l*10k):	478	208

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
DAVID R. FELAND

Remarks:


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003059  
Location = STEE      8164.12.00 06/09  
10/01/2009      14:03

Test	AC	Time
01 Room Air	RFI*	14:05
02 Room Air	0.000	14:05

\*Invalid Test:  
Inhibited - RFI

Sub Name = TEST, SUBJECT NA  
Sub DOB = 06/15/1957  
Sub Sex = Male      Weight = NA  
Test = DUI      Cit = 0001  
Dr. Lic. = ND/NA  
Lot No = 659357  
Cyl No = 37  
Expiration Date = 09/03/2011  
County = 22      Oper No. = 119404



Operator Signature  
DAVID R. FELAND

Remarks:

Form 106-I8000

80-003059  
P. def 10  
RM





OFFICE OF ATTORNEY GENERAL  
 Crime Laboratory Division  
 2641 East Main Avenue  
 P.O. Box 937  
 Bismarck, North Dakota 58502-0937

TEL 701-328-6159

**ETHANOL BREATH STANDARD CYLINDER INSTALLATION FORM**

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: David Feland

Location: Kidder S.O.

Intoxilyzer Serial No.: 80-003059

Lot Number: 659357

Cylinder Number: 37

Expiration Date: 09/03/11

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0.080 %

Test 2: 0.080 %

Test 3: 0.~~080~~079 %

(ATTACH TEST RECORD)

David Feland  
 Chemical Test Operator Signature

01 Oct 09  
 Date Cylinder Changed

Form 105-G  
 Toxicology Laboratory  
 Crime Laboratory Division

80-003059  
 p. 9 of 10  
 yth

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003059  
Location = STEE      8164.12.00 06/09  
10/01/2009      13:53

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:54
02 Std. Gas	0.080	13:54
03 Room Air	0.000	13:55
04 Std. Gas	0.080	13:55
05 Room Air	0.000	13:56
06 Std. Gas	0.079	13:56
07 Room Air	0.000	13:57


Lot No. = 659357

Cyl No = 37

Exp Date = 09/03/2011

County = 22

Oper No. = 119404



Operator Signature  
DAVID R. FELAND

Remarks:

Form 106-I8000

80-003059  
P. 10 of 10  
FK