



OFFICE OF ATTORNEY GENERAL  
 Crime Laboratory Division  
 2641 East Main Avenue  
 P.O. Box 937  
 Bismarck, North Dakota 58502-0937

TEL 701-328-6159

**ETHANOL BREATH STANDARD CYLINDER INSTALLATION FORM**

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: David Feland

Location: Kidder S.O.

Intoxilyzer Serial No.: 80-003059

Lot Number: 659357

Cylinder Number: 37

Expiration Date: 09/03/11

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0.080 %

Test 2: 0.080 %

Test 3: 0.079 %

(ATTACH TEST RECORD)

David Feland  
 Chemical Test Operator Signature

01 oct 09  
 Date Cylinder Changed

Deb Kashner

14 oct 09

Form 105-G  
 Toxicology Laboratory  
 Crime Laboratory Division

80-003059  
 P. 1 of 2  
 SFM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003059  
Location = STEE      8164.12.00 06/09  
10/01/2009      13:53

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:54
02 Std. Gas	0.080	13:54
03 Room Air	0.000	13:55
04 Std. Gas	0.080	13:55
05 Room Air	0.000	13:56
06 Std. Gas	0.079	13:56
07 Room Air	0.000	13:57


Lot No = 659357

Cyl No = 37

Exp Date = 09/03/2011

County = 22

Oper No. = 119404



Operator Signature  
DAVID R. FELAND

Remarks:

Form 106-I8000

80-003059  
P. 2 of 2 FL