



OFFICE OF ATTORNEY GENERAL
 Crime Laboratory Division
 2641 East Main Avenue
 P.O. Box 937
 Bismarck, North Dakota 58502-0937

TEL 701-328-6159

ETHANOL BREATH STANDARD CYLINDER INSTALLATION FORM

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Officer Pete Nielsen

Location: West Fargo PD

Intoxilyzer Serial No.: 80-003065

Lot Number: 659357

Cylinder Number: 10

Expiration Date: 9-3-11

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0. 081 %

Test 2: 0. 080 %

Test 3: 0. 081 %

(ATTACH TEST RECORD)

[Signature]
 Chemical Test Operator Signature

10-1-09
 Date Cylinder Changed

Deb Kashner

19 Oct 09

Form 105-G
 Toxicology Laboratory
 Crime Laboratory Division

80-003065
 P. Nielsen

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003065
Location = WFPD 8164.12.00 06/09
10/02/2009 13:51

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:52
02 Std. Gas	0.081	13:52
03 Room Air	0.000	13:52
04 Std. Gas	0.080	13:53
05 Room Air	0.000	13:53
06 Std. Gas	0.081	13:54
07 Room Air	0.000	13:54

Lot No = 659357

Cyl No = 10

Exp Date = 09/03/2011

County = 09

Oper No. = 039617



Operator Signature
PETER D. NIELSEN

Remarks:

Form 106-I8000

80-003065
P. D. Nielsen