



OFFICE OF ATTORNEY GENERAL  
 Crime Laboratory Division  
 2641 East Main Avenue  
 P.O. Box 937  
 Bismarck, North Dakota 58502-0937

TEL 701-328-6159

**ETHANOL BREATH STANDARD CYLINDER INSTALLATION FORM**

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Darcy W Aberle

Location: Williston LEC

Intoxilyzer Serial No.: 80-004196

Lot Number: 659357

Cylinder Number: 45

Expiration Date: 09/03/2011

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0. 080 %

Test 2: 0. 079 %

(ATTACH TEST RECORD)

Test 3: 0. 080 %

Darcy W Aberle  
 Chemical Test Operator Signature

10/01/09  
 Date Cylinder Changed

Deb Kashner  
 Form 105-G  
 Toxicology Laboratory  
 Crime Laboratory Division

06 Oct 09

80-004196  
 P. 1 of 2 yr

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004196  
Location = WILL      8164.12.00 06/09  
10/01/2009      17:32

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:33
02 Std. Gas	0.080	17:33
03 Room Air	0.000	17:34
04 Std. Gas	0.079	17:34
05 Room Air	0.000	17:35
06 Std. Gas	0.080	17:35
07 Room Air	0.000	17:35

Lot No = 659357

Cyl No = 45

Exp Date = 09/03/2011

County = 53

Oper No. = 010101

*Darcy W Aberle*

Operator Signature  
DARCY W. ABERLE

Remarks:

Test 1

ACA

for Form 105-G  
Form 106-I8000

80-004196  
P. 2 of 2 FM