



**ETHANOL BREATH STANDARD CYLINDER REPORT**  
 OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION  
 Toxicology Section/Breath Alcohol Program  
 SFN 59282 (5-2011)

Chemical Test Operator Name (Print) <i>Richard Johnson</i>		
Location <i>WASH</i>	Intoxilyzer® Serial Number <i>80-003057</i>	
Gas Lot Number <i>16512080 A2</i>	Gas Cylinder Number <i>54</i>	Gas Expiration Date <i>8/1/2014</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).  
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).  
 A. *0.079* AC  
 B. *0.078* AC  
 C. *0.078* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
  - A. Ethanol Breath Standard Cylinder Report (SFN59282, Form 105-G).
  - B. ACA Test Record.

Chemical Test Operator Signature <i>Richard Johnson</i>	Date <i>10/1/2013</i>
Reviewed By (Crime Laboratory Use Only) <i>Deb Kashner</i>	Date <i>11 Oct 2013</i>

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003057  
Location = WASH      8164.13.00 06/09  
10/01/2013      14:01

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:01
02 Std. Gas	0.079	14:02
03 Room Air	0.000	14:02
04 Std. Gas	0.078	14:03
05 Room Air	0.000	14:03
06 Std. Gas	0.078	14:04
07 Room Air	0.000	14:04


Lot No = 16512080A2

Cyl No = 54

Exp Date = 08/01/2014

County = 28

Oper No. = 118707

  
Operator Signature  
RICHARD JOHNSON

Remarks:

Form 106-I8000