



OFFICE OF ATTORNEY GENERAL
 Crime Laboratory Division
 2641 East Main Avenue
 P.O. Box 937
 Bismarck, North Dakota 58502-0937

TEL 701-328-6159

ETHANOL BREATH STANDARD CYLINDER INSTALLATION FORM

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: David J. Leis

Location: Cavalier County SD Langdon

Intoxilyzer Serial No.: 80-004200

Lot Number: 659358

Cylinder Number: 53

Expiration Date: 09/03/11

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0. 081 %

Test 2: 0. 080 %

Test 3: 0. 080 %

(ATTACH TEST RECORD)

David J. Leis
 Chemical Test Operator Signature

10/02/09
 Date Cylinder Changed

Deb Kashner

13 Oct 09

Form 105-G
 Toxicology Laboratory
 Crime Laboratory Division

80-004200
 P. 1 of 2 KM


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004200
Location = LANG 8164.12.00 06/09
10/02/2009 14:17

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:18
02 Std. Gas	0.081	14:18
03 Room Air	0.000	14:19
04 Std. Gas	0.080	14:19
05 Room Air	0.000	14:19
06 Std. Gas	0.080	14:20
07 Room Air	0.000	14:20

Lot No = 659358
Cyl No = 53
Exp Date = 05/28/2011
County = 10 Oper No. = 067518


Operator Signature
DAVID J. ZEIS

Remarks:

Form 106-I8000

80-004200
P. 2 of 2 AL